

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PO107564.D Time Analyzed: 17:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.521	5.423	5.623	452.620	500.000	-9.5
Aroclor-1016-2	5.544	5.445	5.645	444.440	500.000	-11.1
Aroclor-1016-3	5.606	5.507	5.707	429.440	500.000	-14.1
Aroclor-1016-4	5.703	5.604	5.804	443.320	500.000	-11.3
Aroclor-1016-5	5.998	5.899	6.099	468.170	500.000	-6.4
Aroclor-1260-1	7.126	7.026	7.226	532.120	500.000	6.4
Aroclor-1260-2	7.383	7.282	7.482	496.450	500.000	-0.7
Aroclor-1260-3	7.746	7.643	7.843	509.430	500.000	1.9
Aroclor-1260-4	7.970	7.868	8.068	502.340	500.000	0.5
Aroclor-1260-5	8.284	8.181	8.381	476.130	500.000	-4.8
Decachlorobiphenyl	10.067	9.958	10.158	43.470	50.000	-13.1
Tetrachloro-m-xylene	4.371	4.274	4.474	51.160	50.000	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL01 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PO107564.D Time Analyzed: 17:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	516.190	500.000	3.2
Aroclor-1016-2	4.744	4.645	4.845	527.840	500.000	5.6
Aroclor-1016-3	4.920	4.821	5.021	510.100	500.000	2.0
Aroclor-1016-4	4.962	4.862	5.062	472.970	500.000	-5.4
Aroclor-1016-5	5.175	5.076	5.276	551.240	500.000	10.2
Aroclor-1260-1	6.205	6.107	6.307	512.570	500.000	2.5
Aroclor-1260-2	6.393	6.294	6.494	527.590	500.000	5.5
Aroclor-1260-3	6.547	6.448	6.648	479.320	500.000	-4.1
Aroclor-1260-4	7.018	6.918	7.118	454.610	500.000	-9.1
Aroclor-1260-5	7.258	7.159	7.359	470.350	500.000	-5.9
Decachlorobiphenyl	8.641	8.539	8.739	41.100	50.000	-17.8
Tetrachloro-m-xylene	3.645	3.545	3.745	51.240	50.000	2.5