

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PO107576.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	516.630	500.000	3.3
Aroclor-1016-2	5.545	5.445	5.645	501.620	500.000	0.3
Aroclor-1016-3	5.607	5.507	5.707	502.840	500.000	0.6
Aroclor-1016-4	5.704	5.604	5.804	521.720	500.000	4.3
Aroclor-1016-5	5.999	5.899	6.099	530.540	500.000	6.1
Aroclor-1260-1	7.126	7.026	7.226	492.210	500.000	-1.6
Aroclor-1260-2	7.384	7.282	7.482	502.650	500.000	0.5
Aroclor-1260-3	7.746	7.643	7.843	517.710	500.000	3.5
Aroclor-1260-4	7.970	7.868	8.068	510.880	500.000	2.2
Aroclor-1260-5	8.285	8.181	8.381	533.790	500.000	6.8
Decachlorobiphenyl	10.068	9.958	10.158	45.800	50.000	-8.4
Tetrachloro-m-xylene	4.371	4.274	4.474	51.830	50.000	3.7

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PO107576.D Time Analyzed: 20:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.725	4.626	4.826	513.890	500.000	2.8
Aroclor-1016-2	4.743	4.645	4.845	524.040	500.000	4.8
Aroclor-1016-3	4.920	4.821	5.021	507.240	500.000	1.4
Aroclor-1016-4	4.961	4.862	5.062	470.080	500.000	-6.0
Aroclor-1016-5	5.174	5.076	5.276	537.820	500.000	7.6
Aroclor-1260-1	6.205	6.107	6.307	504.210	500.000	0.8
Aroclor-1260-2	6.392	6.294	6.494	543.000	500.000	8.6
Aroclor-1260-3	6.546	6.448	6.648	510.980	500.000	2.2
Aroclor-1260-4	7.018	6.918	7.118	512.450	500.000	2.5
Aroclor-1260-5	7.258	7.159	7.359	542.060	500.000	8.4
Decachlorobiphenyl	8.641	8.539	8.739	47.940	50.000	-4.1
Tetrachloro-m-xylene	3.643	3.545	3.745	51.350	50.000	2.7