

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PP068318.D Time Analyzed: 13:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.926	5.819	6.019	492.820	500.000	-1.4
Aroclor-1016-2	5.947	5.842	6.042	499.970	500.000	0.0
Aroclor-1016-3	6.011	5.905	6.105	554.660	500.000	10.9
Aroclor-1016-4	6.110	6.003	6.203	535.520	500.000	7.1
Aroclor-1016-5	6.404	6.298	6.498	460.110	500.000	-8.0
Aroclor-1260-1	7.527	7.422	7.622	463.250	500.000	-7.4
Aroclor-1260-2	7.779	7.675	7.875	466.490	500.000	-6.7
Aroclor-1260-3	8.141	8.037	8.237	486.930	500.000	-2.6
Aroclor-1260-4	8.380	8.275	8.475	482.310	500.000	-3.5
Aroclor-1260-5	8.717	8.612	8.812	481.070	500.000	-3.8
Decachlorobiphenyl	10.676	10.569	10.769	49.830	50.000	-0.3
Tetrachloro-m-xylene	4.762	4.655	4.855	50.980	50.000	2.0

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL03 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PP068318.D Time Analyzed: 13:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.161	5.061	5.261	452.360	500.000	-9.5
Aroclor-1016-2	5.181	5.081	5.281	464.930	500.000	-7.0
Aroclor-1016-3	5.361	5.262	5.462	461.260	500.000	-7.7
Aroclor-1016-4	5.401	5.301	5.501	468.330	500.000	-6.3
Aroclor-1016-5	5.619	5.520	5.720	468.640	500.000	-6.3
Aroclor-1260-1	6.662	6.564	6.764	457.830	500.000	-8.4
Aroclor-1260-2	6.848	6.750	6.950	458.900	500.000	-8.2
Aroclor-1260-3	7.006	6.908	7.108	461.830	500.000	-7.6
Aroclor-1260-4	7.480	7.383	7.583	466.930	500.000	-6.6
Aroclor-1260-5	7.718	7.622	7.822	465.360	500.000	-6.9
Decachlorobiphenyl	9.212	9.119	9.319	48.250	50.000	-3.5
Tetrachloro-m-xylene	4.054	3.952	4.152	47.470	50.000	-5.1