

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PP068333.D Time Analyzed: 18:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.925	5.819	6.019	511.120	500.000	2.2
Aroclor-1016-2	5.947	5.842	6.042	519.840	500.000	4.0
Aroclor-1016-3	6.011	5.905	6.105	511.480	500.000	2.3
Aroclor-1016-4	6.109	6.003	6.203	493.890	500.000	-1.2
Aroclor-1016-5	6.403	6.298	6.498	495.560	500.000	-0.9
Aroclor-1260-1	7.526	7.422	7.622	472.810	500.000	-5.4
Aroclor-1260-2	7.779	7.675	7.875	467.190	500.000	-6.6
Aroclor-1260-3	8.140	8.037	8.237	481.420	500.000	-3.7
Aroclor-1260-4	8.379	8.275	8.475	479.970	500.000	-4.0
Aroclor-1260-5	8.716	8.612	8.812	477.350	500.000	-4.5
Decachlorobiphenyl	10.675	10.569	10.769	49.950	50.000	-0.1
Tetrachloro-m-xylene	4.761	4.655	4.855	52.760	50.000	5.5

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: P4620 SAS No.: P4620 SDG NO.: P4620

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/08/2024 10/08/2024

Client Sample No.: CCAL04 Date Analyzed: 10/31/2024

Lab Sample No.: AR1660CCC500 Data File : PP068333.D Time Analyzed: 18:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.159	5.061	5.261	443.170	500.000	-11.4
Aroclor-1016-2	5.179	5.081	5.281	456.260	500.000	-8.7
Aroclor-1016-3	5.359	5.262	5.462	463.590	500.000	-7.3
Aroclor-1016-4	5.398	5.301	5.501	473.650	500.000	-5.3
Aroclor-1016-5	5.617	5.520	5.720	465.480	500.000	-6.9
Aroclor-1260-1	6.660	6.564	6.764	449.730	500.000	-10.1
Aroclor-1260-2	6.846	6.750	6.950	452.000	500.000	-9.6
Aroclor-1260-3	7.004	6.908	7.108	452.640	500.000	-9.5
Aroclor-1260-4	7.478	7.383	7.583	451.050	500.000	-9.8
Aroclor-1260-5	7.716	7.622	7.822	445.310	500.000	-10.9
Decachlorobiphenyl	9.209	9.119	9.319	47.940	50.000	-4.1
Tetrachloro-m-xylene	4.052	3.952	4.152	47.460	50.000	-5.1