

Surrogate Summary

SDG No.: **P4620**

Client: **Weston Solutions, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		60	140
		Decachlorobiphenyl	1	20	23.4	117		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
		Decachlorobiphenyl	2	20	22.8	114		60	140
I.BLK-PO107568.D	PIBLK-PO107568.D	Tetrachloro-m-xylene	1	20	19.7	98		60	140
		Decachlorobiphenyl	1	20	18.3	92		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	17.1	85		60	140
P4620-16	P001-CF13-01	Tetrachloro-m-xylene	1	20	16.8	84		10	157
		Decachlorobiphenyl	1	20	17.2	86		10	173
		Tetrachloro-m-xylene	2	20	21.6	108		10	157
		Decachlorobiphenyl	2	20	16.3	81		10	173
P4620-18	P001-CF14-01	Tetrachloro-m-xylene	1	20	23.2	116		10	157
		Decachlorobiphenyl	1	20	19.5	98		10	173
		Tetrachloro-m-xylene	2	20	23.7	119		10	157
		Decachlorobiphenyl	2	20	18.7	93		10	173
P4620-20	P001-CF15-01	Tetrachloro-m-xylene	1	20	19.9	99		10	157
		Decachlorobiphenyl	1	20	17.2	86		10	173
		Tetrachloro-m-xylene	2	20	21.4	107		10	157
		Decachlorobiphenyl	2	20	16.5	82		10	173
P4620-25	P001-CF16-01	Tetrachloro-m-xylene	1	20	20.3	101		10	157
		Decachlorobiphenyl	1	20	17.6	88		10	173
		Tetrachloro-m-xylene	2	20	20.9	104		10	157
		Decachlorobiphenyl	2	20	17.1	85		10	173
P4620-26MS	P001-CF16-01MS	Tetrachloro-m-xylene	1	20	21.2	106		10	157
		Decachlorobiphenyl	1	20	16.0	80		10	173
		Tetrachloro-m-xylene	2	20	20.5	103		10	157
		Decachlorobiphenyl	2	20	15.6	78		10	173
P4620-27MSD	P001-CF16-01MSD	Tetrachloro-m-xylene	1	20	21.3	106		10	157
		Decachlorobiphenyl	1	20	16.1	81		10	173
		Tetrachloro-m-xylene	2	20	20.6	103		10	157
		Decachlorobiphenyl	2	20	15.9	79		10	173
P4620-30	P001-CF16-02	Tetrachloro-m-xylene	1	20	21.3	106		10	157
		Decachlorobiphenyl	1	20	18.5	93		10	173
		Tetrachloro-m-xylene	2	20	21.7	109		10	157
		Decachlorobiphenyl	2	20	18.4	92		10	173
I.BLK-PO107580.D	PIBLK-PO107580.D	Tetrachloro-m-xylene	1	20	20.0	100		60	140
		Decachlorobiphenyl	1	20	19.4	97		60	140
		Tetrachloro-m-xylene	2	20	19.6	98		60	140
		Decachlorobiphenyl	2	20	19.8	99		60	140
I.BLK-PP067586.D	PIBLK-PP067586.D	Tetrachloro-m-xylene	1	20	21.1	105		60	140

Surrogate Summary

SDG No.: **P4620**

Client: **Weston Solutions, Inc.**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP067586.D	PIBLK-PP067586.D	Decachlorobiphenyl	1	20	23.6	118		60	140
		Tetrachloro-m-xylene	2	20	22.1	110		60	140
I.BLK-PP068322.D	PIBLK-PP068322.D	Decachlorobiphenyl	2	20	23.5	117		60	140
		Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	16.5	83		60	140
		Decachlorobiphenyl	2	20	18.4	92		60	140
PB164566BL	PB164566BL	Tetrachloro-m-xylene	1	20	20.3	101		10	157
		Decachlorobiphenyl	1	20	21.9	110		10	173
		Tetrachloro-m-xylene	2	20	18.1	91		10	157
		Decachlorobiphenyl	2	20	21.2	106		10	173
PB164566BS	PB164566BS	Tetrachloro-m-xylene	1	20	20.5	102		10	157
		Decachlorobiphenyl	1	20	22.5	112		10	173
		Tetrachloro-m-xylene	2	20	19.0	95		10	157
		Decachlorobiphenyl	2	20	21.5	107		10	173
PB164566TB	PB164566TB	Tetrachloro-m-xylene	1	20	20.7	103		10	157
		Decachlorobiphenyl	1	20	22.3	112		10	173
		Tetrachloro-m-xylene	2	20	18.3	91		10	157
		Decachlorobiphenyl	2	20	21.0	105		10	173
P4620-02	P001-CF06-01	Tetrachloro-m-xylene	1	20	22.6	113		10	157
		Decachlorobiphenyl	1	20	22.0	110		10	173
		Tetrachloro-m-xylene	2	20	20.4	102		10	157
		Decachlorobiphenyl	2	20	20.1	100		10	173
P4620-04	P001-CF07-01	Tetrachloro-m-xylene	1	20	22.9	115		10	157
		Decachlorobiphenyl	1	20	19.8	99		10	173
		Tetrachloro-m-xylene	2	20	20.9	104		10	157
		Decachlorobiphenyl	2	20	20.3	102		10	173
P4620-06	P001-CF08-01	Tetrachloro-m-xylene	1	20	20.7	103		10	157
		Decachlorobiphenyl	1	20	30.2	151		10	173
		Tetrachloro-m-xylene	2	20	19.9	99		10	157
		Decachlorobiphenyl	2	20	20.5	102		10	173
P4620-08	P001-CF09-01	Tetrachloro-m-xylene	1	20	21.1	106		10	157
		Decachlorobiphenyl	1	20	21.9	110		10	173
		Tetrachloro-m-xylene	2	20	19.8	99		10	157
		Decachlorobiphenyl	2	20	21.3	107		10	173
P4620-10	P001-CF10-01	Tetrachloro-m-xylene	1	20	21.6	108		10	157
		Decachlorobiphenyl	1	20	24.8	124		10	173
		Tetrachloro-m-xylene	2	20	20.0	100		10	157
		Decachlorobiphenyl	2	20	23.9	119		10	173
P4620-12	P001-CF11-01	Tetrachloro-m-xylene	1	20	18.9	95		10	157
		Decachlorobiphenyl	1	20	16.4	82		10	173

Surrogate Summary

SDG No.: P4620

Client: Weston Solutions, Inc.

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P4620-12	P001-CF11-01	Tetrachloro-m-xylene	2	20	17.8	89		10	157
		Decachlorobiphenyl	2	20	16.3	81		10	173
P4620-14	P001-CF12-01	Tetrachloro-m-xylene	1	20	20.5	102		10	157
		Decachlorobiphenyl	1	20	20.4	102		10	173
		Tetrachloro-m-xylene	2	20	19.7	98		10	157
		Decachlorobiphenyl	2	20	20.3	101		10	173
I.BLK-PP068337.D	PIBLK-PP068337.D	Tetrachloro-m-xylene	1	20	18.6	93		60	140
		Decachlorobiphenyl	1	20	19.7	98		60	140
		Tetrachloro-m-xylene	2	20	16.5	82		60	140
		Decachlorobiphenyl	2	20	18.5	92		60	140