

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51835 MA No. : SDG No.: MBH6M7
 Initial Calibration Verification Source : MP83168
 Continuing Calibration Verification Source : MP83170
 Run Batch: LB133364 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV077				ID: CCV057				ID: CCV058		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	95		5.0	5.3	106		5.2	104	