EPA SAMPLE NO.

MBH6N6S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech C	onsulting Group	Contract:	EPW14030		
Lab Code:	ACE	Case No.: 51835	MA No. :		SDG No.:	МВН6М7
Matrix :	Water		Analytical	Method:	CVAA	
% Solids:						
Concentrat	ion Units (µ	ıg/L, mg/L or mg/kg dry	weight): ug/l	Ĺ		

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	2.6		0.11	J	1.0	244	*