


LB133364

SFAM01.1 INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
B= 2.3484e-004 *slope*
C= 8.8856e-002 *intercept*
Rho= 0.9996024
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.090	0.090	3	0.000	3					=
0.05	0.050					442	0				24
0.20	0.200	0.248	0.048	678	0.0 %	678					-8
2.50	2.500	2.307	-0.193	9447	0.0 %	9447					+1
5.00	5.000	5.072	0.072	21219	0.0 %	21219					-1
7.50	7.500	7.413	-0.087	31190	0.0 %	31190					1
10.0	10.000	10.070	0.070	42501	0.0 %	42501					

LB133364
INSTRUMENT ID: CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0		3 -		0 SFAM01.1	PPB	11/8/2024 15:55	Std
0.2 S01	678	-		0.2 SFAM01.1	PPB	11/8/2024 15:57	Std
2.5 S02	9447	-		2.5 SFAM01.1	PPB	11/8/2024 15:59	Std
5 S03	21219	-		5 SFAM01.1	PPB	11/8/2024 16:02	Std
7.5 S04	31190	-		7.5 SFAM01.1	PPB	11/8/2024 16:04	Std
10 S05	42501	-		10 SFAM01.1	PPB	11/8/2024 16:12	Std
ICV077	ICV077	15752	3.788 -	SFAM01.1	PPB	11/8/2024 16:15	SMPL
ICB077	ICB077	-212	0.0391 -	SFAM01.1	PPB	11/8/2024 16:17	SMPL
CCV057	CCV057	22147	5.2898 -	SFAM01.1	PPB	11/8/2024 16:19	SMPL
CCB057	CCB057	-156	0.0522 -	SFAM01.1	PPB	11/8/2024 16:21	SMPL
PB164805TB	LEB805	-23	0.0835 -	SFAM01.1	PPB	11/8/2024 16:29	SMPL
P4623-02	MBH6M8	284	0.1555 -	SFAM01.1	PPB	11/8/2024 16:33	SMPL
P4623-03	MBH6M9	48	0.1001 -	SFAM01.1	PPB	11/8/2024 16:36	SMPL
P4623-04	MBH6N0	48	0.1001 -	SFAM01.1	PPB	11/8/2024 16:38	SMPL
P4623-05	MBH6N1	161	0.1267 -	SFAM01.1	PPB	11/8/2024 16:40	SMPL
P4623-06	MBH8M6	46	0.0997 -	SFAM01.1	PPB	11/8/2024 16:42	SMPL
P4623-07	MBH6N2	71	0.1055 -	SFAM01.1	PPB	11/8/2024 16:45	SMPL
P4623-08	MBH6N3	100	0.1123 -	SFAM01.1	PPB	11/8/2024 16:47	SMPL
P4623-09	MBH6N4	139	0.1215 -	SFAM01.1	PPB	11/8/2024 16:49	SMPL
P4623-10	MBH6N5	76	0.1067 -	SFAM01.1	PPB	11/8/2024 16:52	SMPL
P4623-11	MBH6N6	100	0.1123 -	SFAM01.1	PPB	11/8/2024 16:54	SMPL
P4623-12	MBH6N6D	17	0.0928 -	SFAM01.1	PPB	11/8/2024 16:56	SMPL
P4623-13	MBH6N6S	10482	2.5504 -	SFAM01.1	PPB	11/8/2024 16:58	SMPL
P4623-14	MBH6N7	72	0.1058 -	SFAM01.1	PPB	11/8/2024 17:01	SMPL
P4623-01	MBH6M7	36	0.0973 -	SFAM01.1	PPB	11/8/2024 17:03	SMPL
PB164805BL	PBW805	10	0.0912 -	SFAM01.1	PPB	11/8/2024 17:05	SMPL
CCV058	CCV058	21695	5.1837 -	SFAM01.1	PPB	11/8/2024 17:07	SMPL
CCB058	CCB058	-76	0.071 -	SFAM01.1	PPB	11/8/2024 17:10	SMPL