

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE001	TUNE	11/09/24 20:58		Jaswal	OK
2	S00	S0	CAL1	11/09/24 21:53		Jaswal	OK
3	S02	S02	CAL3	11/09/24 21:59		Jaswal	OK
4	S03	S03	CAL4	11/09/24 22:03		Jaswal	OK
5	S04	S04	CAL5	11/09/24 22:05		Jaswal	OK
6	S05	S05	CAL6	11/09/24 22:08		Jaswal	OK
7	S06	S06	CAL7	11/09/24 22:11		Jaswal	OK
8	S07	S07	CAL8	11/09/24 22:14		Jaswal	OK
9	S08	S08	CAL9	11/09/24 22:16		Jaswal	OK
10	ICV001	ICV001	ICV	11/09/24 22:40	Inst. paused for calibration review	Jaswal	OK
11	ICB001	ICB001	ICB	11/09/24 22:43		Jaswal	OK
12	ICSA001	ICSA001	ICSA	11/09/24 22:47		Jaswal	OK
13	ICSAB001	ICSAB001	ICSAB	11/09/24 22:50		Jaswal	OK
14	CCV001	CCV001	CCV	11/09/24 23:02	Inst. paused for review	Jaswal	OK
15	CCB001	CCB001	CCB	11/09/24 23:06		Jaswal	OK
16	PB164720BL	PBS720	MB	11/09/24 23:16		Jaswal	OK
17	PB164720BS	LCS720	LCS	11/09/24 23:19		Jaswal	OK
18	PB164583BL	PBS583	MB	11/09/24 23:23		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

19	PB164583BS	LCS583	LCS	11/09/24 23:26		Jaswal	OK
20	PB164583BSD	LCSD583	LCSD	11/09/24 23:29		Jaswal	OK
21	P4624-01	MBH6M7	SAM	11/09/24 23:33	INT_STD 89Y(1) Fail	Jaswal	Dilution
22	P4624-02	MBH6M8	SAM	11/09/24 23:36	INT_STD 6Li Fail	Jaswal	Dilution
23	P4624-03	MBH6M9	SAM	11/09/24 23:39	INT_STD 6Li Fail	Jaswal	Dilution
24	P4624-04	MBH6N0	SAM	11/09/24 23:42	INT_STD 6Li Fail	Jaswal	Dilution
25	P4624-05	MBH6N1	SAM	11/09/24 23:45	INT_STD 6Li Fail	Jaswal	Dilution
26	P4624-06	MBH8M6	SAM	11/09/24 23:48	INT_STD 6Li Fail	Jaswal	Dilution
27	P4624-07	MBH6N2	SAM	11/09/24 23:51	INT_STD 6Li Fail	Jaswal	Dilution
28	P4624-08	MBH6N3	SAM	11/09/24 23:54		Jaswal	OK
29	P4624-09	MBH6N4	SAM	11/09/24 23:57		Jaswal	OK
30	P4624-10	MBH6N5	SAM	11/10/24		Jaswal	OK
31	P4624-11	MBH6N6	SAM	11/10/24 00:03	INT_STD 6Li Fail	Jaswal	OK
32	P4624-12	MBH6N6D	DUP	11/10/24 00:06	INT_STD 6Li Fail	Jaswal	OK
33	P4624-11L	MBH6N6L	SD	11/10/24 00:09	INT_STD 6Li Fail	Jaswal	OK
34	P4624-13	MBH6N6S	MS	11/10/24 00:13	INT_STD 6Li Fail	Jaswal	OK
35	P4624-14	MBH6N7	SAM	11/10/24 00:16		Jaswal	OK
36	P4521-01	MYE4N5	SAM	11/10/24 00:19		Jaswal	OK
37	P4521-02	MYE4N6	SAM	11/10/24 00:22		Jaswal	OK
38	P4521-03	MYE4N7	SAM	11/10/24 00:25		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

39	P4521-04	MYE4N8	SAM	11/10/24 00:28		Jaswal	OK
40	P4521-05	MYE4N9	SAM	11/10/24 00:31		Jaswal	OK
41	P4521-06	MYE4P0	SAM	11/10/24 00:34		Jaswal	OK
42	P4521-07	MYE4P1	SAM	11/10/24 00:37	INT_STD 6Li Fail	Jaswal	Dilution
43	P4521-08	MYE4P2	SAM	11/10/24 00:40	INT_STD 45Sc(2),89Y(1&2) Fail, Ba,Co,Ni,V High	Jaswal	Dilution
44	P4521-09	MYE4P2D	DUP	11/10/24 00:43	INT_STD 45Sc(2),89Y(1&2) Fail, Ba,Co,Ni,V High	Jaswal	Dilution
45	CCV002	CCV002	CCV	11/10/24 01:00		Jaswal	OK
46	CCB002	CCB002	CCB	11/10/24 01:02		Jaswal	OK
47	P4521-08L	MYE4P2L	SD	11/10/24 01:06	INT_STD 45Sc(2),89Y(1&2) Fail, Ba,Co,Ni,V High	Jaswal	Dilution
48	P4521-10	MYE4P2S	MS	11/10/24 01:09	INT_STD 45Sc(2),89Y(1&2) Fail, Ba,Co,Ni,V High	Jaswal	Dilution
49	P4521-08MS	MYE4P2SRE	MS	11/10/24 01:12	INT_STD 45Sc(2),89Y(1&2) Fail, Ba,Co,Ni,V High	Jaswal	Dilution
50	P4521-11	MYE4P3	SAM	11/10/24 01:14	INT_STD 6Li Fail	Jaswal	Dilution
51	P4521-12	MYE4P4	SAM	11/10/24 01:17	INT_STD 6Li Fail	Jaswal	Dilution
52	P4521-13	MYE4P5	SAM	11/10/24 01:20	INT_STD 6Li Fail	Jaswal	Dilution
53	P4521-14	MYE4Q2	SAM	11/10/24 01:23	INT_STD 6Li Fail	Jaswal	Dilution
54	P4521-15	MYE4Q3	SAM	11/10/24 01:26	INT_STD 6Li Fail	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

55	P4521-16	MYE4Q4	SAM	11/10/24 01:29		Jaswal	OK
56	P4521-17	MYE4Q5	SAM	11/10/24 01:32		Jaswal	OK
57	P4521-18	MYE4Q6	SAM	11/10/24 01:35	INT_STD 6Li Fail	Jaswal	Dilution
58	P4521-19	MYE4Q7	SAM	11/10/24 01:38		Jaswal	OK
59	P4521-20	MYE4Q8	SAM	11/10/24 01:42		Jaswal	OK
60	P4521-21	MYE4Q9	SAM	11/10/24 01:45		Jaswal	OK
61	P4521-22	MYE4R0	SAM	11/10/24 01:48		Jaswal	OK
62	P4624-01DL	MBH6M7	SAM	11/10/24 01:55	2X For INT_STD 89Y(1)	Jaswal	Confirms
63	P4624-02DL	MBH6M8	SAM	11/10/24 01:58	2X For INT_STD 6Li	Jaswal	Confirms
64	P4624-03DL	MBH6M9	SAM	11/10/24 02:01	2X For INT_STD 6Li	Jaswal	Confirms
65	P4624-04DL	MBH6N0	SAM	11/10/24 02:04	2X For INT_STD 6Li	Jaswal	Confirms
66	P4624-05DL	MBH6N1	SAM	11/10/24 02:07	2X For INT_STD 6Li	Jaswal	Confirms
67	P4624-06DL	MBH8M6	SAM	11/10/24 02:10	2X For INT_STD 6Li	Jaswal	Confirms
68	P4624-07DL	MBH6N2	SAM	11/10/24 02:14	2X For INT_STD 6Li	Jaswal	Confirms
69	P4624-08DL	MBH6N3	SAM	11/10/24 02:17	Not Required	Jaswal	Not Ok
70	P4624-09DL	MBH6N4	SAM	11/10/24 02:20	Not Required	Jaswal	Not Ok
71	P4624-10DL	MBH6N5	SAM	11/10/24 02:23	Not Required	Jaswal	Not Ok
72	P4624-11DL	MBH6N6	SAM	11/10/24 02:26	2X For INT_STD 6Li , Still Fail So Report From Original	Jaswal	Not Ok

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

73	P4624-12DL	MBH6N6D	DUP	11/10/24 02:29	2X For INT_STD 6Li , Still Fail So Report From Original	Jaswal	Not Ok
74	P4624-11LDL	MBH6N6L	SD	11/10/24 02:32	2X For INT_STD 6Li , Still Fail So Report From Original	Jaswal	Not Ok
75	P4624-13DL	MBH6N6S	MS	11/10/24 02:36	2X For INT_STD 6Li , Still Fail So Report From Original	Jaswal	Not Ok
76	P4624-14DL	MBH6N7	SAM	11/10/24 02:39	Not Required	Jaswal	Not Ok
77	CCV003	CCV003	CCV	11/10/24 02:42		Jaswal	OK
78	CCB003	CCB003	CCB	11/10/24 02:57	Inst. paused for review	Jaswal	OK
79	P4521-01DL	MYE4N5	SAM	11/10/24 03:00	Not Required	Jaswal	Not Ok
80	P4521-02DL	MYE4N6	SAM	11/10/24 03:03	Not Required	Jaswal	Not Ok
81	P4521-03DL	MYE4N7	SAM	11/10/24 03:06	Not Required	Jaswal	Not Ok
82	P4521-04DL	MYE4N8	SAM	11/10/24 03:09	Not Required	Jaswal	Not Ok
83	P4521-05DL	MYE4N9	SAM	11/10/24 03:12	Not Required	Jaswal	Not Ok
84	P4521-06DL	MYE4P0	SAM	11/10/24 03:15	Not Required	Jaswal	Not Ok
85	P4521-07DL	MYE4P1	SAM	11/10/24 03:19	2X For INT_STD 6Li	Jaswal	Confirms
86	P4521-08DL	MYE4P2	SAM	11/10/24 03:22	2X For INT_STD 45Sc(2),89Y(1&2), 2X For Ba,Co,Ni,V	Jaswal	Confirms
87	P4521-09DL	MYE4P2D	DUP	11/10/24 03:25	2X For INT_STD 45Sc(2),89Y(1&2), 2X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM

STD. NAME	STD REF.#
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016
ICV Standard	MP83041
CCV Standard	MP83042
ICSA Standard	MP83043,MP83044
CRI Standard	
LCS Standard	MP83046
Chk Standard	MP83050,MP83049

88	P4521-08LDL	MYE4P2L	SD	11/10/24 03:28	2X For INT_STD 45Sc(2),89Y(1&2), 2X For Ba,Co,Ni,V	Jaswal	Confirms
89	P4521-10DL	MYE4P2S	MS	11/10/24 03:31	2X For INT_STD 45Sc(2),89Y(1&2), 2X For Ba,Co,Ni,V	Jaswal	Confirms
90	P4521-08MSDL	MYE4P2SRE	MS	11/10/24 03:34	2X For INT_STD 45Sc(2),89Y(1&2), 2X For Ba,Co,Ni,V	Jaswal	Confirms
91	P4521-11DL	MYE4P3	SAM	11/10/24 03:37	2X For INT_STD 6Li	Jaswal	Confirms
92	P4521-12DL	MYE4P4	SAM	11/10/24 03:40	2X For INT_STD 6Li	Jaswal	Confirms
93	P4521-13DL	MYE4P5	SAM	11/10/24 03:43	2X For INT_STD 6Li	Jaswal	Confirms
94	P4521-14DL	MYE4Q2	SAM	11/10/24 03:46	2X For INT_STD 6Li	Jaswal	Confirms
95	P4521-15DL	MYE4Q3	SAM	11/10/24 03:49	2X For INT_STD 6Li	Jaswal	Confirms
96	P4521-16DL	MYE4Q4	SAM	11/10/24 03:52	Not Required	Jaswal	Not Ok
97	P4521-17DL	MYE4Q5	SAM	11/10/24 03:55	Not Required	Jaswal	Not Ok
98	P4521-18DL	MYE4Q6	SAM	11/10/24 03:58	2X For INT_STD 6Li	Jaswal	Confirms
99	P4521-19DL	MYE4Q7	SAM	11/10/24 04:01	Not Required	Jaswal	Not Ok
100	P4521-20DL	MYE4Q8	SAM	11/10/24 04:04	Not Required	Jaswal	Not Ok
101	P4521-21DL	MYE4Q9	SAM	11/10/24 04:07	Not Required	Jaswal	Not Ok
102	P4521-22DL	MYE4R0	SAM	11/10/24 04:11	Not Required	Jaswal	Not Ok
103	P4521-08DL2	MYE4P2	SAM	11/10/24 04:14	Not Required	Jaswal	Not Ok
104	P4521-09DL2	MYE4P2D	DUP	11/10/24 04:17	Not Required	Jaswal	Not Ok
105	CCV004	CCV004	CCV	11/10/24 04:20		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB133380

Review By	Mohan Bera	Review On	11/11/2024 4:25:52 AM
Supervise By	Sarabjit Jaswal	Supervise On	11/11/2024 4:27:25 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83014,MP83039,MP83038,MP83036,MP83034,MP83035,MP83033,MP83032,MP83016		
ICV Standard	MP83041		
CCV Standard	MP83042		
ICSA Standard	MP83043,MP83044		
CRI Standard			
LCS Standard	MP83046		
Chk Standard	MP83050,MP83049		

106	CCB004	CCB004	CCB	11/10/24 04:23		Jaswal	OK
107	P4521-08LDL2	MYE4P2L	SD	11/10/24 04:26	Not Required	Jaswal	Not Ok
108	P4521-10DL2	MYE4P2S	MS	11/10/24 04:29	Not Required	Jaswal	Not Ok
109	P4521-08MSDL2	MYE4P2SRE	MS	11/10/24 04:32	Not Required	Jaswal	Not Ok
110	CCV005	CCV005	CCV	11/10/24 04:35		Jaswal	OK
111	CCB005	CCB005	CCB	11/10/24 04:38		Jaswal	OK