

Preparation Blank Summary

Client: LEO Ingwer, Inc.

SDG No.: P4632

Project: Waste Water 2024

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: Cyanide	PB164714BL mg/L	< 0.0025	0.0025	U	0.00093	0.005	11/11/2024