

Duplicate Sample Summary

Client:	LEO Ingwer, Inc.	SDG No.:	P4632
Project:	Waste Water 2024	Sample ID:	P4762-01
Client ID:	EFF-WWDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.0027	J	0.0028	J	1	4		11/11/2024

Duplicate Sample Summary

Client:	LEO Ingwer, Inc.	SDG No.:	P4632
Project:	Waste Water 2024	Sample ID:	P4762-01
Client ID:	EFF-WWMSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Cyanide	mg/L	+/-20	0.047		0.047		1	0		11/11/2024