

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	10/25/24
Project:	NJ Air PT	Date Received:	10/30/24
Client Sample ID:	E070-1100DL2	SDG No.:	P4633
Lab Sample ID:	P4633-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL041715.D	4		11/21/24 15:06	VL112124

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
TARGETS							
75-45-6	Chlorodifluoromethane	0.35	1.24	UD	1.24	7.07	ug/m3
75-71-8	Dichlorodifluoromethane	0.84	4.15	UD	4.15	9.89	ug/m3
64-17-5	Ethanol	1.30	2.45	UDQ	2.45	3.77	ug/m3
74-87-3	Chloromethane	0.39	0.81	UD	0.81	4.13	ug/m3
75-01-4	Vinyl Chloride	0.060	0.15	UD	0.15	0.31	ug/m3
141-78-6	Ethyl Acetate	0.32	1.27	UD	1.27	7.93	ug/m3
74-83-9	Bromomethane	0.56	2.17	UD	2.17	7.77	ug/m3
75-00-3	Chloroethane	49.2	130	D	1.79	5.28	ug/m3
109-99-9	Tetrahydrofuran	0.52	1.53	UD	1.53	5.90	ug/m3
75-69-4	Trichlorofluoromethane	0.64	3.60	UD	3.60	11.2	ug/m3
76-13-1	1,1,2-Trichlorotrifluoroethane	0.64	4.91	UD	4.91	15.3	ug/m3
76-14-2	Dichlorotetrafluoroethane	0.35	2.45	UD	2.45	14.0	ug/m3
593-60-2	Bromoethene	0.48	2.10	UD	2.10	8.74	ug/m3
115-07-1	Propene	0.60	1.03	UD	1.03	3.44	ug/m3
75-65-0	tert-Butyl alcohol	0.76	2.30	UD	2.30	6.06	ug/m3
142-82-5	Heptane	22.2	91.0	D	1.80	8.20	ug/m3
67-63-0	Isopropyl Alcohol	0.28	0.69	UD	0.69	4.92	ug/m3
75-35-4	1,1-Dichloroethene	19.0	75.3	D	2.22	7.93	ug/m3
67-64-1	Acetone	0.96	2.28	UD	2.28	4.75	ug/m3
75-15-0	Carbon Disulfide	0.64	1.99	UD	1.99	6.23	ug/m3
1634-04-4	Methyl tert-Butyl Ether	0.48	1.73	UD	1.73	7.21	ug/m3
75-09-2	Methylene Chloride	0.76	2.64	UD	2.64	6.95	ug/m3
156-60-5	trans-1,2-Dichloroethene	0.76	3.01	UD	3.01	7.93	ug/m3
108-05-4	Vinyl Acetate	0.64	2.30	UD	2.30	7.20	ug/m3
75-34-3	1,1-Dichloroethane	0.52	2.10	UD	2.10	8.09	ug/m3
110-82-7	Cyclohexane	0.88	3.03	UD	3.03	6.88	ug/m3
78-93-3	2-Butanone	0.48	1.42	UD	1.42	5.90	ug/m3
56-23-5	Carbon Tetrachloride	26.6	167	D	0.25	0.75	ug/m3
156-59-2	cis-1,2-Dichloroethene	0.36	1.43	UD	1.43	7.93	ug/m3
67-66-3	Chloroform	0.33	1.61	UD	1.61	9.77	ug/m3
71-55-6	1,1,1-Trichloroethane	0.040	0.22	UD	0.22	0.65	ug/m3
540-84-1	2,2,4-Trimethylpentane	0.40	1.87	UD	1.87	9.34	ug/m3
71-43-2	Benzene	0.35	1.12	UD	1.12	6.39	ug/m3
107-06-2	1,2-Dichloroethane	0.36	1.46	UD	1.46	8.09	ug/m3
79-01-6	Trichloroethene	0.070	0.38	UDQ	0.38	0.64	ug/m3

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	10/25/24
Project:	NJ Air PT	Date Received:	10/30/24
Client Sample ID:	E070-1100DL2	SDG No.:	P4633
Lab Sample ID:	P4633-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL041715.D	4		11/21/24 15:06	VL112124

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
78-87-5	1,2-Dichloropropane	0.44	2.03	UD	2.03	9.24	ug/m3
75-27-4	Bromodichloromethane	17.3	116	D	2.08	13.4	ug/m3
108-10-1	4-Methyl-2-Pentanone	0.37	1.52	UD	1.52	8.20	ug/m3
108-88-3	Toluene	41.2	155	D	1.66	7.54	ug/m3
10061-02-6	t-1,3-Dichloropropene	0.24	1.09	UD	1.09	9.08	ug/m3
10061-01-5	cis-1,3-Dichloropropene	0.26	1.18	UD	1.18	9.08	ug/m3
79-00-5	1,1,2-Trichloroethane	0.28	1.53	UD	1.53	10.9	ug/m3
591-78-6	2-Hexanone	0.32	1.31	UD	1.31	8.18	ug/m3
124-48-1	Dibromochloromethane	0.26	2.22	UD	2.22	17.0	ug/m3
106-93-4	1,2-Dibromoethane	0.29	2.23	UD	2.23	3.07	ug/m3
127-18-4	Tetrachloroethene	0.060	0.41	UD	0.41	0.81	ug/m3
108-90-7	Chlorobenzene	0.30	1.38	UD	1.38	9.21	ug/m3
100-41-4	Ethyl Benzene	0.48	2.08	UD	2.08	8.69	ug/m3
179601-23-1	m/p-Xylene	43.6	189	D	3.65	17.4	ug/m3
1330-20-7	Total Xylenes	78.1	339	D	5.73	26.1	ug/m3
95-47-6	o-Xylene	34.5	150	D	2.08	8.69	ug/m3
100-42-5	Styrene	0.44	1.87	UD	1.87	8.52	ug/m3
75-25-2	Bromoform	0.23	2.38	UD	2.38	20.7	ug/m3
98-82-8	Isopropylbenzene	0.35	1.72	UD	1.72	9.83	ug/m3
79-34-5	1,1,2,2-Tetrachloroethane	0.040	0.27	UDQ	0.27	0.82	ug/m3
95-49-8	2-Chlorotoluene	0.44	2.28	UD	2.28	10.4	ug/m3
108-67-8	1,3,5-Trimethylbenzene	0.44	2.16	UD	2.16	9.83	ug/m3
95-63-6	1,2,4-Trimethylbenzene	0.31	1.52	UD	1.52	9.83	ug/m3
541-73-1	1,3-Dichlorobenzene	0.32	1.92	UD	1.92	12.0	ug/m3
106-46-7	1,4-Dichlorobenzene	0.21	1.26	UD	1.26	12.0	ug/m3
95-50-1	1,2-Dichlorobenzene	27.0	162	D	1.80	12.0	ug/m3
120-82-1	1,2,4-Trichlorobenzene	18.5	137	D	2.52	14.8	ug/m3
87-68-3	Hexachloro-1,3-Butadiene	0.34	3.63	UD	3.63	21.3	ug/m3
106-99-0	1,3-Butadiene	0.52	1.15	UD	1.15	4.42	ug/m3
91-20-3	Naphthalene	0.30	1.57	UDQ	1.57	2.10	ug/m3
622-96-8	4-Ethyltoluene	0.48	2.36	UD	2.36	9.83	ug/m3
110-54-3	Hexane	0.44	1.55	UD	1.55	7.05	ug/m3
107-05-1	Allyl Chloride	0.60	1.88	UD	1.88	6.26	ug/m3
123-91-1	1,4-Dioxane	0.84	3.03	UD	3.03	7.21	ug/m3
100-44-7	Benzyl Chloride	0.72	3.73	UDQ	3.73	10.4	ug/m3
80-62-6	Methyl Methacrylate	0.34	1.39	UD	1.39	8.19	ug/m3

Report of Analysis

Client:	Chemtech Consulting Group	Date Collected:	10/25/24
Project:	NJ Air PT	Date Received:	10/30/24
Client Sample ID:	E070-1100DL2	SDG No.:	P4633
Lab Sample ID:	P4633-01DL2	Matrix:	Air
Analytical Method:	TO-15	Test:	VOCMS Group3
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL041715.D	4		11/21/24 15:06	VL112124

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL	Units
SURROGATES							
460-00-4	1-Bromo-4-Fluorobenzene	10.6			65 - 135	106%	SPK: 10
INTERNAL STANDARDS							
74-97-5	Bromochloromethane	1060000		5.23			
540-36-3	1,4-Difluorobenzene	2920000		6.687			
3114-55-4	Chlorobenzene-d5	2650000		11.50			

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements