

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QC Batch ID # VU111224

Review By	Semsettin Yesilyurt	Review On	11/13/2024 4:19:55 PM
Supervise By	Maresh Dadoda	Supervise On	11/13/2024 4:52:46 PM
SubDirectory	VU111224	HP Acquire Method	MSVOA_U
		HP Processing Method	524U111224DW.M

STD. NAME	STD REF.#
Tune/Reschk	VP131396
Initial Calibration Stds	VP131397,VP131398,VP131399,VP131400,VP131401,VP131402
CCC	VP131404
Internal Standard/PEM	VP129503
ICV/I.BLK	VP131403
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VU061594.D	12 Nov 2024 07:51		MD/SY	Ok
2	VSTDICC0.5	VSTDICC0.5	VU061595.D	12 Nov 2024 08:30	%D fail for Comp. #87,89,90 in 0.5PPB	MD/SY	Ok,M
3	VSTDICC001	VSTDICC001	VU061596.D	12 Nov 2024 09:32		MD/SY	Ok,M
4	VSTDICC002	VSTDICC002	VU061597.D	12 Nov 2024 09:57		MD/SY	Ok,M
5	VSTDICC005	VSTDICC005	VU061598.D	12 Nov 2024 10:21		MD/SY	Ok,M
6	VSTDICCC010	VSTDICCC010	VU061599.D	12 Nov 2024 10:46	Comp. #10,62,66,78,80,81,84,86,87,89, 90 are on Linear Regression	MD/SY	Ok,M
7	VSTDICC015	VSTDICC015	VU061600.D	12 Nov 2024 11:10		MD/SY	Ok,M
8	VSTDICV010	ICVVU111224	VU061601.D	12 Nov 2024 12:20		MD/SY	Ok,M
9	VU1112WBL01	VU1112WBL01	VU061602.D	12 Nov 2024 13:15		MD/SY	Ok
10	P4635-01	PT-THM-WS	VU061603.D	12 Nov 2024 14:44	Need 4X,	MD/SY	Dilution
11	VIBLK	VIBLK	VU061604.D	12 Nov 2024 15:19		MD/SY	Ok
12	P4635-01DL	PT-THM-WSDL	VU061605.D	12 Nov 2024 16:11	Not match	MD/SY	Not Ok
13	P4635-01DL	PT-THM-WSDL	VU061606.D	12 Nov 2024 16:35	Need Further Dilution	MD/SY	Not Ok
14	VU1112WBS01	VU1112WBS01	VU061607.D	12 Nov 2024 17:24		MD/SY	Ok,M
15	VU1112WBSD01	VU1112WBSD01	VU061608.D	12 Nov 2024 17:49		MD/SY	Ok,M
16	P4635-01DL	PT-THM-WSDL	VU061609.D	12 Nov 2024 18:13		MD/SY	Ok
17	P4635-01DL	PT-THM-WSDL	VU061610.D	12 Nov 2024 18:38	Not Required	MD/SY	Not Ok

Instrument ID: MSVOA_U

Daily Analysis Runlog For Sequence/QC Batch ID # VU111224

Review By	Semsettin Yesilyurt	Review On	11/13/2024 4:19:55 PM		
Supervise By	Mahesh Dadoda	Supervise On	11/13/2024 4:52:46 PM		
SubDirectory	VU111224	HP Acquire Method	MSVOA_U	HP Processing Method	524U111224DW.M
STD. NAME	STD REF.#				
Tune/Reschk	VP131396				
Initial Calibration Stds	VP131397,VP131398,VP131399,VP131400,VP131401,VP131402				
CCC	VP131404				
Internal Standard/PEM	VP129503				
ICV/I.BLK	VP131403				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	VSTDCCC010	VSTDCCC010EC	VU061611.D	12 Nov 2024 19:02		MD/SY	Ok,M
----	------------	--------------	------------	-------------------	--	-------	------

M : Manual Integration