

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4635

Client: Chemtech Consulting Group

Analytical Method: SW524.2

Datafile : VU061608.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VU1112WBSD01	Chloroform	2	1.90	ug/L	95	10		70	130	20
	Bromodichloromethane	2	1.90	ug/L	95	10		70	130	20
	Dibromochloromethane	2	1.80	ug/L	90	11		70	130	20
	Bromoform	2	1.80	ug/L	90	11		70	130	20