

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

| | |
|----|---|
| J | Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL). |
| U | Indicates the analyte was analyzed for, but not detected. |
| ND | Indicates the analyte was analyzed for, but not detected |
| E | Indicates the reported value is estimated because of the presence of interference |
| M | Indicates Duplicate injection precision not met. |
| N | Indicates the spiked sample recovery is not within control limits. |
| S | Indicates the reported value was determined by the Method of Standard Addition (MSA). |
| * | Indicates that the duplicate analysis is not within control limits. |
| + | Indicates the correlation coefficient for the MSA is less than 0.995. |
| D | Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range. |
| M | Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed |
| OR | Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis. |
| Q | Indicates the LCS did not meet the control limits requirements |
| H | Sample Analysis Out Of Hold Time |

LAB CHRONICLE

| | |
|---------------------------------|--|
| OrderID: P4638 | OrderDate: 10/30/2024 12:05:00 PM |
| Client: Aramark Uniforms | Project: Monthly 2024 |
| Contact: Jose Liceaga | Location: L11 |

| LabID | ClientID | Matrix | Test | Method | Sample Date | Prep Date | Anal Date | Received |
|----------|----------|--------|------|----------|-------------------|-----------|-------------------|----------|
| P4638-01 | GRAB | WATER | | | 10/30/24 11:24 | | | 10/30/24 |
| | | | TPH | 1664A | | | 10/31/24 09:00 | |
| P4638-02 | COMP | WATER | | | 10/30/24 11:26 | | | 10/30/24 |
| | | | BOD5 | SM5210 B | | | 10/31/24 14:45 | |
| | | | TSS | SM2540 D | | | 11/04/24 09:30 | |



SAMPLE DATA

Report of Analysis

| | | | |
|-------------------|------------------|-----------------|----------------|
| Client: | Aramark Uniforms | Date Collected: | 10/30/24 11:24 |
| Project: | Monthly 2024 | Date Received: | 10/30/24 |
| Client Sample ID: | GRAB | SDG No.: | P4638 |
| Lab Sample ID: | P4638-01 | Matrix: | WATER |
| | | % Solid: | 0 |

| Parameter | Conc. | Qua. | DF | MDL | LOQ / CRQL | Units | Prep Date | Date Ana. | Ana Met. |
|-----------|-------|------|----|------|------------|-------|-----------|----------------|----------|
| TPH | 0.50 | J | 1 | 0.40 | 5.00 | mg/L | | 10/31/24 09:00 | 1664A |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits

Report of Analysis

| | | | |
|-------------------|------------------|-----------------|----------------|
| Client: | Aramark Uniforms | Date Collected: | 10/30/24 11:26 |
| Project: | Monthly 2024 | Date Received: | 10/30/24 |
| Client Sample ID: | COMP | SDG No.: | P4638 |
| Lab Sample ID: | P4638-02 | Matrix: | WATER |
| | | % Solid: | 0 |

| Parameter | Conc. | Qua. | DF | MDL | LOQ / CRQL | Units | Prep Date | Date Ana. | Ana Met. |
|-----------|-------|------|----|------|------------|-------|-----------|----------------|--------------|
| BOD5 | 777 | | 1 | 0.17 | 2.00 | mg/L | | 10/31/24 14:45 | SM 5210 B-16 |
| TSS | 734 | | 1 | 1.00 | 4.00 | mg/L | | 11/04/24 09:30 | SM 2540 D-15 |

Comments: _____

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits



QC RESULT SUMMARY

Preparation Blank Summary

Client: Aramark Uniforms

SDG No.: P4638

Project: Monthly 2024

| Analyte | Units | Result | Acceptance Limits | Conc Qual | MDL | RDL | Analysis Date |
|--------------------|---------------------------|----------|----------------------|--------------|------|-----|------------------|
| Sample ID: TPH | LB133226BL mg/L | < 2.5000 | 2.5000 | U | 0.4 | 5.0 | 10/31/2024 |
| Sample ID: BOD5 | LB133238BL mg/L | < 0.2000 | 0.2000 | U | 0.17 | 2.0 | 10/31/2024 |
| Sample ID: TSS | LB133266BL mg/L | < 2.0000 | 2.0000 | U | 1 | 4 | 11/04/2024 |

Duplicate Sample Summary

| | |
|--|--|
| Client: Aramark Uniforms Project: Monthly 2024 Client ID: LB133226BSD | SDG No.: P4638 Sample ID: LB133226BS Percent Solids for Spike Sample: 0 |
|--|--|

| Analyte | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/ AD | Qual | Analysis Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| TPH | mg/L | +/-18 | 16.8 | | 17.0 | | 1 | 1.18 | | 10/31/2024 |

Duplicate Sample Summary

| | | | |
|-------------------|------------------|---|----------|
| Client: | Aramark Uniforms | SDG No.: | P4638 |
| Project: | Monthly 2024 | Sample ID: | P4638-02 |
| Client ID: | COMPDUP | Percent Solids for Spike Sample: | 0 |

| Analyte | Units | Acceptance Limit | Sample Result | Conc. Qualifier | Duplicate Result | Conc. Qualifier | Dilution Factor | RPD/ AD | Qual | Analysis Date |
|---------|-------|---------------------|------------------|--------------------|---------------------|--------------------|--------------------|------------|------|------------------|
| BOD5 | mg/L | +/-20 | 777 | | 795 | | 1 | 2.32 | | 10/31/2024 |
| TSS | mg/L | +/-5 | 734 | | 740 | | 1 | 0.81 | | 11/04/2024 |

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4638

Project: Monthly 2024

Run No.: LB133226

| Analyte | Units | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB133226BS | | | | | | | |
| TPH | mg/L | 20.0 | 16.8 | | 84 | 1 | 78-114 | 10/31/2024 |

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4638

Project: Monthly 2024

Run No.: LB133226

| Analyte | Units | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|-------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB133226BSD | | | | | | | |
| TPH | mg/L | 20.0 | 17.0 | | 85 | 1 | 78-114 | 10/31/2024 |

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4638

Project: Monthly 2024

Run No.: LB133238

| Analyte | Units | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB133238BS | | | | | | | |
| BOD5 | mg/L | 198 | 199 | | 100 | 1 | 84.6-115.4 | 10/31/2024 |

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: P4638

Project: Monthly 2024

Run No.: LB133266

| Analyte | Units | True Value | Result | Conc. Qualifier | % Recovery | Dilution Factor | Acceptance Limit %R | Analysis Date |
|-----------|------------|------------|--------|-----------------|------------|-----------------|---------------------|---------------|
| Sample ID | LB133266BS | | | | | | | |
| TSS | mg/L | 550 | 538 | | 98 | 1 | 90-110 | 11/04/2024 |



RAW DATA

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB133226
Analysis Date: 10/31/2024
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 10/31/2024
Extraction IN Time: 07:35
Extraction OUT Time: 08:15
Thermometer ID: EXT OVEN#3

| Dish # | Lab ID | Client ID | Matrix | pH | Sample Vol (ml) | Final Volume (ml) | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Silica Gel Weight (g) | Weight After Drying (g) | Final Weight After Drying (g) | Change Weight (g) | Result in ppm |
|--------|-------------|-------------|--------|-----|-----------------|-------------------|-----------------------|-----------------------------|-----------------------|-------------------------|-------------------------------|-------------------|---------------|
| 1 | LB133226BL | LB133226BL | WATER | 1.3 | 1000 | 100 | 2.8563 | 2.8563 | 3.02 | 2.8564 | 2.8564 | 0.0001 | 0.1 |
| 2 | LB133226BS | LB133226BS | WATER | 1.3 | 1000 | 100 | 2.9784 | 2.9784 | 3.03 | 2.9952 | 2.9952 | 0.0168 | 16.8 |
| 3 | LB133226BSD | LB133226BSD | WATER | 1.3 | 1000 | 100 | 3.0256 | 3.0256 | 3.02 | 3.0426 | 3.0426 | 0.0170 | 17 |
| 4 | P4491-01 | GRAB | WATER | 1.6 | 1000 | 100 | 3.0573 | 3.0573 | 3.01 | 3.0976 | 3.0976 | 0.0403 | 40.3 |
| 5 | P4638-01 | GRAB | WATER | 1.3 | 1000 | 100 | 3.1026 | 3.1026 | 3.03 | 3.1031 | 3.1031 | 0.0005 | 0.5 |
| 6 | P4641-01 | 402 | WATER | 1.6 | 1000 | 100 | 3.1060 | 3.1060 | 3.04 | 3.1457 | 3.1457 | 0.0397 | 39.7 |



QC Batch# LB133226

Test: TPH

Analysis Date: 10/31/2024

Chemicals Used:

| Chemical Name | Chemical Lot # |
|----------------|----------------|
| HEXANE | W3110 |
| pH Paper 0-14 | M4909 |
| Sodium Sulfate | EP2554 |
| 1:1 HCL | WP108566 |
| Silica Gel | W3079 |
| Sand | NA |

Standards Used:

| Standard Name | Amount Used | Standard Lot # |
|---------------|-------------|----------------|
| LCSW | 5.00 ML | WP108567 |
| LCSWD | 5.00 ML | WP108568 |
| MS/MSD | NA | NA |

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:26

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:00

Bal Check Time: 07:40 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:00

Out Time1: 10:25

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:01

1.0000 gram Balance: 1.0003 (0.9950-1.0050) In Time2: 11:30

Bal Check Time: 12:36 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 12:35

Out Time2: 12:00

WORKLIST(Hardcopy Internal Chain)

✓ 133226

WorkList Name : tph p4491

WorkList ID : 184962

Department : Wet-Chemistry

Date : 10-31-2024 07:25:02

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|----------------------|----------|-----------------------------|--------------|--------|
| P4491-01 | GRAB | Water | TPH | Conc H2SO4 to pH < 2 | ARAM01 | K11 | 10/23/2024 | 1664A |
| P4638-01 | GRAB | Water | TPH | Conc H2SO4 to pH < 2 | ARAM01 | L11 | 10/30/2024 | 1664A |
| P4641-01 | 402 | Water | TPH | Conc H2SO4 to pH < 2 | PSEG04 | K51 | 10/30/2024 | 1664A |

Date/Time 10/31/24 07:30

Raw Sample Received by: SP Cee

Raw Sample Relinquished by: afm

Date/Time 10/31/24

Raw Sample Received by: afm

Raw Sample Relinquished by: afm

BOD5 LOG

ANALYST: rubin
Inst Id :DO METER
LB :LB133238

SUPERVISOR: Iwona

Analysis Date: 10/31/2024

MANGANOUS SULFATE SOLUTION: W3103

Alkaline Iodide Azide: W3109

Sodium Thiosulfate, 0.025N: W3105

NaOH, 1N: WP108662

IncubatorID: INCUBATOR #3

GuageID: 0511062

Zero DO: WP110006

QC BATCH ID: LB133238

BOD Water: WP110512

Starch: W3149

Sulfuric acid, 1N: WP110386

POLYSEED: WP110514

GGA: WP110513

Chlorine Strips: W2965

pH Strips: W3104

| Lab SampleID | Client ID | Bottle No. | VOL. ML | Initial Reading (ML) | Final Reading (ML) | Difference | Average |
|--------------|-----------|------------|---------|----------------------|--------------------|------------|---------|
| WINKLER 1 | WINKLER 1 | 1 | 300 | 0.0 | 9.4 | 9.4 | 9.4 |
| WINKLER 2 | WINKLER 2 | 2 | 300 | 9.6 | 19.00 | 9.4 | 9.4 |

Meter Calibration1: 8.79 Zero DO Reading1: 0.10 mg/L (<=0.2 Criteria)

Barometric Pressure1: 771 mmHg DO Meter BOD fluid reading for winkler comparison: 9.47

After Incubation

Meter Calibration2: 8.72 Zero DO Reading2: 0.14 mg/L (<=0.2 Criteria)

Barometric Pressure2: 771 mmHg

QC BATCH ID: LB133238

INCUBATOR TEMP IN(C): 20.5

INCUBATOR TEMP OUT(C): 20.3

TIME IN: 14:45

TIME OUT: 10:15

DATE IN: 10/31/2024

DATE OUT: 11/05/2024

| Lab SampleID | Bottle No. | Check CL | Initial PH | Final PH | Temp °C | Sam Vol. (mL) | D.O.1 Initial | D.O.2 Final | Depletion | BOD Result (mg/L) | Avg Result (mg/L) | Comment |
|--------------|------------|----------|------------|----------|---------|---------------|---------------|-------------|-----------|-------------------|-------------------|---------|
| LB133238BL | 1 | No | 6.72 | N/A | 20.70 | 300 | 9.46 | 9.45 | 0.01 | 0.01 | 0.01 | |
| POLYSEED | 1 | | | | | 10 | 9.45 | 7.25 | 2.2 | 0.44 | 0.51 | |
| POLYSEED | 2 | | | | | 15 | 9.41 | 5.38 | 4.03 | 0.54 | | |
| POLYSEED | 3 | | | | | 20 | 9.33 | 3.91 | 5.42 | 0.54 | | |
| GGA | 1 | | | | | 6 | 9.40 | 5.09 | 4.31 | 190 | 198.5 | |
| GGA | 2 | | | | | 6 | 9.36 | 4.89 | 4.47 | 198 | | |
| GGA | 3 | | | | | 6 | 9.35 | 4.69 | 4.66 | 207.5 | | |
| P4638-02 | 1 | No | 6.83 | N/A | 20.00 | 0.5 | 9.44 | 7.40 | 2.04 | 918 | 776.5 | |
| P4638-02 | 2 | | | | | 1 | 9.37 | 5.75 | 3.62 | 933 | | |
| P4638-02 | 3 | | | | | 2 | 9.34 | 3.91 | 5.43 | 738 | | |
| P4638-02 | 4 | | | | | 3 | 9.31 | 3.63 | 5.68 | 517 | | |
| P4638-02DUP | 1 | No | 6.83 | N/A | 20.00 | 0.5 | 9.42 | 7.28 | 2.14 | 978 | 794.75 | |
| P4638-02DUP | 2 | | | | | 1 | 9.36 | 5.72 | 3.64 | 939 | | |
| P4638-02DUP | 3 | | | | | 2 | 9.34 | 3.87 | 5.47 | 744 | | |
| P4638-02DUP | 4 | | | | | 3 | 9.29 | 3.60 | 5.69 | 518 | | |

NOTE: 2ml POLYSEED added to GGA and all the Samples, but not in Blank.

NOTE (For, CBOD5): 0.16 g Nitrification Inhibitor added to GGA and all the Samples, but not in Blank.

WORKLIST(Hardcopy Internal Chain)

66133238

WorkList Name : bod5-1031

WorkList ID : 184978

Department : Wet-Chemistry

Date : 10-31-2024 10:15:18

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|----------|
| P4638-02 | COMP | Water | BOD5 | Cool 4 deg C | ARAM01 | L11 | 10/30/2024 | SM5210 B |

Date/Time 10/31/2024 13:00
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

Date/Time 10/31/2024 14:30
Raw Sample Received by: RM CWC
Raw Sample Relinquished by: RM CWC

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: Niha

Date: 11/01/2024

Run Number: LB133266

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 11/01/2024 14:00 TEMP1 OUT: 104 °C 11/01/2024 15:00
 TEMP2 IN: 103 °C 11/01/2024 15:30 TEMP2 OUT: 104 °C 11/01/2024 16:30
 TEMP3 IN: 104 °C 11/04/2024 09:30 TEMP3 OUT: 103 °C 11/04/2024 11:00
 TEMP4 IN: 104 °C 11/04/2024 11:30 TEMP4 OUT: 103 °C 11/04/2024 13:00

| Dish # | Lab ID | Client ID | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Sample Volume (ml) | 1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | 2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Weight (g) | Result mg/L |
|--------|-------------|------------|-----------------------|-----------------------------|--------------------|---|---|---|------------|-------------|
| 1 | LB133266BL | LB133266BL | 1.4012 | 1.4012 | 100 | 1.4012 | 1.4012 | 1.4012 | 0.0000 | 0 |
| 2 | LB133266BS | LB133266BS | 1.4132 | 1.4132 | 100 | 1.4670 | 1.4670 | 1.4670 | 0.0538 | 538 |
| 3 | P4638-02 | COMP | 1.3988 | 1.3988 | 50 | 1.4355 | 1.4355 | 1.4355 | 0.0367 | 734 |
| 4 | P4638-02DUP | COMPDUP | 1.4021 | 1.4021 | 50 | 1.4391 | 1.4391 | 1.4391 | 0.0370 | 740 |
| 5 | P4649-01 | MC0JZ7 | 1.3994 | 1.3994 | 1000 | 1.6624 | 1.6624 | 1.6624 | 0.2630 | 263 |
| 6 | P4649-02 | MC0JZ9 | 1.3908 | 1.3908 | 1000 | 1.3938 | 1.3938 | 1.3938 | 0.0030 | 3 |
| 7 | P4649-03 | MC0K01 | 1.4168 | 1.4168 | 1000 | 1.4184 | 1.4184 | 1.4184 | 0.0016 | 1.6 |
| 8 | P4649-04 | MC0K04 | 1.4040 | 1.4040 | 1000 | 1.4046 | 1.4046 | 1.4046 | 0.0006 | 0.6 |
| 9 | P4649-05 | MC0K06 | 1.4134 | 1.4134 | 1000 | 1.4138 | 1.4138 | 1.4138 | 0.0004 | 0.4 |
| 10 | P4649-06 | MC0K08 | 1.4151 | 1.4151 | 1000 | 1.4155 | 1.4155 | 1.4155 | 0.0004 | 0.4 |
| 11 | P4649-07 | MC0K10 | 1.4021 | 1.4021 | 1000 | 1.4032 | 1.4032 | 1.4032 | 0.0011 | 1.1 |
| 12 | P4649-08 | MC0K12 | 1.3961 | 1.3961 | 1000 | 1.3976 | 1.3976 | 1.3976 | 0.0015 | 1.5 |
| 13 | P4649-09 | MC0K14 | 1.4068 | 1.4068 | 1000 | 1.4074 | 1.4074 | 1.4074 | 0.0006 | 0.6 |
| 14 | P4649-10 | MC0K16 | 1.4059 | 1.4059 | 1000 | 1.4065 | 1.4065 | 1.4065 | 0.0006 | 0.6 |
| 15 | P4649-11 | MC0K18 | 1.3990 | 1.3990 | 1000 | 1.4111 | 1.4111 | 1.4111 | 0.0121 | 12.1 |

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: Iwona

ANALYST: Niha

Date: 11/01/2024

Run Number: LB133266

BalanceID: WC SC-6

OvenID: WC OVEN-1

FilterID: 17416528

ThermometerID: WET OVEN#1

TEMP1 IN: 103 °C 11/01/2024 14:00 TEMP1 OUT: 104 °C 11/01/2024 15:00
 TEMP2 IN: 103 °C 11/01/2024 15:30 TEMP2 OUT: 104 °C 11/01/2024 16:30
 TEMP3 IN: 104 °C 11/04/2024 09:30 TEMP3 OUT: 103 °C 11/04/2024 11:00
 TEMP4 IN: 104 °C 11/04/2024 11:30 TEMP4 OUT: 103 °C 11/04/2024 13:00

| Dish # | Lab ID | Client ID | Empty Dish Weight (g) | Final Empty Dish Weight (g) | Sample Volume (ml) | 1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | 2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g) | Weight (g) | Result mg/L |
|--------|----------|-------------|-----------------------|-----------------------------|--------------------|---|---|---|------------|-------------|
| 16 | P4649-12 | MC0K20 | 1.4052 | 1.4052 | 1000 | 1.4055 | 1.4055 | 1.4055 | 0.0003 | 0.3 |
| 17 | P4649-13 | MC0K22 | 1.3714 | 1.3714 | 1000 | 1.3772 | 1.3772 | 1.3772 | 0.0058 | 5.8 |
| 18 | P4674-01 | OUTFALL-001 | 1.4101 | 1.4101 | 750 | 1.4905 | 1.4905 | 1.4905 | 0.0804 | 107.2 |

A = Sample Volume (ml)
 B = Final Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

LB133266

WorkList Name : TSS-10312024 WorkList ID : 184987 Department : Wet-Chemistry Date : 10-31-2024 11:45:53

| Sample | Customer Sample | Matrix | Test | Preservative | Customer | Raw Sample Storage Location | Collect Date | Method |
|----------|-----------------|--------|------|--------------|----------|-----------------------------|--------------|----------|
| P4674-01 | OUTFALL-001 | Water | TSS | Cool 4 deg C | RSBE01 | K51 | 10/31/2024 | SM2540 D |
| P4649-01 | MC0JZ7 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-02 | MC0JZ9 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-03 | MC0K01 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-04 | MC0K04 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-05 | MC0K06 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-06 | MC0K08 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-13 | MC0K22 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-07 | MC0K10 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-08 | MC0K12 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-09 | MC0K14 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-10 | MC0K16 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-11 | MC0K18 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4649-12 | MC0K20 | Water | TSS | Cool 4 deg C | TETR16 | K61 | 10/29/2024 | SM2540 D |
| P4638-02 | COMP | Water | TSS | Cool 4 deg C | ARAM01 | L11 | 10/30/2024 | SM2540 D |

Date/Time 11.04.2024, 09:00
 Raw Sample Received by: NF(wc)
 Raw Sample Relinquished by: CP

Date/Time 11.04.2024, 11:00
 Raw Sample Received by: CP
 Raw Sample Relinquished by: NF(wc)

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB133226

| | | | |
|------------------|---|--------------|-----------------------|
| Review By | jignesh | Review On | 10/31/2024 9:39:17 AM |
| Supervise By | Iwona | Supervise On | 10/31/2024 9:56:19 AM |
| SubDirectory | LB133226 | Test | TPH |
| STD. NAME | STD REF.# | | |
| ICAL Standard | N/A | | |
| ICV Standard | N/A | | |
| CCV Standard | N/A | | |
| ICSA Standard | N/A | | |
| CRI Standard | N/A | | |
| LCS Standard | N/A | | |
| Chk Standard | W3110,M4909,EP2554,WP108566,W3079,NA,WP108567,WP108568,NA | | |

| Sr# | SampleId | ClientID | QcType | Date | Comment | Operator | Status |
|-----|-------------|-------------|--------|----------------|---------|----------|--------|
| 1 | LB133226BL | LB133226BL | MB | 10/31/24 09:00 | | jignesh | OK |
| 2 | LB133226BS | LB133226BS | LCS | 10/31/24 09:00 | | jignesh | OK |
| 3 | LB133226BSD | LB133226BSD | LCSD | 10/31/24 09:00 | | jignesh | OK |
| 4 | P4491-01 | GRAB | SAM | 10/31/24 09:00 | | jignesh | OK |
| 5 | P4638-01 | GRAB | SAM | 10/31/24 09:00 | | jignesh | OK |
| 6 | P4641-01 | 402 | SAM | 10/31/24 09:00 | | jignesh | OK |

Instrument ID: DO METER

Daily Analysis Runlog For Sequence/QC Batch ID # LB133238

| | | | |
|------------------|--|--------------|----------------------|
| Review By | rubina | Review On | 11/5/2024 2:38:02 PM |
| Supervise By | Iwona | Supervise On | 11/5/2024 3:05:20 PM |
| SubDirectory | LB133238 | Test | BOD5 |
| STD. NAME | STD REF.# | | |
| ICAL Standard | N/A | | |
| ICV Standard | N/A | | |
| CCV Standard | N/A | | |
| ICSA Standard | N/A | | |
| CRI Standard | N/A | | |
| LCS Standard | N/A | | |
| Chk Standard | WP110512,W3149,WP110386,W3103,W3109,W3105,WP110514,WP110513,WP108662 | | |

| Sr# | SampleId | ClientID | QcType | Date | Comment | Operator | Status |
|-----|-------------|------------|--------|----------------|-----------------------|----------|--------|
| 1 | LB133238BL | LB133238BL | MB | 10/31/24 14:45 | | rubina | OK |
| 2 | LB133238BS | LB133238BS | LCS | 10/31/24 14:45 | | rubina | OK |
| 3 | P4638-02 | COMP | SAM | 10/31/24 14:45 | Intermediate dilution | rubina | OK |
| 4 | P4638-02DUP | COMPDUP | DUP | 10/31/24 14:45 | Intermediate dilution | rubina | OK |

Instrument ID: WC SC-3

Daily Analysis Runlog For Sequence/QC Batch ID # LB133266

| | | | |
|------------------|------------------|--------------|----------------------|
| Review By | Niha | Review On | 11/4/2024 2:44:16 PM |
| Supervise By | Iwona | Supervise On | 11/4/2024 4:46:27 PM |
| SubDirectory | LB133266 | Test | TSS |
| STD. NAME | STD REF.# | | |
| ICAL Standard | N/A | | |
| ICV Standard | N/A | | |
| CCV Standard | N/A | | |
| ICSA Standard | N/A | | |
| CRI Standard | N/A | | |
| LCS Standard | N/A | | |
| Chk Standard | N/A | | |

| Sr# | SampleId | ClientID | QcType | Date | Comment | Operator | Status |
|-----|-------------|-------------|--------|----------------|---------|----------|--------|
| 1 | LB133266BL | LB133266BL | MB | 11/04/24 09:30 | | Niha | OK |
| 2 | LB133266BS | LB133266BS | LCS | 11/04/24 09:30 | | Niha | OK |
| 3 | P4638-02 | COMP | SAM | 11/04/24 09:30 | | Niha | OK |
| 4 | P4638-02DUP | COMPDUP | DUP | 11/04/24 09:30 | | Niha | OK |
| 5 | P4649-01 | MC0JZ7 | SAM | 11/04/24 09:30 | | Niha | OK |
| 6 | P4649-02 | MC0JZ9 | SAM | 11/04/24 09:30 | | Niha | OK |
| 7 | P4649-03 | MC0K01 | SAM | 11/04/24 09:30 | | Niha | OK |
| 8 | P4649-04 | MC0K04 | SAM | 11/04/24 09:30 | | Niha | OK |
| 9 | P4649-05 | MC0K06 | SAM | 11/04/24 09:30 | | Niha | OK |
| 10 | P4649-06 | MC0K08 | SAM | 11/04/24 09:30 | | Niha | OK |
| 11 | P4649-07 | MC0K10 | SAM | 11/04/24 09:30 | | Niha | OK |
| 12 | P4649-08 | MC0K12 | SAM | 11/04/24 09:30 | | Niha | OK |
| 13 | P4649-09 | MC0K14 | SAM | 11/04/24 09:30 | | Niha | OK |
| 14 | P4649-10 | MC0K16 | SAM | 11/04/24 09:30 | | Niha | OK |
| 15 | P4649-11 | MC0K18 | SAM | 11/04/24 09:30 | | Niha | OK |
| 16 | P4649-12 | MC0K20 | SAM | 11/04/24 09:30 | | Niha | OK |
| 17 | P4649-13 | MC0K22 | SAM | 11/04/24 09:30 | | Niha | OK |
| 18 | P4674-01 | OUTFALL-001 | SAM | 11/04/24 09:30 | | Niha | OK |

Prep Standard - Chemical Standard Summary

Order ID : P4638

Test : BOD5,TPH,TSS

Prepbatch ID :

Sequence ID/Qc Batch ID: LB133226, LB133238, LB133266,

Standard ID :

EP2554, WP108566, WP108567, WP108568, WP108662, WP110386, WP110512, WP110513, WP110514,

Chemical ID :

E3551, E3726, M4909, M5673, M5943, W2606, W2653, W2654, W2817, W2871, W3009, W3030, W3079, W3082, W3103, W3105, W3109, W3110, W3112, W3113, W3117, W3149,

Extractions STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|----------------------|------------------------|------------------|------------------------|--------------------|-------------------------------------|------------------|---------------------------------|
| 3923 | Baked Sodium Sulfate | EP2554 | 10/26/2024 | 01/03/2025 | RUPESHKUMAR SHAH | Extraction_SC ALE_2 (EX-SC-2) | None | Rajesh Parikh 10/26/2024 |

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|--------------------------------|
| 229 | 1:1 HCL | WP108566 | 06/27/2024 | 10/24/2024 | Jignesh Parikh | None | None | Iwona Zarych 06/27/2024 |

FROM 500.00000ml of M5943 + 500.00000ml of W2606 = Final Quantity: 1.000 L

Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|--------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------|
| 2470 | 1664A SPIKING SOLN | WP108567 | 06/27/2024 | 12/25/2024 | Jignesh Parikh | None | None | Iwona Zarych |
| | | | | | | | | 06/27/2024 |

FROM 1000.00000ml of E3726 + 4.00000gram of W2817 + 4.00000gram of W2871 = Final Quantity: 1000.000 ml

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|----------------------|
| 3374 | 1664A QCS spiking solution-SS | WP108568 | 06/27/2024 | 12/25/2024 | Jignesh Parikh | WETCHEM_S CALE_4 (WC SC-4) | None | Iwona Zarych |
| | | | | | | | | 06/27/2024 |

FROM 1000.00000ml of E3726 + 4.00000gram of W3009 + 4.00000gram of W3082 = Final Quantity: 1000.000 ml



| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|--|----------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|--------------------------------|
| 1571 | Sodium hydroxide, 1N | WP108662 | 07/09/2024 | 01/09/2025 | Rubina Mughal | WETCHEM_S CALE_5 (WC SC-5) | None | Iwona Zarych 07/11/2024 |
| <u>FROM</u> 4.00000gram of W3113 + 96.00000ml of W3112 = Final Quantity: 100.000 ml | | | | | | | | |

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|--|-------------------|--------------------------|------------------|------------------------|--------------------|----------------|---------------------------|----------------------------|
| 1841 | Sulfuric Acid, 1N | WP110386 | 10/24/2024 | 04/24/2025 | Rubina Mughal | None | WETCHEM_FIPETTE_3 (WC) | Iwona Zarych 10/24/2024 |
| <u>FROM</u> 2.80000ml of M5673 + 97.20000ml of W3112 = Final Quantity: 100.000 ml | | | | | | | | |

Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|--------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------|
| 127 | BOD Dilution fluid | WP110512 | 10/31/2024 | 11/01/2024 | Rubina Mughal | None | None | Iwona Zarych |
| | | | | | | | | 11/04/2024 |

FROM 18.00000L of W3112 + 3.00000PILLOW of W3117 = Final Quantity: 18.000 L

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|------------------|-----------------------------------|--------------------------|------------------|------------------------|--------------------|----------------------------------|------------------|----------------------|
| 129 | Glutamic acid-glucose mix for BOD | WP110513 | 10/31/2024 | 11/01/2024 | Rubina Mughal | WETCHEM_S CALE_7 (WC SC-6) | None | Iwona Zarych |
| | | | | | | | | 11/04/2024 |

FROM 0.15000gram of W2653 + 0.15000gram of W2654 + 1000.00000ml of W3112 = Final Quantity: 1000.000 ml

Wet Chemistry STANDARD PREPARATION LOG

| <u>Recipe ID</u> | <u>NAME</u> | <u>NO.</u> | <u>Prep Date</u> | <u>Expiration Date</u> | <u>Prepared By</u> | <u>ScaleID</u> | <u>PipetteID</u> | <u>Supervised By</u> |
|--|-----------------------|--------------------------|------------------|------------------------|--------------------|----------------|------------------|----------------------------|
| 128 | polyseed seed control | WP110514 | 10/31/2024 | 11/01/2024 | Rubina Mughal | None | None | Iwona Zarych 11/04/2024 |
| <p><u>FROM</u> 1.00000PILLOW of W3030 + 300.00000ml of WP110512 = Final Quantity: 300.000 ml</p> | | | | | | | | |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1 | 313201 | 01/03/2025 | 01/03/2024 / Rajesh | 07/20/2023 / Rajesh | E3551 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|--|-------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9254-03 / Acetone, Ultra Resi (cs/4x4L) | 1234 | 12/25/2024 | 02/26/2024 / Rajesh | 02/23/2024 / Rajesh | E3726 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 140440 / TEST PAPERS,PH,0-2.5,.2SENSI, 100PK | HC908519 | 08/31/2024 | 11/28/2022 / jaswal | 08/09/2021 / jaswal | M4909 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L) | 23D2462010 | 03/20/2028 | 09/21/2023 / mohan | 09/05/2023 / mohan | M5673 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9530-33 / Hydrochloric Acid, Instra-Analyzed (cs/6x2.5L) | 22G2862015 | 12/24/2024 | 06/24/2024 / Al-Terek | 06/21/2024 / Al-Terek | M5943 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------|---------------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | DIW / DI Water | Daily Lab-Certified | 10/24/2024 | 10/24/2019 / apatel | 10/24/2019 / apatel | W2606 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AC156212500 / GLUTAMIC ACID BIOCHEM REG, 250G | A0405990 | 01/24/2030 | 01/24/2020 / apatel | 01/24/2020 / apatel | W2653 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | D16-500 / DEXTROSE ANHYDROUS ACS REAGENT, 500G(New) | 186122A | 01/24/2030 | 01/24/2020 / apatel | 01/24/2020 / apatel | W2654 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | A12244 / Stearic acid, 98%, 100 g | U20E006 | 04/02/2026 | 04/02/2021 / apatel | 04/02/2021 / apatel | W2817 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-----------------------------|------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | H223-57 / Hexadecane, 99.0% | 0000266903 | 05/04/2027 | 09/07/2021 / apatel | 08/26/2021 / apatel | W2871 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|-----------------------------|----------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | H223-57 / Hexadecane, 99.0% | SHBP8192 | 02/27/2028 | 02/27/2023 / lwona | 02/27/2023 / lwona | W3009 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|----------------------|--------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 136742-80 / POLYSEED | 282211 | 11/30/2024 | 10/30/2024 / lwona | 05/10/2023 / lwona | W3030 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|-----------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 04667-2.5 / Silica Gel (60-200 mesh), 2.5 KG | 072154301 | 01/30/2029 | 05/07/2024 / jignesh | 01/30/2024 / jignesh | W3079 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|-----------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | A12244 / Stearic acid, 98%, 100 g | U23E020 | 02/26/2029 | 02/26/2024 / lwona | 02/26/2024 / lwona | W3082 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | 4620-32 / MANGANOUS SULFATE SOLUTION-364 | 2403J02 | 03/31/2026 | 04/22/2024 / lwona | 04/22/2024 / lwona | W3103 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL69870-8 / SODIUM THIOSULFATE,0.025N,4LIT RE | 4403S13 | 09/30/2025 | 04/22/2024 / lwona | 04/22/2024 / lwona | W3105 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL04100-4 / Alkaline Iodide Azide, 1 L | 1405D67 | 04/30/2026 | 05/23/2024 / lwona | 05/23/2024 / lwona | W3109 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---|--------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L) | 235898 | 02/28/2029 | 06/27/2024 / jignesh | 06/26/2024 / jignesh | W3110 |

CHEMICAL RECEIPT LOG BOOK

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|------------------|---------------------|---------------------|-----------------|-------------------------|-----------------------------|----------------|
| Seidler Chemical | DIW / DI Water | Daily Lab-Certified | 07/03/2029 | 07/03/2024 / lwona | 07/03/2024 / lwona | W3112 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|--|------------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | PC19510-7 / Sodium Hydroxide Pellets 12 Kg | 23B1556310 | 12/31/2025 | 07/08/2024 / lwona | 07/08/2024 / lwona | W3113 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|----------|--|-------|-----------------|-------------------------|-----------------------------|----------------|
| HACH | 1486266 / BOD Nutrient Buffer Pillows, 6 mL concentrate to make 6 L, 50/pk | A4032 | 04/30/2029 | 10/02/2024 / rubina | 07/12/2024 / lwona | W3117 |

| Supplier | ItemCode / ItemName | Lot # | Expiration Date | Date Opened / Opened By | Received Date / Received By | Chemtech Lot # |
|-----------------------------|---------------------------------|---------|-----------------|-------------------------|-----------------------------|----------------|
| PCI Scientific Supply, Inc. | AL70850-8 / Starch Solution, 4L | 4408P62 | 08/31/2026 | 10/16/2024 / lwona | 10/16/2024 / lwona | W3149 |

Hexadecane, 99.0%



Material No.: H223-57
Batch No.: 0000266903
Manufactured Date: 2020/05/05
Retest Date: 2027/05/04
Revision No: 1

Certificate of Analysis

| Test | Specification | Result |
|--|----------------|--------|
| Assay ($\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$) (by GC) | $\geq 99.0 \%$ | 99.3 |
| Infrared Spectrum | Passes Test | PT |

For Laboratory, Research or Manufacturing Use

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC



Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



ACROS ORGANICS part of Thermo Fisher Scientific





Version 0

Molecular weight 147.13

Molecular formula C5 H9 N O4

CAS No 56-86-0

Linear formula HO2CCH2CH2CH(NH2)CO2H

Flash point (°C)


Certificate of Analysis

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Acros Organics expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to human or animals. It is the responsibility of the purchaser, formulator or those performing further manufacturing to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

| | | | |
|------------------------------|-------------------------|------------------------------------|---------------|
| Catalog Number | 15621 | Quality Test / Release Date | 13 March 2019 |
| Lot Number | A0405990 | Suggested Retest Date | March 2022 |
| Description | L(+)-Glutamic acid, 99% | | |
| Country of Origin | CHINA | | |
| Declaration of Origin | plant | | |

| | |
|-----------------------|---|
| Origin Comment | The product is made by fermentation of sugar molasses |
|-----------------------|---|

| Result Name | Specifications | Test Value |
|---------------------------|--|--|
| Appearance (Color) | White | White |
| Appearance (Form) | Powder | Powder |
| Infrared spectrum | Conforms | Conforms |
| Titration with NaOH | 98.5 to 100.5 % (On dried substance) | 99.32 % (On dried substance) |
| Loss on drying | ≤0.5 % (105°C, 3 hrs) | 0.002 % (105°C, 3 hrs) |
| Heavy metals (as Pb) | ≤10 ppm | ≤10 ppm |
| Sulfated ash | ≤0.1 % | 0.08 % |
| Other amino acids | not detectable | not detectable |
| Specific optical rotation | +30.5° to +32.5° (20°C, 589 nm) (on dried substance) | +32° (20°C, 589 nm) (on dried substance) |
| Specific optical rotation | (c=10, 2N HCl) | (c=10, 2N HCl) |
| Chloride (Cl) | ≤200 ppm | ≤200 ppm |
| Iron (Fe) | ≤30 ppm | ≤10 ppm |
| Sulfate (SO4) | ≤300 ppm | ≤200 ppm |
| Ammonium (NH4) | ≤200 ppm | ≤200 ppm |
| Arsenic oxide (As2O3) | ≤1 ppm | ≤1 ppm |



A handwritten signature in black ink, which appears to read "L. Van den Broek".

L. Van den Broek, QA Manager

Issued: 24 January 2020

Acros Organics

ENA23, zone 1, nr 1350, Janssen Pharmaceuticaaan 3a, B-2440 Geel, Belgium

Tel +32 14/57.52.11 - Fax +32 14/59.34.34 Internet: <http://www.acros.com>

1 Reagent Lane, Fair Lawn, NJ 07410, USA Fax 201-796-1329

W2817

REC. 04/02/2021

Product Name: Stearic acid, 98%, Thermo Scientific Chemicals
Catalog Number: A12244.14

CAS Number: 57-11-4
Molecular Formula: C₁₈H₃₆O₂
Molecular Weight: 284.48
InChI Key: QIQXTHQIDYTRH-UHFFFAOYSA-N
SMILES: CCCCCCCCCCCCCCCC(O)=O
Synonym: stearic acid acide stearique hydrofol acid 1855 hydrofol acid 1655 industrene 5016
stearic acid, ion(1-) (8Cl) glycon TP glycon DP acidum stearinicum hydrofol acid 150

Product Specification

Appearance (Color): White
Form: Crystals or powder or crystalline powder or flakes or waxy solid
Assay (Silylated GC): ≥97.5%
Melting Point (clear melt): 67.0-74.0°C

Date Of Print: 11/30/2023

Product Specifications are subject to amendment and may change over time. Data contained is accurate as of the date printed.

W3009
rec. 2/27/2023 12

Product Name:

Hexadecane - ReagentPlus®, 99%

Certificate of Analysis

Product Number:

H6703

Batch Number:

SHBP8192

 $\text{CH}_3(\text{CH}_2)_{14}\text{CH}_3$

Brand:

SIAL

CAS Number:

544-76-3

MDL Number:

MFCD00008998

Formula:

C16H34

Formula Weight:

226.44 g/mol

Quality Release Date:

04 AUG 2022

| Test | Specification | Result |
|----------------------------|-----------------------|-----------|
| Appearance (Color) | Colorless or White | Colorless |
| Appearance (Form) | Liquid or Solid | Liquid |
| Infrared Spectrum | Conforms to Structure | Conforms |
| Refractive index at 20 ° C | 1.432 - 1.436 | 1.435 |
| Purity (GC) | ≥ 98.5 % | 99.3 % |
| Color Test | ≤ 20 APHA | < 5 APHA |


Larry Coers, Director

Quality Control

Sheboygan Falls, WI US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.



W 3029

W 3030

Rec. 05/11/23

12



CERTIFICATE OF ANALYSIS

PO BOX 130549 Spring, TX 77393
Phone: (281) 298-9410 Fax: (281) 298-9411

FINISHED PRODUCT, LOT NUMBER, MFG. /EXP DATE:

PolySeed® • Part No. P-110 • Lot 282211 • Mfg. Date: 11/2022 • Exp. Date: 11/2024

FORMULATION:

The formulation for this product contains a range of naturally occurring microorganisms, which are known to be non-pathogenic to man or animals.

VIABLE COUNT, FINAL TEST RESULT:

The product has been fully tested in accordance with Finished Product Specifications and contains a minimum viable count of 4.00×10^9 cfu/g.

GLUCOSE/GLUTAMIC-ACID RESULTS:

Tested results within acceptable range 198 ± 30.5 mg/L (167.5 - 228.5 mg/L). GGA Lot# J317-19 – Average Test Result: 205.3

See www.polyseed.com for details.

SEED CONTROL FACTOR:

Tested results within acceptable range 0.6 – 1.0 see www.polyseed.com for details

SALMONELLA TEST RESULT:

The product has been shown to be Salmonella negative using procedures recommended in the Microbiology Laboratory Guidebook, published by the USDA Food Safety and Inspection Service.

The purpose of this document is to assure that the Finished Product conforms to the above specification.

Signature: _____

Quality Control Department

Date: 11/28/2022

POLYSEED.Ref.1.19

Revised Jan 22



Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

| | | | |
|-------------------|---|-----------------------------|------------|
| Catalog Number | D16 | Quality Test / Release Date | 03/19/2019 |
| Lot Number | 186122A | | |
| Description | DEXTROSE, ANHYDROUS, A.C.S. | | |
| Country of Origin | United States | Suggested Retest Date | Mar/2022 |
| Chemical Origin | Organic - Plant | | |
| BSE/TSE Comment | No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product. | | |
| Chemical Comment | | | |

| N/A | | | |
|--------------------------|------------------|---------------------------------|------------------------|
| Result Name | Units | Specifications | Test Value |
| APPEARANCE | | REPORT | White, granular powder |
| TITRATABLE ACID | MEQ/G | <= 0.002 | <0.002 |
| STARCH | | = PASS TEST | pass test |
| SPECIFIC ROTATION @ 25 C | DEGREES (+ OR -) | Inclusive Between +52.5 - +53.0 | 53.0 |
| SULFATE & SULFITE | % | <= 0.005 | <0.005 |
| IRON (Fe) | ppm | <= 5 | <5 |
| CHLORIDE | % | <= 0.01 | <0.01 |
| IGNITION RESIDUE | % | <= 0.02 | <0.02 |
| IDENTIFICATION | PASS/FAIL | = PASS TEST | pass test |
| HEAVY METALS (as Pb) | ppm | <= 5 | <5 |
| LOSS ON DRYING @ 105 C | % | <= 0.2 | <0.2 |
| INSOLUBLE MATTER | % | <= 0.005 | 0.002 |

Jerisa Bailey-Wyche

Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

| | | | |
|------------------------|-----------------------------------|---------------|---------------------------------|
| PRODUCT : | SODIUM SULFATE CRYSTALS ANHYDROUS | | |
| QUALITY : | ACS (CODE RMB3375) | FORMULA : | Na ₂ SO ₄ |
| SPECIFICATION NUMBER : | 6399 | RELEASE DATE: | ABR/21/2023 |
| LOT NUMBER : | 313201 | | |

| TEST | SPECIFICATIONS | LOT VALUES |
|--|----------------|-------------|
| Assay (Na ₂ SO ₄) | Min. 99.0% | 99.7 % |
| pH of a 5% solution at 25°C | 5.2 - 9.2 | 6.1 |
| Insoluble matter | Max. 0.01% | 0.005 % |
| Loss on ignition | Max. 0.5% | 0.1 % |
| Chloride (Cl) | Max. 0.001% | <0.001 % |
| Nitrogen compounds (as N) | Max. 5 ppm | <5 ppm |
| Phosphate (PO ₄) | Max. 0.001% | <0.001 % |
| Heavy metals (as Pb) | Max. 5 ppm | <5 ppm |
| Iron (Fe) | Max. 0.001% | <0.001 % |
| Calcium (Ca) | Max. 0.01% | 0.002 % |
| Magnesium (Mg) | Max. 0.005% | 0.001 % |
| Potassium (K) | Max. 0.008% | 0.003 % |
| Extraction-concentration suitability | Passes test | Passes test |
| Appearance | Passes test | Passes test |
| Identification | Passes test | Passes test |
| Solubility and foreign matter | Passes test | Passes test |
| Retained on US Standard No. 10 sieve | Max. 1% | 0.1 % |
| Retained on US Standard No. 60 sieve | Min. 94% | 97.3 % |
| Through US Standard No. 60 sieve | Max. 5% | 2.5 % |
| Through US Standard No. 100 sieve | Max. 10% | 0.1 % |

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

 **avantor**™



Material No.: 9673-33
Batch No.: 23D2462010
Manufactured Date: 2023-03-22
Retest Date: 2028-03-20
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|--|---------------|-------------|
| ACS – Assay (H ₂ SO ₄) | 95.0 – 98.0 % | 96.1 % |
| Appearance | Passes Test | Passes Test |
| ACS – Color (APHA) | ≤ 10 | 5 |
| ACS – Residue after Ignition | ≤ 3 ppm | < 1 ppm |
| ACS – Substances Reducing Permanganate (as SO ₂) | ≤ 2 ppm | < 2 ppm |
| Ammonium (NH ₄) | ≤ 1 ppm | 1 ppm |
| Chloride (Cl) | ≤ 0.1 ppm | < 0.1 ppm |
| Nitrate (NO ₃) | ≤ 0.2 ppm | < 0.1 ppm |
| Phosphate (PO ₄) | ≤ 0.5 ppm | < 0.1 ppm |
| Trace Impurities – Aluminum (Al) | ≤ 30.0 ppb | < 5.0 ppb |
| Arsenic and Antimony (as As) | ≤ 4.0 ppb | < 2.0 ppb |
| Trace Impurities – Boron (B) | ≤ 10.0 ppb | 8.5 ppb |
| Trace Impurities – Cadmium (Cd) | ≤ 2.0 ppb | < 0.3 ppb |
| Trace Impurities – Chromium (Cr) | ≤ 6.0 ppb | < 0.4 ppb |
| Trace Impurities – Cobalt (Co) | ≤ 0.5 ppb | < 0.3 ppb |
| Trace Impurities – Copper (Cu) | ≤ 1.0 ppb | < 0.1 ppb |
| Trace Impurities – Gold (Au) | ≤ 10.0 ppb | 0.5 ppb |
| Heavy Metals (as Pb) | ≤ 500.0 ppb | < 100.0 ppb |
| Trace Impurities – Iron (Fe) | ≤ 50.0 ppb | 1.3 ppb |
| Trace Impurities – Lead (Pb) | ≤ 0.5 ppb | < 0.5 ppb |
| Trace Impurities – Magnesium (Mg) | ≤ 7.0 ppb | 0.8 ppb |
| Trace Impurities – Manganese (Mn) | ≤ 1.0 ppb | < 0.4 ppb |
| Trace Impurities – Mercury (Hg) | ≤ 0.5 ppb | < 0.1 ppb |
| Trace Impurities – Nickel (Ni) | ≤ 2.0 ppb | 0.3 ppb |
| Trace Impurities – Potassium (K) | ≤ 500.0 ppb | < 2.0 ppb |
| Trace Impurities – Selenium (Se) | ≤ 50.0 ppb | < 0.1 ppb |
| Trace Impurities – Silicon (Si) | ≤ 100.0 ppb | 31.5 ppb |
| Trace Impurities – Silver (Ag) | ≤ 1.0 ppb | < 0.3 ppb |

>>> Continued on page 2 >>>

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium



Material No.: 9673-33
Batch No.: 23D2462010

| Test | Specification | Result |
|-----------------------------------|------------------|-----------|
| Trace Impurities – Sodium (Na) | ≤ 500.0 ppb | 5.4 ppb |
| Trace Impurities – Strontium (Sr) | ≤ 5.0 ppb | < 0.2 ppb |
| Trace Impurities – Tin (Sn) | ≤ 5.0 ppb | < 0.8 ppb |
| Trace Impurities – Zinc (Zn) | ≤ 5.0 ppb | 0.4 ppb |

For Laboratory, Research, or Manufacturing Use

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

avantor™



MS943 MS944
MS945 MS946

Material No.: 9530-33
Batch No.: 22G2862015
Manufactured Date: 2022-06-15
Retest Date: 2027-06-14
Revision No.: 0

Certificate of Analysis

| Test | Specification | Result |
|---|---------------|-------------|
| ACS – Assay (as HCl) (by acid–base titrn) | 36.5 – 38.0 % | 37.9 % |
| ACS – Color (APHA) | ≤ 10 | 5 |
| ACS – Residue after Ignition | ≤ 3 ppm | < 1 ppm |
| ACS – Specific Gravity at 60°/60°F | 1.185 – 1.192 | 1.191 |
| ACS – Bromide (Br) | ≤ 0.005 % | < 0.005 % |
| ACS – Extractable Organic Substances | ≤ 5 ppm | < 1 ppm |
| ACS – Free Chlorine (as Cl ₂) | ≤ 0.5 ppm | < 0.5 ppm |
| Phosphate (PO ₄) | ≤ 0.05 ppm | < 0.03 ppm |
| Sulfate (SO ₄) | ≤ 0.5 ppm | < 0.3 ppm |
| Sulfite (SO ₃) | ≤ 0.8 ppm | 0.3 ppm |
| Ammonium (NH ₄) | ≤ 3 ppm | < 1 ppm |
| Trace Impurities – Arsenic (As) | ≤ 0.010 ppm | < 0.003 ppm |
| Trace Impurities – Aluminum (Al) | ≤ 10.0 ppb | 1.3 ppb |
| Arsenic and Antimony (as As) | ≤ 5.0 ppb | < 3.0 ppb |
| Trace Impurities – Barium (Ba) | ≤ 1.0 ppb | 0.2 ppb |
| Trace Impurities – Beryllium (Be) | ≤ 1.0 ppb | < 0.2 ppb |
| Trace Impurities – Bismuth (Bi) | ≤ 10.0 ppb | < 1.0 ppb |
| Trace Impurities – Boron (B) | ≤ 20.0 ppb | < 5.0 ppb |
| Trace Impurities – Cadmium (Cd) | ≤ 1.0 ppb | < 0.3 ppb |
| Trace Impurities – Calcium (Ca) | ≤ 50.0 ppb | 163.0 ppb |
| Trace Impurities – Chromium (Cr) | ≤ 1.0 ppb | 0.7 ppb |
| Trace Impurities – Cobalt (Co) | ≤ 1.0 ppb | < 0.3 ppb |
| Trace Impurities – Copper (Cu) | ≤ 1.0 ppb | < 0.1 ppb |
| Trace Impurities – Gallium (Ga) | ≤ 1.0 ppb | < 0.2 ppb |
| Trace Impurities – Germanium (Ge) | ≤ 3.0 ppb | < 2.0 ppb |
| Trace Impurities – Gold (Au) | ≤ 4.0 ppb | 0.6 ppb |
| Heavy Metals (as Pb) | ≤ 100 ppb | < 50 ppb |
| Trace Impurities – Iron (Fe) | ≤ 15 ppb | 6 ppb |

>>> Continued on page 2 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantorsm**



Material No.: 9530-33
Batch No.: 22G2862015

| Test | Specification | Result |
|--|---------------|------------|
| Trace Impurities – Lead (Pb) | ≤ 1.0 ppb | < 0.5 ppb |
| Trace Impurities – Lithium (Li) | ≤ 1.0 ppb | < 0.2 ppb |
| Trace Impurities – Magnesium (Mg) | ≤ 10.0 ppb | 2.9 ppb |
| Trace Impurities – Manganese (Mn) | ≤ 1.0 ppb | < 0.4 ppb |
| Trace Impurities – Mercury (Hg) | ≤ 0.5 ppb | 0.1 ppb |
| Trace Impurities – Molybdenum (Mo) | ≤ 10.0 ppb | < 3.0 ppb |
| Trace Impurities – Nickel (Ni) | ≤ 4.0 ppb | < 0.3 ppb |
| Trace Impurities – Niobium (Nb) | ≤ 1.0 ppb | 0.8 ppb |
| Trace Impurities – Potassium (K) | ≤ 9.0 ppb | < 2.0 ppb |
| Trace Impurities – Selenium (Se), For Information Only | | < 1.0 ppb |
| Trace Impurities – Silicon (Si) | ≤ 100.0 ppb | < 10.0 ppb |
| Trace Impurities – Silver (Ag) | ≤ 1.0 ppb | 0.5 ppb |
| Trace Impurities – Sodium (Na) | ≤ 100.0 ppb | 2.3 ppb |
| Trace Impurities – Strontium (Sr) | ≤ 1.0 ppb | < 0.2 ppb |
| Trace Impurities – Tantalum (Ta) | ≤ 1.0 ppb | 1.6 ppb |
| Trace Impurities – Thallium (Tl) | ≤ 5.0 ppb | < 2.0 ppb |
| Trace Impurities – Tin (Sn) | ≤ 5.0 ppb | 4.0 ppb |
| Trace Impurities – Titanium (Ti) | ≤ 1.0 ppb | 1.5 ppb |
| Trace Impurities – Vanadium (V) | ≤ 1.0 ppb | < 0.2 ppb |
| Trace Impurities – Zinc (Zn) | ≤ 5.0 ppb | 0.8 ppb |
| Trace Impurities – Zirconium (Zr) | ≤ 1.0 ppb | 0.3 ppb |

>>> Continued on page 3 >>>

Hydrochloric Acid, 36.5–38.0%
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis

 **avantorsm**




Material No.: 9530-33
Batch No.: 22G2862015

| Test | Specification | Result |
|------|---------------|--------|
|------|---------------|--------|

For Laboratory, Research, or Manufacturing Use
Product Information (not specifications):
Appearance (clear, fuming liquid)
Meets ACS Specifications
Storage Condition: Store below 25 °C.

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

Certificate of Analysis

Product information

Product: Silica 60, 0.063 - 0.200 mm
REF: 815330.25
LOT: 072154301

W 3049
SP

Technical data

Material: Synthetic amorphous silica (irregular shaped)
Description: White powder

| Parameter | Specifications | Result |
|---|-------------------|--------|
| Specific surface (m ² /g, N ₂ adsorption) : | 450 - 550 | 537 |
| Particle size distribution (screen analysis) : | < 63 µm max. 5 % | 0.3 |
| | > 200 µm max. 5 % | 0.1 |
| pH value : | 6.0 - 7.5 | 7 |
| Water content (%) : | < 7 | 3.6 |
| Pore volume (mL/g, N ₂ adsorption) : | 0.65 - 0.85 | 0.82 |
| Mean pore size (Å, N ₂ adsorption) : | 50 - 70 | 62 |

Expiry

This product has no stated expiration date or shelf life.

We recommend to use the product within a time period of 5 years after date of QC release.

This time period is valid only if the product is stored under dry and frost-free conditions.

After 5 years we recommend retesting the adsorbent to make sure that the expected performance is still given.

Confirmation

Hereby we confirm, that the above mentioned product has successfully passed our quality control system in accordance with ISO 9001 and meets the specific quality criteria.

This document has been produced electronically and is valid without a signature.

Date of measurement: 16.02.2023 22:00

Certificate of analysis

W3082 Received on 2/26/2026 by IZ

Product No.: A12244
Product: Stearic acid, 98%
Lot No.: U23E020

Appearance White flakes
Assay 98.7 %

This document has been electronically generated and does not require a signature.

Order our products online www.alfa.com

ThermoFisher
S C I E N T I F I C



Certificate of Analysis

Manganous Sulfate Solution, 364 g/L**Lot Number:** 2403J02**Product Number:** 4620**Manufacture Date:** MAR 15, 2024**Expiration Date:** MAR 2026

| Name | CAS# | Grade |
|-------------------------------|------------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Manganous Sulfate Monohydrate | 10034-96-5 | Reagent |
| Sulfuric Acid | 7664-93-9 | ACS |

| Test | Specification | Result |
|-----------------------------|---------------|---------|
| Appearance | Pink liquid | Passed |
| Assay (by Refractive Index) | 360-368 g/L | 367 g/L |

| Specification | Reference |
|----------------------------|-----------------|
| Manganous Sulfate Solution | ASTM (D 888 A) |
| Manganous Sulfate Solution | ASTM (D 888 A) |
| Manganous Sulfate Solution | APHA (4500-O E) |
| Manganous Sulfate Solution | APHA (4500-O F) |
| Manganous Sulfate Solution | APHA (4500-O D) |
| Manganous Sulfate Solution | APHA (4500-O E) |
| Manganous Sulfate Solution | APHA (4500-O F) |
| Manganous Sulfate Solution | APHA (4500-O D) |
| Manganous Sulfate Solution | APHA (4500-O C) |
| Manganous Sulfate Solution | APHA (4500-O C) |
| Manganous Sulfate Solution | EPA (360.2) |
| Manganous Sulfate Solution | EPA (360.2) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 4620-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jose Pena (03/15/2024)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Sodium Thiosulfate, 0.0250 Normal (N/40)

Lot Number: 4403S13

Product Number: 7900

Manufacture Date: MAR 29, 2024

Expiration Date: SEP 2025

This product is specially formulated to increase its stability. A preservative is added to prevent bacterial contamination. However, all Sodium Thiosulfate solutions are subject to slow chemical deterioration and should be restandardized periodically.

| Name | CAS# | Grade |
|---------------------------------|-------------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Thiosulfate Pentahydrate | 10102-17-7 | ACS |
| Organic Preservative | Proprietary | |
| Sodium Carbonate | 497-19-8 | ACS |

| Test | Specification | Result | NIST SRM# |
|-------------------------------------|---------------------------|-------------------|-----------|
| Appearance | Colorless liquid | Passed | |
| Assay (vs. Potassium Iodate/Starch) | 0.02499-0.02501 N at 20°C | 0.02501 N at 20°C | 136 |

| Specification | Reference |
|--|---------------------|
| Standard Sodium Thiosulfate Solution, 0.0250 N | APHA (4500-S2- F) |
| Standard Sodium Thiosulfate Titrant | APHA (4500-O D) |
| Standard Sodium Thiosulfate Titrant | APHA (4500-O E) |
| Standard Sodium Thiosulfate Titrant | APHA (4500-O F) |
| Standard Sodium Thiosulfate Titrant, 0.025 N | APHA (4500-CI B) |
| Standard Sodium Thiosulfate Titrant | APHA (4500-O C) |
| Standard Sodium Thiosulfate Titrant, 0.025 M | APHA (5530 C) |
| Standard Sodium Thiosulfate Solution (0.025 N) | EPA (SW-846) (9031) |
| Standard Sodium Thiosulfate solution (0.025 N) | EPA (SW-846) (9034) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 7900-1 | 4 L natural poly | 18 months |
| 7900-16 | 500 mL natural poly | 18 months |
| 7900-1CT | 4 L Cubitainer® | 18 months |
| 7900-32 | 1 L natural poly | 18 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Paul Brandon (03/29/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials --
Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Certificate of Analysis

Alkaline-Iodide-Azide, Pomeroy Formulation for Dissolved Oxygen (DO) Analysis

Lot Number: 1405D67**Product Number:** 535**Manufacture Date:** APR 05, 2024**Expiration Date:** APR 2026

This solution is intended for use with samples with high Dissolved Oxygen content (above 15 mg/L) and for samples with high concentrations of organic material.

| Name | CAS# | Grade |
|------------------|------------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Iodide | 7681-82-5 | ACS |
| Sodium Hydroxide | 1310-73-2 | ACS |
| Sodium Azide | 26628-22-8 | Reagent |

| Test | Specification | Result |
|-------------|------------------|--------|
| Appearance | Colorless liquid | Passed |
| Free Iodine | To Pass Test | Passed |

| Specification | Reference |
|---|----------------|
| Alkaline Iodide-Sodium Azide Solution II | ASTM (D 888 A) |
| Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured. | |

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 535-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Heidi J Green (04/05/2024)
Operations Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

W3110
58
operate!
06/27/2024

Certificate of Analysis

1 Reagent Lane
Fair Lawn, NJ 07410
201.796.7100 tel
201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

| | | | |
|-------------------|---|-----------------------------|------------|
| Catalog Number | H303 | Quality Test / Release Date | 02/23/2024 |
| Lot Number | 235898 | | |
| Description | HEXANES - OPTIMA | | |
| Country of Origin | United States | Suggested Retest Date | Feb/2029 |
| Chemical Origin | Organic - non animal | | |
| BSE/TSE Comment | No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product. | | |

| N/A | | | |
|-----------------------------|------------|---------------------------------|-------------------------|
| Result Name | Units | Specifications | Test Value |
| APPEARANCE | | REPORT | Clear, colorless liquid |
| ASSAY (N-HEXANE) | % | >= 60 | 73 |
| ASSAY (SUM C6 HYDROCARBONS) | % | >= 99.9 | >99.9 |
| COLOR | APHA | <= 5 | <5 |
| DENSITY AT 25 DEGREES C | GM/ML | Inclusive Between 0.653 - 0.673 | 0.670 |
| EVAPORATION RESIDUE | ppm | <= 1 | 0.3 |
| FLUORESCENCE BACKGROUND | ppb | <= 1 | <1 |
| IDENTIFICATION | PASS/FAIL | = PASS TEST | PASS TEST |
| OPTICAL ABS AT 195 NM | ABS. UNITS | <= 1 | 0.64 |
| OPTICAL ABS AT 210 NM | ABS. UNITS | <= 0.25 | 0.16 |
| OPTICAL ABS AT 220 NM | ABS. UNITS | <= 0.07 | 0.06 |
| OPTICAL ABS AT 254 NM | ABS. UNITS | <= 0.005 | 0.002 |
| PESTICIDE RESIDUE ANALYSIS | NG/L | <= 10 | <10 |
| REFRACTIVE INDEX @ 25 DEG C | | Inclusive Between 1.375 - 1.385 | 1.380 |
| SUITABILITY FOR GC/MS | | = PASS TEST | PASS TEST |
| SULFUR COMPOUNDS | % | <= 0.005 | <0.005 |
| THIOPHENE | PASS/FAIL | = PASS TEST | PASS TEST |
| WATER (H2O) | % | <= 0.01 | <0.01 |
| WATER-SOLUBLE TITRABLE ACID | MEQ/G | <= 0.0003 | 0.0001 |

Harout Sahagian
Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.



Certificate of Analysis



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

| TEST | SPECIFICATION | ANALYSIS | DISPOSITION |
|--------------------|---------------|----------|-------------|
| Calcium | <= 0.005 % | <0.005 % | PASS |
| Chloride | <= 0.005 % | 0.002 % | PASS |
| Heavy Metals | <= 0.002 % | <0.002 % | PASS |
| Iron | <= 0.001 % | <0.001 % | PASS |
| Magnesium | <= 0.002 % | <0.002 % | PASS |
| Mercury | <= 0.1 ppm | <0.1 ppm | PASS |
| Nickel | <= 0.001 % | <0.001 % | PASS |
| Nitrogen Compounds | <= 0.001 % | <0.001 % | PASS |
| Phosphate | <= 0.001 % | <0.001 % | PASS |
| Potassium | <= 0.02 % | <0.02 % | PASS |
| Purity | >= 97.0 % | 99.2 % | PASS |
| Sodium Carbonate | <= 1.0 % | 0.5 % | PASS |
| Sulfate | <= 0.003 % | <0.003 % | PASS |

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

Spec Set: 0583ACS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.

Certificate of Analysis List

For request number 2018139

| Catalog Number Entered | Lot Number Entered | Related Catalog Number | Related Lot Code | Description |
|---------------------------------------|-----------------------------------|---------------------------------------|---------------------------------|-----------------------------|
| 1486227 | 4032 | 1486266 | 4032 | BOD Nutrient Buffer Pillows |

Total Enclosures: 1

***Certificate of Analysis******This is a Component of 1486266 lot A4032***

Page 1

COMMODITY: **BOD Nutrient Buffer Pillows**COMMODITY NUMBER: **1486227**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A4032****3/5/2024****3/13/2024**

| <i>TEST</i> | <i>SPECIFICATIONS</i> | <i>RESULTS</i> |
|---|------------------------------|-----------------------|
| Ammonia Concentration of a diluted pillow | 0.57 to 0.79 ppm | 0.710 ppm |
| Calcium Concentration of a diluted pillow | 0.93 to 1.29 ppm | 1.060 ppm |
| Iron Concentration of a diluted pillow | 0.27 to 0.36 ppm | 0.298 ppm |
| Magnesium Concentration of a diluted pillow | 0.35 to 0.48 ppm | 0.430 ppm |
| Phosphorus Concentration of a diluted pillow | 7.6 to 10.3 ppm | 8.05 ppm |
| pH in a 6 L of DI water | 7.1 to 7.6 | 7.32 |
| Five Day Change in Dissolved Oxygen Concentration | -0.2 to 0.2 ppm | -0.18 ppm |
| Sterility | To Pass | Passed |

The expiration date is Mar 2029

Certified by

Scott Als
Analytical Services Chemist



Certificate of Analysis

Starch Indicator, 0.5% (w/v), Mercury Free, for Iodometric Titrations

Lot Number: 4408P62

Product Number: 8000

Manufacture Date: AUG 28, 2024

Expiration Date: AUG 2026

This product is Mercury-free.

| Name | CAS# | Grade |
|-----------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Starch, soluble | 9005-84-9 | ACS |
| Salicylic Acid | 69-72-7 | ACS |

| Test | Specification | Result |
|---------------------|---|--------|
| Appearance | White translucent liquid | Passed |
| Suitability for Use | Colorless (Iodine absent) - Blue (Iodine present) | Passed |

| Specification | Reference |
|---------------------------|---------------------|
| Starch Solution | APHA (4500-S2- F) |
| Starch Indicator Solution | APHA (4500-CI B) |
| Starch Indicator | APHA (4500-SO32- B) |
| Starch indicator solution | APHA (2350 B) |
| Starch indicator solution | APHA (2350 E) |
| Starch Solution | APHA (510 B) |
| Starch Solution | APHA (5530 C) |
| Starch Indicator | APHA (4500-CI C) |
| Starch Indicator | EPA (345.1) |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 8000-1 | 4 L natural poly | 24 months |
| 8000-16 | 500 mL natural poly | 24 months |
| 8000-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in blue ink that reads "Paul Brandon". The signature is fluid and cursive, with the first name "Paul" and last name "Brandon" clearly distinguishable.

Paul Brandon (08/28/2024)
Production Manager

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.



SHIPPING DOCUMENTS

CHEMTECH

CHAIN OF CUSTODY RECORD

284 Sheffield Street, Mountainside, NJ 07092
(908) 789-8900 • Fax (908) 789-8922
www.chemtech.net

CHEMTECH PROJECT NO. **P4638**
QUOTE NO.
COC Number **2042236**

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: **Aramark uniforms**
ADDRESS: **740 Frelinghuysen Ave**
CITY **Newark** STATE **NJ** ZIP **07114**
ATTENTION: **Jarrod Mills**
PHONE: **973-826-1101** FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: **monthly**
PROJECT NO.: LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*
*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS DAYS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT _____

Handwritten notes:
104
1305
155
method 1631
method 1631

PRESERVATIVES

COMMENTS

| CHEMTECH SAMPLE ID | PROJECT SAMPLE IDENTIFICATION | SAMPLE MATRIX | SAMPLE TYPE | | SAMPLE COLLECTION | | # OF BOTTLES | | | | | | | | | | COMMENTS ← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER |
|--------------------------|----------------------------------|------------------|----------------|------|----------------------|-------|--------------|---|---|---|---|---|---|---|---|---|--|
| | | | COMP | GRAB | DATE | TIME | | 1 | 2 | 3 | 4 | 5 | 6 | 7 | 8 | 9 | |
| 1. | Grab Comp | W | ✓ | | 10-30-24 | 11:24 | 1 | ✓ | | | | | | | | | |
| 2. | | W | ✓ | | 10-30-24 | 11:26 | 2 | ✓ | ✓ | | | | | | | | |
| 3. | | | | | | | | | | | | | | | | | |
| 4. | | | | | | | | | | | | | | | | | |
| 5. | | | | | | | | | | | | | | | | | |
| 6. | | | | | | | | | | | | | | | | | |
| 7. | | | | | | | | | | | | | | | | | |
| 8. | | | | | | | | | | | | | | | | | |
| 9. | | | | | | | | | | | | | | | | | |
| 10. | | | | | | | | | | | | | | | | | |

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

Handwritten: 34°C

| | | | |
|---|---|--|---|
| RELINQUISHED BY SAMPLER: 1. <i>[Signature]</i> | DATE/TIME: 1130 10-30-24 | RECEIVED BY: 1. <i>[Signature]</i> 1130 10-30-24 | Conditions of bottles or coolers at receipt: <input type="checkbox"/> COMPLIANT <input type="checkbox"/> NON COMPLIANT <input type="checkbox"/> COOLER TEMP _____ °C Comments: _____ |
| RELINQUISHED BY SAMPLER: 2. <i>[Signature]</i> | DATE/TIME: _____ | RECEIVED BY: 2. _____ | |
| RELINQUISHED BY SAMPLER: 3. <i>[Signature]</i> | DATE/TIME: 1530 10-30-24 | RECEIVED BY: 3. _____ | |

Page _____ of _____

CLIENT: ☐ Hand Delivered ☐ Other _____
CHEMTECH: ☐ Picked Up ☐ Field Sampling

Shipment Complete
☐ YES ☐ NO



284 Sheffield Street, Mountainside NJ 07092 (908)-789-8900 Fax : 908 789 8922

Laboratory Certification

| Certified By | License No. |
|----------------------|------------------|
| | |
| CAS EPA CLP Contract | 68HERH20D0011 |
| | |
| Connecticut | PH-0830 |
| | |
| DOD ELAP (ANAB) | L2219 |
| | |
| Maine | 2024021 |
| | |
| Maryland | 296 |
| | |
| New Hampshire | 255424 Rev 1 |
| | |
| New Jersey | 20012 |
| | |
| New York | 11376 |
| | |
| Pennsylvania | 68-00548 |
| | |
| Soil Permit | 525-24-234-08441 |
| | |
| Texas | T104704488 |