

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: P4646 SAS No.: P4646 SDG No.: P4646
 Instrument ID: MSVOA_N Calibration Date(s): 10/31/2024 10/31/2024
 Heated Purge: (Y/N) N Calibration Time(s): 12:08 18:13
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN084613.D		RRF020 = VN084614.D		RRF050 = VN084615.D		
		RRF100 = VN084616.D		RRF150 = VN084617.D		RRFCAL = VN084626.D		
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRFCAL	RRF	% RSD
Acrolein	0.345	0.350	0.313	0.326	0.347		0.336	4.8
Acrylonitrile	0.838	0.902	0.889	0.892	0.937		0.892	4
1,2-Dichloroethane-d4	2.444	2.473	2.377	2.353	2.413		2.412	2
Toluene-d8	1.451	1.442	1.416	1.393	1.398		1.420	1.8
4-Bromofluorobenzene	0.482	0.506	0.531	0.538	0.526		0.517	4.4

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.