

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GARD04

Lab Code: CHEM Case No.: P4646 SAS No.: P4646 SDG No.: P4646

Instrument ID: MSVOA_N Calibration Date/Time: 10/31/2024 12:32

Lab File ID: VN084614.D Init. Calib. Date(s): 10/31/2024 10/31/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 12:08 18:13

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.336	0.350		4.17	
Acrylonitrile	0.892	0.902		1.12	
1,2-Dichloroethane-d4	2.412	2.473		2.53	
Toluene-d8	1.420	1.442		1.55	
4-Bromofluorobenzene	0.517	0.506		-2.13	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.