

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4646

Client: Garden State Laboratories, Inc.

Analytical Method: SW8260-Low

Datafile : VX043641.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1031WBSD01	Dichlorodifluoromethane	20	16.4	ug/L	82	5		69	116	20
	Chloromethane	20	15.2	ug/L	76	4		65	116	20
	Vinyl chloride	20	17.2	ug/L	86	0		65	117	20
	Bromomethane	20	17.6	ug/L	88	2		58	125	20
	Chloroethane	20	18.8	ug/L	94	7		56	128	20
	Trichlorofluoromethane	20	20.2	ug/L	101	2		73	115	20
	1,1,2-Trichlorotrifluoroethane	20	20.2	ug/L	101	1		80	112	20
	1,1-Dichloroethene	20	19.3	ug/L	97	3		74	110	20
	Acetone	100	86.1	ug/L	86	5		60	125	20
	Carbon disulfide	20	15.9	ug/L	79	1		64	112	20
	Methyl tert-butyl Ether	20	18.0	ug/L	90	5		78	114	20
	Methyl Acetate	20	17.6	ug/L	88	5		67	125	20
	Methylene Chloride	20	17.6	ug/L	88	3		72	114	20
	trans-1,2-Dichloroethene	20	18.4	ug/L	92	2		75	108	20
	1,1-Dichloroethane	20	18.3	ug/L	92	3		78	112	20
	Cyclohexane	20	19.2	ug/L	96	2		75	110	20
	2-Butanone	100	89.4	ug/L	89	7		65	122	20
	Carbon Tetrachloride	20	19.8	ug/L	99	2		77	113	20
	cis-1,2-Dichloroethene	20	18.6	ug/L	93	3		77	110	20
	Bromochloromethane	20	16.7	ug/L	84	4		70	124	20
	Chloroform	20	18.3	ug/L	92	2		79	113	20
	1,1,1-Trichloroethane	20	18.8	ug/L	94	1		80	108	20
	Methylcyclohexane	20	20.5	ug/L	103	0		72	115	20
	Benzene	20	19.4	ug/L	97	0		82	109	20
	1,2-Dichloroethane	20	18.4	ug/L	92	3		80	115	20
	Trichloroethene	20	20.1	ug/L	101	0		77	113	20
	1,2-Dichloropropane	20	18.7	ug/L	94	3		83	111	20
	Bromodichloromethane	20	19.0	ug/L	95	4		83	110	20
	4-Methyl-2-Pentanone	100	96.6	ug/L	97	6		74	118	20
	Toluene	20	20.8	ug/L	104	2		82	110	20
	t-1,3-Dichloropropene	20	18.6	ug/L	93	3		79	110	20
	cis-1,3-Dichloropropene	20	19.7	ug/L	99	4		82	110	20
	1,1,2-Trichloroethane	20	20.0	ug/L	100	2		83	112	20
	2-Hexanone	100	96.6	ug/L	97	7		73	117	20
	Dibromochloromethane	20	19.5	ug/L	98	6		82	110	20
	1,2-Dibromoethane	20	19.8	ug/L	99	4		81	110	20
	Tetrachloroethene	20	21.3	ug/L	106	2		67	123	20
	Chlorobenzene	20	19.9	ug/L	100	1		82	109	20
	Ethyl Benzene	20	20.3	ug/L	102	2		83	109	20
	m/p-Xylenes	40	41.3	ug/L	103	1		82	110	20
	o-Xylene	20	20.2	ug/L	101	3		83	109	20
	Styrene	20	20.6	ug/L	103	1		80	111	20
	Bromoform	20	18.8	ug/L	94	3		79	109	20
	Isopropylbenzene	20	20.5	ug/L	103	1		83	112	20
	1,1,2,2-Tetrachloroethane	20	19.9	ug/L	100	6		76	118	20
	1,3-Dichlorobenzene	20	20.3	ug/L	102	1		82	108	20
	1,4-Dichlorobenzene	20	20.2	ug/L	101	1		82	107	20
	1,2-Dichlorobenzene	20	20.5	ug/L	103	1		82	109	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4646

Client: Garden State Laboratories, Inc.

Analytical Method: SW8260-Low

Datafile : VX043641.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1031WBSD01	1,2-Dibromo-3-Chloropropane	20	17.1	ug/L	86	2		68	112	20
	1,2,4-Trichlorobenzene	20	20.2	ug/L	101	1		75	113	20
	1,2,3-Trichlorobenzene	20	19.9	ug/L	100	0		76	114	20