

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GARD04
 Lab Code: CHEM Case No.: P4646 SAS No.: P4646 SDG NO.: P4646
 Lab File ID: VX043637.D Date Analyzed: 10/31/2024
 Instrument ID: MSVOA_X Time Analyzed: 09:26
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	102825	12.018				
UPPER LIMIT	205650	12.518				
LOWER LIMIT	51412.5	11.518				
EPA SAMPLE NO.						
241030069-01-VOA	73845	12.02				
241030043-05-TRIP-BLANK	72782	12.02				
VX1031WBL01	95694	12.02				
VX1031WBS01	135150	12.02				
VX1031WBSD01	129133	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.