

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/01/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 11:10 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	8.61	8.59	8.49	8.69	-0.02
Aroclor-1268-2 (2)	8.70	8.68	8.58	8.78	-0.02
Aroclor-1268-3 (3)	8.93	8.91	8.81	9.01	-0.02
Aroclor-1268-4 (4)	9.34	9.31	9.21	9.41	-0.03
Aroclor-1268-5 (5)	9.75	9.72	9.62	9.82	-0.03
Tetrachloro-m-xylene	4.38	4.37	4.27	4.47	-0.01
Decachlorobiphenyl	10.09	10.06	9.96	10.16	-0.03

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/01/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 11:10 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1268-1 (1)	7.55	7.54	7.44	7.64	-0.01
Aroclor-1268-2 (2)	7.61	7.61	7.51	7.71	0.00
Aroclor-1268-3 (3)	7.82	7.81	7.71	7.91	-0.01
Aroclor-1268-4 (4)	8.10	8.10	8.00	8.20	0.00
Aroclor-1268-5 (5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene	3.64	3.65	3.55	3.75	0.01
Decachlorobiphenyl	8.65	8.64	8.54	8.74	-0.01