

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/01/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 23:12 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1268-1 (1)   | 8.59       | 8.59      | 8.49      | 8.69  | 0.00       |
| Aroclor-1268-2 (2)   | 8.69       | 8.68      | 8.58      | 8.78  | -0.01      |
| Aroclor-1268-3 (3)   | 8.92       | 8.91      | 8.81      | 9.01  | 0.00       |
| Aroclor-1268-4 (4)   | 9.32       | 9.31      | 9.21      | 9.41  | -0.01      |
| Aroclor-1268-5 (5)   | 9.73       | 9.72      | 9.62      | 9.82  | -0.01      |
| Tetrachloro-m-xylene | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   | 10.07      | 10.06     | 9.96      | 10.16 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/01/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 23:12 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1268-1       | (1) | 7.54       | 7.54      | 7.44      | 7.64 | 0.00       |
| Aroclor-1268-2       | (2) | 7.61       | 7.61      | 7.51      | 7.71 | 0.00       |
| Aroclor-1268-3       | (3) | 7.81       | 7.81      | 7.71      | 7.91 | 0.00       |
| Aroclor-1268-4       | (4) | 8.10       | 8.10      | 8.00      | 8.20 | 0.00       |
| Aroclor-1268-5       | (5) | 8.39       | 8.39      | 8.29      | 8.49 | 0.00       |
| Tetrachloro-m-xylene |     | 3.64       | 3.65      | 3.55      | 3.75 | 0.01       |
| Decachlorobiphenyl   |     | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |