

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/05/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 01:37 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|-------|------------|
|                      |     |            |           | FROM      | TO    |            |
| Aroclor-1254-1       | (1) | 6.38       | 6.38      | 6.28      | 6.48  | 0.00       |
| Aroclor-1254-2       | (2) | 6.60       | 6.59      | 6.49      | 6.69  | -0.01      |
| Aroclor-1254-3       | (3) | 6.96       | 6.96      | 6.86      | 7.06  | 0.00       |
| Aroclor-1254-4       | (4) | 7.25       | 7.24      | 7.14      | 7.34  | -0.01      |
| Aroclor-1254-5       | (5) | 7.67       | 7.66      | 7.56      | 7.76  | -0.01      |
| Tetrachloro-m-xylene |     | 4.37       | 4.37      | 4.27      | 4.47  | 0.00       |
| Decachlorobiphenyl   |     | 10.07      | 10.06     | 9.96      | 10.16 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

Continuing Calib Date: 11/05/2024 Initial Calibration Date(s): 10/15/2024 10/16/2024

Continuing Calib Time: 01:37 Initial Calibration Time(s): 18:27 02:36

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1254-1       | (1) | 5.52       | 5.53      | 5.43      | 5.63 | 0.01       |
| Aroclor-1254-2       | (2) | 5.67       | 5.67      | 5.57      | 5.77 | 0.00       |
| Aroclor-1254-3       | (3) | 6.07       | 6.08      | 5.98      | 6.18 | 0.01       |
| Aroclor-1254-4       | (4) | 6.30       | 6.30      | 6.20      | 6.40 | 0.00       |
| Aroclor-1254-5       | (5) | 6.72       | 6.72      | 6.62      | 6.82 | 0.00       |
| Tetrachloro-m-xylene |     | 3.64       | 3.64      | 3.54      | 3.74 | 0.00       |
| Decachlorobiphenyl   |     | 8.64       | 8.64      | 8.54      | 8.74 | 0.00       |