

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107617.D Time Analyzed: 22:07

| COMPOUND             | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|--------|-----------|--------|--------------------|-------------------|------|
|                      |        | FROM      | TO     |                    |                   |      |
| Aroclor-1016-1       | 5.521  | 5.423     | 5.623  | 520.120            | 500.000           | 4.0  |
| Aroclor-1016-2       | 5.544  | 5.445     | 5.645  | 502.650            | 500.000           | 0.5  |
| Aroclor-1016-3       | 5.606  | 5.507     | 5.707  | 505.120            | 500.000           | 1.0  |
| Aroclor-1016-4       | 5.703  | 5.604     | 5.804  | 534.940            | 500.000           | 7.0  |
| Aroclor-1016-5       | 5.998  | 5.899     | 6.099  | 518.960            | 500.000           | 3.8  |
| Aroclor-1260-1       | 7.127  | 7.026     | 7.226  | 502.860            | 500.000           | 0.6  |
| Aroclor-1260-2       | 7.384  | 7.282     | 7.482  | 522.550            | 500.000           | 4.5  |
| Aroclor-1260-3       | 7.746  | 7.643     | 7.843  | 528.660            | 500.000           | 5.7  |
| Aroclor-1260-4       | 7.971  | 7.868     | 8.068  | 552.120            | 500.000           | 10.4 |
| Aroclor-1260-5       | 8.285  | 8.181     | 8.381  | 522.950            | 500.000           | 4.6  |
| Decachlorobiphenyl   | 10.071 | 9.958     | 10.158 | 47.880             | 50.000            | -4.2 |
| Tetrachloro-m-xylene | 4.371  | 4.274     | 4.474  | 51.090             | 50.000            | 2.2  |

### CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL02 Date Analyzed: 11/01/2024

Lab Sample No.: AR1660CCC500 Data File : PO107617.D Time Analyzed: 22:07

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D   |
|----------------------|-------|-----------|-------|--------------------|-------------------|------|
|                      |       | FROM      | TO    |                    |                   |      |
| Aroclor-1016-1       | 4.723 | 4.626     | 4.826 | 507.760            | 500.000           | 1.6  |
| Aroclor-1016-2       | 4.742 | 4.645     | 4.845 | 522.490            | 500.000           | 4.5  |
| Aroclor-1016-3       | 4.918 | 4.821     | 5.021 | 502.610            | 500.000           | 0.5  |
| Aroclor-1016-4       | 4.960 | 4.862     | 5.062 | 466.890            | 500.000           | -6.6 |
| Aroclor-1016-5       | 5.173 | 5.076     | 5.276 | 501.120            | 500.000           | 0.2  |
| Aroclor-1260-1       | 6.204 | 6.107     | 6.307 | 491.540            | 500.000           | -1.7 |
| Aroclor-1260-2       | 6.392 | 6.294     | 6.494 | 520.280            | 500.000           | 4.1  |
| Aroclor-1260-3       | 6.545 | 6.448     | 6.648 | 503.430            | 500.000           | 0.7  |
| Aroclor-1260-4       | 7.016 | 6.918     | 7.118 | 504.530            | 500.000           | 0.9  |
| Aroclor-1260-5       | 7.258 | 7.159     | 7.359 | 527.090            | 500.000           | 5.4  |
| Decachlorobiphenyl   | 8.639 | 8.539     | 8.739 | 50.530             | 50.000            | 1.1  |
| Tetrachloro-m-xylene | 3.643 | 3.545     | 3.745 | 51.220             | 50.000            | 2.4  |