

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/04/2024

Lab Sample No.: AR1660CCC500 Data File : PO107624.D Time Analyzed: 09:39

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.522	5.423	5.623	489.910	500.000	-2.0
Aroclor-1016-2	5.544	5.445	5.645	475.890	500.000	-4.8
Aroclor-1016-3	5.607	5.507	5.707	470.250	500.000	-6.0
Aroclor-1016-4	5.704	5.604	5.804	490.430	500.000	-1.9
Aroclor-1016-5	5.999	5.899	6.099	493.770	500.000	-1.2
Aroclor-1260-1	7.128	7.026	7.226	487.280	500.000	-2.5
Aroclor-1260-2	7.384	7.282	7.482	506.380	500.000	1.3
Aroclor-1260-3	7.746	7.643	7.843	507.530	500.000	1.5
Aroclor-1260-4	7.971	7.868	8.068	529.190	500.000	5.8
Aroclor-1260-5	8.285	8.181	8.381	519.740	500.000	3.9
Decachlorobiphenyl	10.067	9.958	10.158	48.500	50.000	-3.0
Tetrachloro-m-xylene	4.372	4.274	4.474	49.380	50.000	-1.2

CALIBRATION VERIFICATION SUMMARY

Contract: BSIG01

Lab Code: CHEM Case No.: P4647 SAS No.: P4647 SDG NO.: P4647

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/15/2024 10/15/2024

Client Sample No.: CCAL03 Date Analyzed: 11/04/2024

Lab Sample No.: AR1660CCC500 Data File : PO107624.D Time Analyzed: 09:39

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.724	4.626	4.826	517.590	500.000	3.5
Aroclor-1016-2	4.743	4.645	4.845	535.890	500.000	7.2
Aroclor-1016-3	4.919	4.821	5.021	513.970	500.000	2.8
Aroclor-1016-4	4.960	4.862	5.062	479.850	500.000	-4.0
Aroclor-1016-5	5.173	5.076	5.276	510.340	500.000	2.1
Aroclor-1260-1	6.205	6.107	6.307	494.360	500.000	-1.1
Aroclor-1260-2	6.392	6.294	6.494	524.820	500.000	5.0
Aroclor-1260-3	6.546	6.448	6.648	512.150	500.000	2.4
Aroclor-1260-4	7.016	6.918	7.118	512.060	500.000	2.4
Aroclor-1260-5	7.257	7.159	7.359	541.040	500.000	8.2
Decachlorobiphenyl	8.639	8.539	8.739	54.030	50.000	8.1
Tetrachloro-m-xylene	3.644	3.545	3.745	52.550	50.000	5.1