

Laboratory Control Sample Summary

Client: Tetra Tech, EMI

SDG No.: P4648

Project: R36720

Run No.: LB133263

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133263BS							
Oil and Grease	mg/L	20.0	16.8		84	1	78-114	11/03/2024

Laboratory Control Sample Summary

Client: Tetra Tech, EMI

SDG No.: P4648

Project: R36720

Run No.: LB133263

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB133263BSD							
Oil and Grease	mg/L	20.0	17.0		85	1	78-114	11/03/2024