

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4652
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4652 **SAS No.:** P4652
Matrix: Water **Sample ID:** P4661-03 **Client ID:** 102324-ADUP
Percent Solids for Sample: NA **Duplicate ID** P4661-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.17	J	0.13	J	25		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tetra Tech NUS, Inc.</u>	Level:	<u>LOW</u>	SDG No.:	<u>P4652</u>
Contract:	<u>TETR06</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>P4652</u>
Matrix:	<u>Water</u>	Sample ID:	<u>P4661-03MS</u>	Client ID:	<u>102324-AMSD</u>
Percent Solids for Sample:	<u>NA</u>	Duplicate ID	<u>P4661-03MSD</u>	Percent Solids for Spike Sample:	<u>NA</u>

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	3.36		3.26		3		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4652
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4652 **SAS No.:** P4652
Matrix: Water **Sample ID:** P4673-07 **Client ID:** OUTFALL-007DUP
Percent Solids for Sample: NA **Duplicate ID** P4673-07DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	519		521		0		P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	248		247		0		P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	18400		18200		1		P
Chromium	ug/L	20	21.1		21.0		0		P
Cobalt	ug/L	20	15.0	U	15.0	U			P
Copper	ug/L	20	14.2		14.4		1		P
Iron	ug/L	20	892		908		2		P
Lead	ug/L	20	6.00	U	6.00	U			P
Magnesium	ug/L	20	676	J	671	J	1		P
Manganese	ug/L	20	26.8		26.5		1		P
Nickel	ug/L	20	2.30	J	2.41	J	5		P
Potassium	ug/L	20	2620		2670		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	3780		3850		2		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	286		289		1		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4652
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4652 **SAS No.:** P4652
Matrix: Water **Sample ID:** P4673-07MS **Client ID:** OUTFALL-007MSD
Percent Solids for Sample: NA **Duplicate ID** P4673-07MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1650		1640		1		P
Antimony	ug/L	20	409		411		0		P
Arsenic	ug/L	20	413		414		0		P
Barium	ug/L	20	378		376		1		P
Beryllium	ug/L	20	106		105		1		P
Cadmium	ug/L	20	95.9		96.3		0		P
Calcium	ug/L	20	19900		19800		1		P
Chromium	ug/L	20	236		234		1		P
Cobalt	ug/L	20	100		101		1		P
Copper	ug/L	20	168		169		1		P
Iron	ug/L	20	2600		2590		0		P
Lead	ug/L	20	475		476		0		P
Magnesium	ug/L	20	1730		1730		0		P
Manganese	ug/L	20	137		136		1		P
Nickel	ug/L	20	250		251		0		P
Potassium	ug/L	20	8060		8090		0		P
Selenium	ug/L	20	954		964		1		P
Silver	ug/L	20	38.9		38.7		1		P
Sodium	ug/L	20	5890		5950		1		P
Thallium	ug/L	20	947		978		3		P
Vanadium	ug/L	20	167		166		1		P
Zinc	ug/L	20	411		412		0		P