

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **SDG No.:** P4652
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4652 **SAS No.:** P4652

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164645BS							
Aluminum	ug/L	1000	1010		101	86 - 115	P
Antimony	ug/L	400	397		99	88 - 113	P
Arsenic	ug/L	400	395		99	87 - 113	P
Barium	ug/L	100	101		101	88 - 113	P
Beryllium	ug/L	100	102		102	89 - 112	P
Cadmium	ug/L	100	98.5		98	88 - 113	P
Calcium	ug/L	500	515	J	103	87 - 113	P
Chromium	ug/L	200	205		102	90 - 113	P
Cobalt	ug/L	100	99.8		100	89 - 114	P
Copper	ug/L	150	157		105	86 - 114	P
Iron	ug/L	1500	1470		98	87 - 115	P
Lead	ug/L	500	492		98	86 - 113	P
Magnesium	ug/L	1000	1000		100	85 - 113	P
Manganese	ug/L	100	103		103	90 - 114	P
Nickel	ug/L	250	250		100	88 - 113	P
Potassium	ug/L	5000	4740		95	86 - 114	P
Selenium	ug/L	1000	992		99	83 - 114	P
Silver	ug/L	37.5	36.9		98	84 - 115	P
Sodium	ug/L	1500	1360		91	87 - 115	P
Thallium	ug/L	1000	1020		102	85 - 114	P
Vanadium	ug/L	150	153		102	90 - 111	P
Zinc	ug/L	100	103		103	87 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc.

SDG No.: P4652

Contract: TETR06

Lab Code: CHEM

Case No.: P4652

SAS No.: P4652

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164687BS Mercury	ug/L	4.0	3.57		89	82 - 119	CV