

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN103024

Review By	yogesh	Review On	11/4/2024 1:33:07 AM
Supervise By	mohammad	Supervise On	11/4/2024 1:50:10 AM
SubDirectory	BN103024	HP Acquire Method	BNA_N, 8270_HP Processing Method bn103024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN034739.D	30 Oct 2024 08:41		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN034740.D	30 Oct 2024 09:20	Compound#20 removed from 0.1 ppm	RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN034741.D	30 Oct 2024 09:56		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN034742.D	30 Oct 2024 10:32	Compound#20 Kept on LR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN034743.D	30 Oct 2024 11:08	The Calibration is Good For DOD	RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN034744.D	30 Oct 2024 11:44		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN034745.D	30 Oct 2024 12:20		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN034746.D	30 Oct 2024 12:56		RC/JU	Ok
9	SSTDICV0.4	ICVBN103024	BN034747.D	30 Oct 2024 15:21		RC/JU	Ok
10	PB164323BL	PB164323BL	BN034748.D	30 Oct 2024 15:57	END CCC Missing	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110424

Review By	Jagrut	Review On	11/5/2024 4:35:19 PM
Supervise By	mohammad	Supervise On	11/6/2024 3:14:45 AM
SubDirectory	BN110424	HP Acquire Method	BNA_N, 8270_HP Processing Method bn103024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN034836.D	04 Nov 2024 09:30		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN034837.D	04 Nov 2024 10:45		RC/JU	Ok,M
3	PB164594BL	PB164594BL	BN034838.D	04 Nov 2024 11:21		RC/JU	Ok
4	P4652-01	RW5-SP100-50-202410	BN034839.D	04 Nov 2024 14:55	Need 5X Dilution	RC/JU	Dilution
5	P4652-02	RW5-SP201-50-202410	BN034840.D	04 Nov 2024 15:31		RC/JU	Ok,M
6	P4652-03	RW5-SP301-303-50-20	BN034841.D	04 Nov 2024 16:07		RC/JU	Ok
7	P4652-01DL	RW5-SP100-50-202410	BN034842.D	04 Nov 2024 16:43	Internal Standard spiking error	RC/JU	Not Ok
8	P4652-05	RW5-SP201-90-202410	BN034843.D	04 Nov 2024 17:19		RC/JU	Ok
9	P4652-06	RW5-SP301-303-90-20	BN034844.D	04 Nov 2024 17:54		RC/JU	Ok
10	P4652-01DL	RW5-SP100-50-202410	BN034845.D	04 Nov 2024 18:30		RC/JU	Ok
11	P4665-05	BP-VPB-190-GW-558-5	BN034846.D	04 Nov 2024 19:06		RC/JU	Ok
12	P4665-07	BP-VPB-190-GW-598-6	BN034847.D	04 Nov 2024 19:42		RC/JU	Ok
13	SSTDCCC0.4	SSTDCCC0.4EC	BN034848.D	04 Nov 2024 20:18		RC/JU	Ok
14	DFTPP	DFTPP	BN034849.D	04 Nov 2024 20:54		RC/JU	Ok
15	SSTDCCC0.4	SSTDCCC0.4	BN034850.D	04 Nov 2024 21:33		RC/JU	Ok
16	PB164511BL	PB164511BL	BN034851.D	04 Nov 2024 22:09		RC/JU	Ok
17	PB164511BS	PB164511BS	BN034852.D	04 Nov 2024 22:45		RC/JU	Ok
18	PB164511BSD	PB164511BSD	BN034853.D	04 Nov 2024 23:21		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110424

Review By	Jagrut	Review On	11/5/2024 4:35:19 PM
Supervise By	mohammad	Supervise On	11/6/2024 3:14:45 AM
SubDirectory	BN110424	HP Acquire Method	BNA_N, 8270_HP Processing Method bn103024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597		
CCC	SP6601		
Internal Standard/PEM	SP6527		
ICV/I.BLK	SP6548		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P4600-03	BP-VPB-190-GW-438-4	BN034854.D	04 Nov 2024 23:57		RC/JU	Ok
20	P4600-05	BP-VPB-190-GW-478-4	BN034855.D	05 Nov 2024 00:33		RC/JU	Ok
21	P4600-07	BP-VPB-190-GW-518-5	BN034856.D	05 Nov 2024 01:09		RC/JU	Ok
22	P4652-04	RW5-SP100-90-202410	BN034857.D	05 Nov 2024 01:45	Need 5X Dilution	RC/JU	Dilution
23	P4665-02	VPB190-HYD-2024103	BN034858.D	05 Nov 2024 02:21	Surrogate Fail	RC/JU	ReRun
24	PB164594BS	PB164594BS	BN034859.D	05 Nov 2024 02:57		RC/JU	Ok
25	PB164594BSD	PB164594BSD	BN034860.D	05 Nov 2024 03:33		RC/JU	Ok
26	SSTDCCC0.4	SSTDCCC0.4EC	BN034861.D	05 Nov 2024 04:09		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN110524

Review By	yogesh	Review On	11/6/2024 3:10:02 AM
Supervise By	mohammad	Supervise On	11/6/2024 3:16:41 AM
SubDirectory	BN110524	HP Acquire Method	BNA_N
		HP Processing Method	bn103024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN034862.D	05 Nov 2024 08:44		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN034863.D	05 Nov 2024 09:23		RC/JU	Ok
3	PB164594BL	PB164594BL	BN034864.D	05 Nov 2024 10:37		RC/JU	Ok
4	P4652-04DL	RW5-SP100-90-202410	BN034865.D	05 Nov 2024 11:13		RC/JU	Ok
5	P4665-02RE	VPB190-HYD-2024103	BN034866.D	05 Nov 2024 11:49	Surrogate Fail	RC/JU	Confirms
6	SSTDCCC0.4	SSTDCCC0.4EC	BN034867.D	05 Nov 2024 12:46		RC/JU	Ok

M : Manual Integration