

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4652 SAS No.: P4652 SDG NO.: P4652

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 11/05/2024

Lab File ID: BN034863.D Time Analyzed: 09:23

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8979	7.589	27748	10.35	13645	14.22
UPPER LIMIT	17958	8.089	55496	10.851	27290	14.722
LOWER LIMIT	4489.5	7.089	13874	9.851	6822.5	13.722
EPA SAMPLE NO.						
01 PB164594BL	7362	7.59	21619	10.35	10048	14.22
02 RW5-SP100-90-20241030DL	6432	7.59	19375	10.35	9318	14.22

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.