

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51810 MA No.: _____ SDG No.: MC0VG5
Instrument ID: CV1 Start Date: 11/21/2024
Analytical Method: CVAA Run Batch: LB133546
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.060	0	0.2	0.25	26	2.5	2.5	-2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51810 MA No.: _____ SDG No.: MC0VG5
Instrument ID: CV1 Start Date: 11/21/2024
Analytical Method: CVAA Run Batch: LB133546
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.8	-4	7.5	7.6	1	10	10	1