

SDG NARRATIVE

USEPA
SDG # MC0VG5
CASE # 51810
CONTRACT # 68HERH20D0011
SOW# SFAM01.1
LAB NAME: Alliance Technical Group, LLC
LAB CODE: ACE
LAB ORDER ID # P4654

A. Number of Samples and Date of Receipt

19 Soil samples were delivered to the laboratory intact on 10/31/2024, 11/02/2024

B. Parameters

Test requested for Mercury.

C. Cooler Temp

Indicator Bottle: Presence/

Absence Cooler: 2.4°C

D. Detail Documentation (related to Sample Handling Shipping, Analytical Problem, Temp of Cooler etc):

Issue 1 : A "P" or "M" prefix was listed at the beginning of a CLP sample ID.

E. Corrective Action taken for above:

Resolution 1: To maintain COC integrity, ASB requests no changes to the Sample IDs. The laboratory will note the issue in the SDG Narrative and proceed with the analysis of the samples.

F. Analytical Techniques:

All analyses were based on CLP Methodology by method SFAM01.1.

G. Calculation:

Calculation for Hg Soil Sample:

Conversion of Results from µg /L or ppb to mg/kg:



284 Sheffield Street Mountainside, NJ 07092

Concentration (mg/kg) =
$$C \times Vf \times DF / 1000$$

W x S

Where,

C = Instrument response in μ g/L from the calibration curve.

Vf = Final prepared (absorbing solution) volume (mL)

W = Initial aliquot amount (g) (Fraction of Sample amount taken in prep)

S = % Solids / 100 (Fraction of Percent Solids)

DF = Dilution Factor

Example Calculation For Sample MC0VG5:

$$\begin{array}{ll} If \ C &= 1.0674 \ ppb \\ Vf = 100 \ mL \\ W &= 0.60g \\ S &= 0.857(85.7/100) \\ DF &= 1 \end{array}$$

Concentration (mg/kg) =
$$1.0674 \text{ x} \frac{100}{0.60 \text{ x} 0.857} \text{ x } 1 / 1000$$

= 0.207584 mg/kg

= 0.21 mg/kg (Reported Result with Signification)

H. QA/QC

Calibrations met requirements. Blank analyses did not indicate any presence of contamination. Spike sample did meet requirements. Duplicate sample did meet requirements.

I certify that the data package is in compliance with the terms and conditions of the contract both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the Laboratory Director or his designee, as verified by the following signature.

Signature	_ Name: Nimisha Pandya
Date	Title: Document Control Officer
Daic	THE DOCUMENT CONTROL OFFICE