

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51810 MA No.: _____ SDG No.: MC0VD3
Instrument ID: CV1 Start Date: 11/19/2024
Analytical Method: CVAA Run Batch: LB133521
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.21	7	2.5	2.5	1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51810 MA No.: _____ SDG No.: MC0VD3
Instrument ID: CV1 Start Date: 11/19/2024
Analytical Method: CVAA Run Batch: LB133521
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	5.0	-1	7.5	7.4	-1	10	10	1