

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51810 MA No. : SDG No.: MC0VD3
 Initial Calibration Verification Source : MP83287
 Continuing Calibration Verification Source : MP83289
 Run Batch: LB133521 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV092				ID: CCV014				ID: CCV015		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.8	95		5.0	5.1	102		5.1	103	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51810 MA No. : SDG No.: MC0VD3
Initial Calibration Verification Source : MP83287
Continuing Calibration Verification Source : MP83289
Run Batch: LB133521 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV016				ID: CCV017		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	5.2	104		5.2	103	