


LB133521

SFAM01.1 INSTRUMENT ID: CV1



A= 0.0000e+000
B= 2.3766e-004 slope
C= -4.1308e-002 y-intercept
Rho= 0.9999221
Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.003	-0.003	163	0.000	163					
0.05	0.050					442					
0.20	0.200	0.213	0.013	1072	0.0 %	1072	0				
2.50	2.500	2.532	0.032	10828	0.0 %	10828					
5.00	5.000	4.974	-0.026	21102	0.0 %	21102					
7.50	7.500	7.418	-0.082	31388	0.0 %	31388					
10.0	10.000	10.065	0.065	42522	0.0 %	42522					

LB133521

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	163	-		0 SFAM01.1	PPB	11/19/2024 17:34	Std
	0.2 S01	1072	-		0.2 SFAM01.1	PPB	11/19/2024 17:37	Std
	2.5 S02	10828	-		2.5 SFAM01.1	PPB	11/19/2024 17:41	Std
	5 S03	21102	-		5 SFAM01.1	PPB	11/19/2024 17:43	Std
	7.5 S04	31388	-		7.5 SFAM01.1	PPB	11/19/2024 17:45	Std
	10 S05	42522	-		10 SFAM01.1	PPB	11/19/2024 17:48	Std
ICV092	ICV092	16241	3.8186 -		SFAM01.1	PPB	11/19/2024 17:53	SMPL
ICB092	ICB092	30	-0.0342 -		SFAM01.1	PPB	11/19/2024 17:55	SMPL
CCV014	CCV014	21709	5.1181 -		SFAM01.1	PPB	11/19/2024 17:57	SMPL
CCB014	CCB014	-25	-0.0472 -		SFAM01.1	PPB	11/19/2024 18:00	SMPL
PB165119BL	PBS119	173	-0.0002 -		SFAM01.1	PPB	11/19/2024 18:02	SMPL
P4650-01	MC0VD4	2993	0.67 -		SFAM01.1	PPB	11/19/2024 18:04	SMPL
P4650-02	MC0VD7	5417	1.2461 -		SFAM01.1	PPB	11/19/2024 18:07	SMPL
P4650-03	MC0VD8	2271	0.4984 -		SFAM01.1	PPB	11/19/2024 18:09	SMPL
P4650-04	MC0VD9	2564	0.5681 -		SFAM01.1	PPB	11/19/2024 18:11	SMPL
P4650-05	MC0VE3	7722	1.7939 -		SFAM01.1	PPB	11/19/2024 18:13	SMPL
P4650-06	MC0VE4	5687	1.3103 -		SFAM01.1	PPB	11/19/2024 18:16	SMPL
P4650-07	MC0VE5	7092	1.6442 -		SFAM01.1	PPB	11/19/2024 18:18	SMPL
P4650-08	MC0VE6	3671	0.8312 -		SFAM01.1	PPB	11/19/2024 18:20	SMPL
P4650-09	MC0VF2	2056	0.4473 -		SFAM01.1	PPB	11/19/2024 18:23	SMPL
P4650-10	MC0VF3	3990	0.907 -		SFAM01.1	PPB	11/19/2024 18:25	SMPL
P4650-11	MC0VF4	4890	1.1209 -		SFAM01.1	PPB	11/19/2024 18:27	SMPL
P4650-12	MC0VF7	1407	0.2931 -		SFAM01.1	PPB	11/19/2024 18:29	SMPL
P4650-13	MC0VF8	3948	0.897 -		SFAM01.1	PPB	11/19/2024 18:32	SMPL
P4650-14	MC0VF9	3897	0.8849 -		SFAM01.1	PPB	11/19/2024 18:34	SMPL
P4650-15	MC0VG0	2201	0.4818 -		SFAM01.1	PPB	11/19/2024 18:36	SMPL
P4650-16	MC0VG1	2913	0.651 -		SFAM01.1	PPB	11/19/2024 18:38	SMPL
P4650-17	MC0VG2	2961	0.6624 -		SFAM01.1	PPB	11/19/2024 18:41	SMPL
P4650-18	MC0VG3	3415	0.7703 -		SFAM01.1	PPB	11/19/2024 18:43	SMPL
P4650-19	MC0VG4	7091	1.644 -		SFAM01.1	PPB	11/19/2024 18:45	SMPL
CCV015	CCV015	21773	5.1333 -		SFAM01.1	PPB	11/19/2024 18:47	SMPL
CCB015	CCB015	61	-0.0268 -		SFAM01.1	PPB	11/19/2024 18:50	SMPL
P4650-20	MC0VH6	2462	0.5438 -		SFAM01.1	PPB	11/19/2024 18:52	SMPL
P4650-21	MC0VH6D	2477	0.5474 -		SFAM01.1	PPB	11/19/2024 18:54	SMPL
P4650-22	MC0VH6S	15181	3.5667 -		SFAM01.1	PPB	11/19/2024 18:57	SMPL
PB165120BL	PBS120	63	-0.0263 -		SFAM01.1	PPB	11/19/2024 18:59	SMPL
P4655-01	MC0VD3	9318	2.1732 -		SFAM01.1	PPB	11/19/2024 19:01	SMPL
P4655-02	MC0VE1	2571	0.5697 -		SFAM01.1	PPB	11/19/2024 19:03	SMPL
P4655-03	MC0VE2	5288	1.2155 -		SFAM01.1	PPB	11/19/2024 19:06	SMPL
P4655-04	MC0VE7	2428	0.5357 -		SFAM01.1	PPB	11/19/2024 19:08	SMPL
P4655-05	MC0VE8	3423	0.7722 -		SFAM01.1	PPB	11/19/2024 19:10	SMPL
P4655-06	MC0VE9	1573	0.3325 -		SFAM01.1	PPB	11/19/2024 19:12	SMPL
P4655-07	MC0VF0	2125	0.4637 -		SFAM01.1	PPB	11/19/2024 19:15	SMPL
P4655-08	MC0VF1	2674	0.5942 -		SFAM01.1	PPB	11/19/2024 19:17	SMPL
P4655-09	MC0VF5	5542	1.2758 -		SFAM01.1	PPB	11/19/2024 19:19	SMPL
P4655-10	MC0VF6	3627	0.8207 -		SFAM01.1	PPB	11/19/2024 19:21	SMPL

LB133521

INSTRUMENT ID : CV1

P4655-11	MC0VG6	8553	1.9914 -	SFAM01.1	PPB	11/19/2024 19:24	SMPL
P4655-12	MC0VG7	2099	0.4575 -	SFAM01.1	PPB	11/19/2024 19:26	SMPL
P4655-13	MC0VG8	2019	0.4385 -	SFAM01.1	PPB	11/19/2024 19:28	SMPL
P4655-14	MC0VG9	1300	0.2677 -	SFAM01.1	PPB	11/19/2024 19:30	SMPL
P4655-15	MC0VH0	4585	1.0484 -	SFAM01.1	PPB	11/19/2024 19:33	SMPL
P4655-16	MC0VJ1	1861	0.401 -	SFAM01.1	PPB	11/19/2024 19:35	SMPL
CCV016	CCV016	22064	5.2025 -	SFAM01.1	PPB	11/19/2024 19:37	SMPL
CCB016	CCB016	49	-0.0297 -	SFAM01.1	PPB	11/19/2024 19:40	SMPL
P4655-17	MC0VH1	1455	0.3045 -	SFAM01.1	PPB	11/19/2024 19:42	SMPL
P4655-18	MC0VH2	2819	0.6287 -	SFAM01.1	PPB	11/19/2024 19:44	SMPL
P4655-19	MC0VH3	2374	0.5229 -	SFAM01.1	PPB	11/19/2024 19:46	SMPL
P4655-20	MC0VH7	3192	0.7173 -	SFAM01.1	PPB	11/19/2024 19:49	SMPL
P4655-21	MC0VH7D	2739	0.6097 -	SFAM01.1	PPB	11/19/2024 19:51	SMPL
P4655-22	MC0VH7S	13811	3.2411 -	SFAM01.1	PPB	11/19/2024 19:53	SMPL
CCV017	CCV017	21861	5.1543 -	SFAM01.1	PPB	11/19/2024 19:55	SMPL
CCB017	CCB017	32	-0.0337 -	SFAM01.1	PPB	11/19/2024 19:58	SMPL