



Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1000	1000	100	90 - 110	P	11/06/2024	17:34	LB133323



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4658
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4658 **SAS No.:** P4658
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	12.2	12.0	102	80 - 120	P	11/06/2024	17:38	LB133323

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4658
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4658 **SAS No.:** P4658
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	5010	5000	100	90 - 110	P	11/06/2024	18:23	LB133323
CCV02	Lead	4970	5000	99	90 - 110	P	11/06/2024	18:39	LB133323
CCV03	Lead	4910	5000	98	90 - 110	P	11/06/2024	19:32	LB133323
CCV04	Lead	5080	5000	102	90 - 110	P	11/06/2024	20:22	LB133323
CCV05	Lead	5120	5000	102	90 - 110	P	11/06/2024	21:13	LB133323
CCV06	Lead	5030	5000	100	90 - 110	P	11/06/2024	22:11	LB133323
CCV07	Lead	5100	5000	102	90 - 110	P	11/06/2024	23:03	LB133323
CCV08	Lead	5040	5000	101	90 - 110	P	11/06/2024	23:59	LB133323
CCV09	Lead	4900	5000	98	90 - 110	P	11/07/2024	00:53	LB133323
CCV10	Lead	5270	5000	105	90 - 110	P	11/07/2024	01:49	LB133323



		Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1020	1000	102	90 - 110	P	11/07/2024	13:49	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4658
Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4658 **SAS No.:** P4658
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	11.4	12.0	95	80 - 120	P	11/07/2024	14:23	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** P4658

Contract: PORT06 **Lab Code:** CHEM **Case No.:** P4658 **SAS No.:** P4658

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4920	5000	98	90 - 110	P	11/07/2024	15:01	LB133344
CCV02	Lead	4800	5000	96	90 - 110	P	11/07/2024	15:53	LB133344
CCV03	Lead	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
CCV04	Lead	5010	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
CCV05	Lead	4880	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
CCV06	Lead	4930	5000	99	90 - 110	P	11/07/2024	19:08	LB133344
CCV07	Lead	4850	5000	97	90 - 110	P	11/07/2024	19:58	LB133344
CCV08	Lead	4840	5000	97	90 - 110	P	11/07/2024	20:58	LB133344
CCV09	Lead	4780	5000	96	90 - 110	P	11/07/2024	21:51	LB133344
CCV10	Lead	4710	5000	94	90 - 110	P	11/07/2024	22:52	LB133344
CCV11	Lead	4800	5000	96	90 - 110	P	11/07/2024	23:43	LB133344
CCV12	Lead	4820	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
CCV13	Lead	4870	5000	97	90 - 110	P	11/08/2024	00:57	LB133344