

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06

Lab Code: CHEM Case No.: P4658 SAS No.: P4658 SDG No.: P4658

Instrument ID: MSVOA\_N Calibration Date(s): 10/30/2024 10/30/2024

Heated Purge: (Y/N) N Calibration Time(s): 11:46 13:45

GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                   |        | RRF100 = VN084570.D |        | RRF050 = VN084571.D |        | RRF020 = VN084572.D |       |       |  |
|--------------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                                |        | RRF010 = VN084573.D |        | RRF005 = VN084574.D |        | RRF001 = VN084575.D |       |       |  |
| COMPOUND                       | RRF100 | RRF050              | RRF020 | RRF010              | RRF005 | RRF001              | RRF   | % RSD |  |
| Dichlorodifluoromethane        | 0.571  | 0.552               | 0.552  | 0.598               | 0.594  | 0.581               | 0.575 | 3.4   |  |
| Chloromethane                  | 0.658  | 0.672               | 0.725  | 0.871               | 0.995  | 1.789               | 0.952 | 45.2  |  |
| Vinyl Chloride                 | 0.613  | 0.605               | 0.623  | 0.636               | 0.651  | 0.581               | 0.618 | 4     |  |
| Bromomethane                   | 0.292  | 0.296               | 0.310  | 0.336               | 0.405  |                     | 0.328 | 14.2  |  |
| Chloroethane                   | 0.378  | 0.376               | 0.413  | 0.426               | 0.475  | 0.863               | 0.488 | 38.3  |  |
| Trichlorofluoromethane         | 0.971  | 0.959               | 1.017  | 1.022               | 1.070  | 1.071               | 1.018 | 4.7   |  |
| 1,1,2-Trichlorotrifluoroethane | 0.566  | 0.557               | 0.571  | 0.585               | 0.588  | 0.586               | 0.575 | 2.2   |  |
| 1,1-Dichloroethene             | 0.548  | 0.538               | 0.560  | 0.575               | 0.552  | 0.644               | 0.569 | 6.8   |  |
| Acetone                        | 0.209  | 0.204               | 0.213  | 0.223               | 0.241  | 0.338               | 0.238 | 21.3  |  |
| Carbon Disulfide               | 1.604  | 1.603               | 1.700  | 1.714               | 1.784  | 2.117               | 1.753 | 10.9  |  |
| Methyl tert-butyl Ether        | 1.773  | 1.758               | 1.802  | 1.779               | 1.786  | 1.572               | 1.745 | 4.9   |  |
| Methyl Acetate                 | 0.731  | 0.749               | 0.954  | 1.044               | 1.266  | 2.291               | 1.172 | 49.7  |  |
| Methylene Chloride             | 0.604  | 0.602               | 0.633  | 0.658               | 0.714  | 0.600               | 0.635 | 7.1   |  |
| trans-1,2-Dichloroethene       | 0.565  | 0.563               | 0.596  | 0.600               | 0.584  | 0.601               | 0.585 | 2.9   |  |
| 1,1-Dichloroethane             | 1.067  | 1.066               | 1.114  | 1.127               | 1.203  | 1.033               | 1.102 | 5.5   |  |
| Cyclohexane                    | 0.956  | 0.938               | 0.956  | 1.043               | 1.093  |                     | 0.997 | 6.8   |  |
| 2-Butanone                     | 0.316  | 0.315               | 0.348  | 0.338               | 0.370  | 0.334               | 0.337 | 6.1   |  |
| Carbon Tetrachloride           | 0.530  | 0.514               | 0.532  | 0.548               | 0.537  | 0.488               | 0.525 | 4     |  |
| cis-1,2-Dichloroethene         | 0.675  | 0.662               | 0.697  | 0.685               | 0.705  | 0.673               | 0.683 | 2.4   |  |
| Bromochloromethane             | 0.483  | 0.511               | 0.388  | 0.415               | 0.408  | 0.429               | 0.439 | 10.8  |  |
| Chloroform                     | 1.099  | 1.086               | 1.142  | 1.154               | 1.222  | 1.025               | 1.121 | 6     |  |
| 1,1,1-Trichloroethane          | 1.000  | 0.991               | 1.046  | 1.073               | 1.032  | 0.980               | 1.021 | 3.5   |  |
| Methylcyclohexane              | 0.546  | 0.509               | 0.495  | 0.487               | 0.458  | 0.371               | 0.478 | 12.5  |  |
| Benzene                        | 1.494  | 1.448               | 1.509  | 1.507               | 1.546  | 1.540               | 1.507 | 2.4   |  |
| 1,2-Dichloroethane             | 0.488  | 0.494               | 0.493  | 0.492               | 0.503  | 0.459               | 0.488 | 3.1   |  |
| Trichloroethene                | 0.339  | 0.335               | 0.345  | 0.338               | 0.341  | 0.387               | 0.348 | 5.7   |  |
| 1,2-Dichloropropane            | 0.356  | 0.348               | 0.358  | 0.357               | 0.373  | 0.330               | 0.354 | 4     |  |
| Bromodichloromethane           | 0.528  | 0.521               | 0.526  | 0.522               | 0.529  | 0.530               | 0.526 | 0.7   |  |
| 4-Methyl-2-Pentanone           | 0.424  | 0.423               | 0.416  | 0.417               | 0.412  | 0.344               | 0.406 | 7.6   |  |
| Toluene                        | 0.923  | 0.899               | 0.919  | 0.902               | 0.891  | 0.757               | 0.882 | 7.1   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: PORT06  
 Lab Code: CHEM Case No.: P4658 SAS No.: P4658 SDG No.: P4658  
 Instrument ID: MSVOA\_N Calibration Date(s): 10/30/2024 10/30/2024  
 Heated Purge: (Y/N) N Calibration Time(s): 11:46 13:45  
 GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:                |        | RRF100 = VN084570.D |        | RRF050 = VN084571.D |        | RRF020 = VN084572.D |       |       |  |
|-----------------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                             |        | RRF010 = VN084573.D |        | RRF005 = VN084574.D |        | RRF001 = VN084575.D |       |       |  |
| COMPOUND                    | RRF100 | RRF050              | RRF020 | RRF010              | RRF005 | RRF001              | RRF   | % RSD |  |
| t-1,3-Dichloropropene       | 0.552  | 0.543               | 0.538  | 0.532               | 0.547  | 0.555               | 0.544 | 1.6   |  |
| cis-1,3-Dichloropropene     | 0.592  | 0.581               | 0.582  | 0.569               | 0.584  | 0.548               | 0.576 | 2.7   |  |
| 1,1,2-Trichloroethane       | 0.336  | 0.329               | 0.334  | 0.342               | 0.342  | 0.309               | 0.332 | 3.7   |  |
| 2-Hexanone                  | 0.314  | 0.312               | 0.301  | 0.297               | 0.294  | 0.256               | 0.296 | 7.1   |  |
| Dibromochloromethane        | 0.404  | 0.394               | 0.391  | 0.393               | 0.377  | 0.312               | 0.378 | 8.9   |  |
| 1,2-Dibromoethane           | 0.340  | 0.333               | 0.341  | 0.335               | 0.355  | 0.336               | 0.340 | 2.4   |  |
| Tetrachloroethene           | 0.326  | 0.313               | 0.333  | 0.347               | 0.351  | 0.325               | 0.333 | 4.3   |  |
| Chlorobenzene               | 1.068  | 1.061               | 1.149  | 1.123               | 1.165  | 1.146               | 1.119 | 3.9   |  |
| Ethyl Benzene               | 1.957  | 1.891               | 1.928  | 1.880               | 1.849  | 1.697               | 1.867 | 4.9   |  |
| m/p-Xylenes                 | 0.737  | 0.728               | 0.737  | 0.701               | 0.683  | 0.654               | 0.707 | 4.8   |  |
| o-Xylene                    | 0.703  | 0.690               | 0.701  | 0.679               | 0.645  | 0.550               | 0.661 | 8.9   |  |
| Styrene                     | 1.223  | 1.205               | 1.206  | 1.144               | 1.093  | 1.050               | 1.154 | 6.1   |  |
| Bromoform                   | 0.287  | 0.295               | 0.298  | 0.289               | 0.302  | 0.286               | 0.293 | 2.3   |  |
| Isopropylbenzene            | 3.558  | 3.570               | 3.701  | 3.605               | 3.402  | 3.188               | 3.504 | 5.2   |  |
| 1,1,2,2-Tetrachloroethane   | 1.052  | 1.073               | 1.163  | 1.190               | 1.317  | 1.222               | 1.170 | 8.4   |  |
| 1,3-Dichlorobenzene         | 1.642  | 1.668               | 1.770  | 1.802               | 1.884  | 2.264               | 1.838 | 12.3  |  |
| 1,4-Dichlorobenzene         | 1.646  | 1.674               | 1.782  | 1.867               | 1.879  | 2.773               | 1.937 | 21.7  |  |
| 1,2-Dichlorobenzene         | 1.601  | 1.618               | 1.748  | 1.732               | 1.879  | 2.021               | 1.766 | 9.1   |  |
| 1,2-Dibromo-3-Chloropropane | 0.204  | 0.217               | 0.229  | 0.226               | 0.274  | 0.272               | 0.237 | 12.3  |  |
| 1,2,4-Trichlorobenzene      | 0.852  | 0.865               | 0.895  | 0.881               | 0.956  | 1.353               | 0.967 | 19.9  |  |
| 1,2,3-Trichlorobenzene      | 0.832  | 0.841               | 0.953  | 0.877               | 0.939  | 1.189               | 0.939 | 14.1  |  |
| 1,2-Dichloroethane-d4       | 0.689  | 0.721               | 0.708  | 0.722               | 0.771  |                     | 0.722 | 4.2   |  |
| Dibromofluoromethane        | 0.334  | 0.344               | 0.326  | 0.336               | 0.353  |                     | 0.338 | 3.1   |  |
| Toluene-d8                  | 1.267  | 1.303               | 1.216  | 1.231               | 1.217  |                     | 1.247 | 3     |  |
| 4-Bromofluorobenzene        | 0.481  | 0.493               | 0.450  | 0.451               | 0.454  |                     | 0.466 | 4.3   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.