



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110124\
Data File : VN084627.D
Acq On : 01 Nov 2024 09:48
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Title : SW846 8260
Last Update : Mon Oct 28 13:41:34 2024



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1843

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	7000	PASS
75	95	30	60	51.3	18438	PASS
95	95	100	100	100.0	35963	PASS
96	95	5	9	7.2	2604	PASS
173	174	0.00	2	1.5	400	PASS
174	95	50	100	74.6	26821	PASS
175	174	5	9	7.2	1944	PASS
176	174	95	101	95.5	25611	PASS
177	176	5	9	6.9	1760	PASS