

Surrogate Summary

SDG No.: P4658

Client: Portal Partners Tri-Venture

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4658-04	TB	1,2-Dichloroethane-d4	50	50.6	101	70 (74)	130 (125)
		Dibromofluoromethane	50	49.3	99	70 (75)	130 (124)
		Toluene-d8	50	47.6	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	45.3	91	70 (77)	130 (121)
VN1101WBL01	VN1101WBL01	1,2-Dichloroethane-d4	50	51.6	103	70 (74)	130 (125)
		Dibromofluoromethane	50	49.8	100	70 (75)	130 (124)
		Toluene-d8	50	47.4	95	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.8	94	70 (77)	130 (121)
VN1101WBS01	VN1101WBS01	1,2-Dichloroethane-d4	50	44.3	89	70 (74)	130 (125)
		Dibromofluoromethane	50	45.1	90	70 (75)	130 (124)
		Toluene-d8	50	46.1	92	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.0	94	70 (77)	130 (121)
VN1101WBSD0	VN1101WBSD01	1,2-Dichloroethane-d4	50	45.0	90	70 (74)	130 (125)
		Dibromofluoromethane	50	46.5	93	70 (75)	130 (124)
		Toluene-d8	50	46.3	93	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.5	93	70 (77)	130 (121)

