

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4658 SAS No.: P4658 SDG NO.: P4658
 Lab File ID: VN084628.D Date Analyzed: 11/01/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	184180	8.22	299507	9.10	271207	11.87
UPPER LIMIT	368360	8.724	599014	9.6	542414	12.365
LOWER LIMIT	92090	7.724	149754	8.6	135604	11.365
EPA SAMPLE NO.						
TB	177716	8.22	313388	9.10	275782	11.87
VN1101WBL01	180035	8.22	318575	9.10	285991	11.87
VN1101WBS01	190995	8.22	320452	9.10	287801	11.87
VN1101WBSD01	179647	8.22	303928	9.10	276373	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.