

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: PORT06
 Lab Code: CHEM Case No.: P4658 SAS No.: P4658 SDG NO.: P4658
 Lab File ID: VN084628.D Date Analyzed: 11/01/2024
 Instrument ID: MSVOA_N Time Analyzed: 10:22
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	136355	13.794				
UPPER LIMIT	272710	14.294				
LOWER LIMIT	68177.5	13.294				
EPA SAMPLE NO.						
TB	122691	13.79				
VN1101WBL01	125664	13.79				
VN1101WBS01	147052	13.79				
VN1101WBSD01	139639	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.