

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: PORT06
Lab Code: CHEM Case No.: P4658 SAS No.: P4658 SDG NO.: P4658
Lab File ID: VN084627.D BFB Injection Date: 11/01/2024
Instrument ID: MSVOA_N BFB Injection Time: 09:48
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.5
75	30.0 - 60.0% of mass 95	51.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	7.2
173	Less than 2.0% of mass 174	1.1 (1.5) 1
174	50.0 - 100.0% of mass 95	74.6
175	5.0 - 9.0% of mass 174	5.4 (7.2) 1
176	95.0 - 101.0% of mass 174	71.2 (95.5) 1
177	5.0 - 9.0% of mass 176	4.9 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN084628.D	11/01/2024	10:22
VN1101WBL01	VN1101WBL01	VN084630.D	11/01/2024	11:29
VN1101WBS01	VN1101WBS01	VN084631.D	11/01/2024	12:08
VN1101WBSD01	VN1101WBSD01	VN084632.D	11/01/2024	12:32
TB	P4658-04	VN084634.D	11/01/2024	13:20