

## **CASE NARRATIVE**

### **ENTACT**

**Project Name: 540 Degraw St, Brooklyn, NY - E9309**

**Project # N/A**

**Chemtech Project # P4660**

**Test Name: TCLP BNA**

### **A. Number of Samples and Date of Receipt:**

12 Solid samples were received on 10/31/2024.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: ASTM Ammonia, ASTM COD, ASTM Leach Extraction, ASTM Oil and Grease, ASTM TS, Corrosivity, Ignitability, Oil and Grease, Paint Filter, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, TCLP BNA, TCLP Extraction, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TCLP Metals Group2, TS, TVS and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_E using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for

WC-TA2-01-C [2,4,6-Tribromophenol - 113%],

WC-WOOD-01-C [2,4,6-Tribromophenol - 115%],

WC-CONCRETE-01-C [2,4,6-Tribromophenol - 121% and 2-Fluorophenol - 117%]

these surrogates did not meet the NJDKQP criteria but met the in-house criteria .

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {P4640-04MS} with File ID: BE101481.D recoveries met the requirements for all compounds except for Hexachlorobutadiene[66%] this compound did not meet the NJDKQP criteria but met the in-house criteria, while 1,4-Dichlorobenzene[52%] this compound did not meet the NJDKQP criteria and in-house criteria, due to matrix interference.

The MSD {P4640-04MSD} with File ID: BE101482.D recoveries met the requirements for all compounds except for Hexachlorobutadiene[68%] this compound did not meet the NJDKQP criteria but met the in-house criteria, while 1,4-Dichlorobenzene[52%] this compound did not meet the NJDKQP criteria and in-house criteria, due to matrix interference.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

**E. Additional Comments:**

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_