


LB133297

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.053	0.053	-109	0.000	-109	0				12
0.2	0.200	0.223	0.023	490	0.0 %	490					-2
2.5	2.500	2.450	-0.050	8334	0.0 %	8334					1
5.0	5.000	5.050	0.050	17494	0.0 %	17494					3
7.5	7.500	7.283	-0.217	25358	0.0 %	25358					1
10.0	10.000	10.143	0.143	35432	0.0 %	35432					

LB133297 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Nov 2024 11:28:30

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-109	-	0.000	7470A	PPB	05 Nov 2024 11:33:52	S	Std
0.2 - 1	502	490	-	0.200	7470A	PPB	05 Nov 2024 11:36:09	S	Std
2.5 - 1	52.5	8334	-	2.500	7470A	PPB	05 Nov 2024 11:38:25	S	Std
5.0 - 1	55	17494	-	5.000	7470A	PPB	05 Nov 2024 11:40:42	S	Std
7.5 - 1	57.5	25358	-	7.500	7470A	PPB	05 Nov 2024 11:42:58	S	Std
10.0 - 1	510	35432	-	10.000	7470A	PPB	05 Nov 2024 11:55:45	S	Std
ICV69 - 1	ICV69	13346	3.8724	-	7470A	PPB	05 Nov 2024 11:58:34	U	SMPL
ICB69 - 1	ICB69	-361	-0.0190	-	7470A	PPB	05 Nov 2024 12:00:49	U	SMPL
CCV26 - 1	CCV26	16471	4.7596	-	7470A	PPB	05 Nov 2024 12:03:05	U	SMPL
CCB26 - 1	CCB26	-363	-0.0195	-	7470A	PPB	05 Nov 2024 12:05:21	U	SMPL
CRA - 1	CRA	459	0.2138	-	7470A	PPB	05 Nov 2024 12:07:37	U	SMPL
HighStd - 1	HighStd	35445	10.1463	-	7470A	PPB	05 Nov 2024 12:09:53	U	SMPL
ChkStd - 1	ChkStd	21940	6.3122	-	7470A	PPB	05 Nov 2024 12:12:08	U	SMPL
PB164660BL - 1	PBW	-510	-0.0613	-	7470A	PPB	05 Nov 2024 12:14:26	U	SMPL
PB164660BS - 1	LCSW	13579	3.9386	-	7470A	PPB	05 Nov 2024 12:16:44	U	SMPL
P4640-04 - 1	MH-3	-367	-0.0207	-	7470A	PPB	05 Nov 2024 12:19:00	U	SMPL
P4643-04 - 1	BP-F9-ADDITIONAL	-129	0.0469	-	7470A	PPB	05 Nov 2024 12:21:18	U	SMPL
P4643-08 - 1	BP-F8	-85	0.0594	-	7470A	PPB	05 Nov 2024 12:23:35	U	SMPL
P4643-12 - 1	TP-9	-89	0.0583	-	7470A	PPB	05 Nov 2024 12:25:51	U	SMPL
P4645-04 - 1	Z-02-WC	-77	0.0617	-	7470A	PPB	05 Nov 2024 12:28:06	U	SMPL
CCV27 - 1	CCV27	17236	4.9768	-	7470A	PPB	05 Nov 2024 12:30:22	U	SMPL
CCB27 - 1	CCB27	-327	-0.0093	-	7470A	PPB	05 Nov 2024 12:32:36	U	SMPL
P4659-04 - 1	MH-2	-77	0.0617	-	7470A	PPB	05 Nov 2024 12:34:53	U	SMPL
P4660-03 - 1	WC-TA2-01-C	-55	0.0679	-	7470A	PPB	05 Nov 2024 12:37:08	U	SMPL
P4660-07 - 1	WC-WOOD-01-C	-73	0.0628	-	7470A	PPB	05 Nov 2024 12:39:24	U	SMPL
P4660-11 - 1	WC-CONCRETE-01-C	-72	0.0631	-	7470A	PPB	05 Nov 2024 12:41:39	U	SMPL
P4667-04 - 1	BP-F-6	-84	0.0597	-	7470A	PPB	05 Nov 2024 12:43:55	U	SMPL
P4667-08 - 1	BP-F-5	-78	0.0614	-	7470A	PPB	05 Nov 2024 12:46:11	U	SMPL
P4667-12 - 1	BP-F-10	-64	0.0654	-	7470A	PPB	05 Nov 2024 12:48:28	U	SMPL
P4667-16 - 1	BP-F-7	-58	0.0671	-	7470A	PPB	05 Nov 2024 12:50:44	U	SMPL
P4679-04 - 1	MH-1	-59	0.0668	-	7470A	PPB	05 Nov 2024 12:53:01	U	SMPL
P4680-04 - 1	BP-F26	0	0.0835	-	7470A	PPB	05 Nov 2024 12:55:16	U	SMPL
CCV28 - 1	CCV28	17106	4.9399	-	7470A	PPB	05 Nov 2024 12:57:32	U	SMPL
CCB28 - 1	CCB28	-300	-0.0016	-	7470A	PPB	05 Nov 2024 12:59:48	U	SMPL
P4680-08 - 1	BP-F25	-103	0.0543	-	7470A	PPB	05 Nov 2024 13:02:04	U	SMPL
P4682-02 - 1	BELL-SHOP-RAGS	-73	0.0628	-	7470A	PPB	05 Nov 2024 13:04:19	U	SMPL
P4684-01 - 1	MECHANIC-ST-SWEEPINGS	-57	0.0673	-	7470A	PPB	05 Nov 2024 13:06:35	U	SMPL
P4684-01DUP - 1	MECHANIC-ST-SWEEPINGS DUP	-28	0.0756	-	7470A	PPB	05 Nov 2024 13:08:50	U	SMPL
P4684-01MS - 1	MECHANIC-ST-SWEEPINGSMS	15929	4.8057	-	7470A	PPB	05 Nov 2024 13:11:06	U	SMPL
P4684-01MSD - 1	MECHANIC-ST-SWEEPINGSMSD	14622	4.2347	-	7470A	PPB	05 Nov 2024 13:13:22	U	SMPL
PB164687BL - 1	PBW	-305	-0.0031	-	7470A	PPB	05 Nov 2024 13:15:40	U	SMPL
PB164687BS - 1	LCSW	12294	3.5738	-	7470A	PPB	05 Nov 2024 13:22:37	U	SMPL
P4652-01 - 1	RW5-SP100-50-20241030	-251	0.0123	-	7470A	PPB	05 Nov 2024 13:24:54	U	SMPL
P4652-03 - 1	RW5-SP301-303-50-20241030	-166	0.0364	-	7470A	PPB	05 Nov 2024 13:27:11	U	SMPL
CCV29 - 1	CCV29	17742	5.1204	-	7470A	PPB	05 Nov 2024 13:32:08	U	SMPL
CCB29 - 1	CCB29	-352	-0.0164	-	7470A	PPB	05 Nov 2024 13:34:24	U	SMPL
P4652-04 - 1	RW5-SP100-90-20241030	-140	0.0438	-	7470A	PPB	05 Nov 2024 13:36:41	U	SMPL
P4652-06 - 1	RW5-SP301-303-90-20241030	-105	0.0537	-	7470A	PPB	05 Nov 2024 13:38:57	U	SMPL
P4661-03 - 1	102324-A	303	0.1696	-	7470A	PPB	05 Nov 2024 13:46:15	U	SMPL
P4661-03DUP - 1	102324-ADUP	171	0.1321	-	7470A	PPB	05 Nov 2024 13:48:30	U	SMPL
P4661-03MS - 1	102324-AMS	11553	3.3634	-	7470A	PPB	05 Nov 2024 13:50:47	U	SMPL
P4661-03MSD - 1	102324-AMSD	11171	3.2550	-	7470A	PPB	05 Nov 2024 13:53:03	U	SMPL
P4662-04DLX10 - 1	102524-D	166	0.1307	-	7470A	PPB	05 Nov 2024 13:59:42	U	SMPL
P4673-01 - 1	OUTFALL-001	165	0.1304	-	7470A	PPB	05 Nov 2024 14:01:58	U	SMPL
P4673-02 - 1	OUTFALL-002	-79	0.0611	-	7470A	PPB	05 Nov 2024 14:04:15	U	SMPL
P4673-03 - 1	OUTFALL-003	-27	0.0759	-	7470A	PPB	05 Nov 2024 14:06:30	U	SMPL
CCV30 - 1	CCV30	18033	5.2031	-	7470A	PPB	05 Nov 2024 14:08:45	U	SMPL
CCB30 - 1	CCB30	-114	0.0512	-	7470A	PPB	05 Nov 2024 14:11:01	U	SMPL
P4673-04 - 1	OUTFALL-004	-110	0.0523	-	7470A	PPB	05 Nov 2024 14:13:18	U	SMPL
P4673-05 - 1	OUTFALL-005	-87	0.0588	-	7470A	PPB	05 Nov 2024 14:15:34	U	SMPL
P4673-06 - 1	OUTFALL-006	31	0.0923	-	7470A	PPB	05 Nov 2024 14:17:49	U	SMPL
P4673-07 - 1	OUTFALL-007	-111	0.0520	-	7470A	PPB	05 Nov 2024 14:20:04	U	SMPL
P4674-01 - 1	OUTFALL-001	-7	0.0815	-	7470A	PPB	05 Nov 2024 14:22:20	U	SMPL
P4684-03 - 1	MECHANIC-ST-OIL-WATER	981	0.3620	-	7470A	PPB	05 Nov 2024 14:24:36	U	SMPL
PB164560TB - 1	PB164560TB	-198	0.0273	-	7470A	PPB	05 Nov 2024 14:26:52	U	SMPL
CCV31 - 1	CCV31	18853	5.4359	-	7470A	PPB	05 Nov 2024 14:29:09	U	SMPL
CCB31 - 1	CCB31	-146	0.0421	-	7470A	PPB	05 Nov 2024 14:31:26	U	SMPL
P4684-01LX5 - 1		-68	0.0642	-	7470A	PPB	05 Nov 2024 14:33:43	U	SMPL
P4684-01A - 1		14276	4.1365	-	7470A	PPB	05 Nov 2024 14:35:59	U	SMPL
P4661-03LX5 - 1		-91	0.0577	-	7470A	PPB	05 Nov 2024 14:38:15	U	SMPL
P4661-03A - 1		11752	3.4199	-	7470A	PPB	05 Nov 2024 14:40:33	U	SMPL
CCV32 - 1	CCV32	17695	5.1071	-	7470A	PPB	05 Nov 2024 14:42:49	U	SMPL
CCB32 - 1	CCB32	-159	0.0384	-	7470A	PPB	05 Nov 2024 14:45:07	U	SMPL