

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV69	Mercury	3.87	4.0	97	90 - 110	CV	11/05/2024	11:58	LB133297

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV26	Mercury	4.76	5.0	95	90 - 110	CV	11/05/2024	12:03	LB133297
CCV27	Mercury	4.98	5.0	100	90 - 110	CV	11/05/2024	12:30	LB133297
CCV28	Mercury	4.94	5.0	99	90 - 110	CV	11/05/2024	12:57	LB133297
CCV29	Mercury	5.12	5.0	102	90 - 110	CV	11/05/2024	13:32	LB133297
CCV30	Mercury	5.20	5.0	104	90 - 110	CV	11/05/2024	14:08	LB133297
CCV31	Mercury	5.44	5.0	109	90 - 110	CV	11/05/2024	14:29	LB133297
CCV32	Mercury	5.11	5.0	102	90 - 110	CV	11/05/2024	14:42	LB133297

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: P4660
Contract: ENTA05 Lab Code: CHEM Case No.: P4660 SAS No.: P4660
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1030	1000	103	90 - 110	P	11/07/2024	13:49	LB133344
	Barium	521	520	100	90 - 110	P	11/07/2024	13:49	LB133344
	Cadmium	509	510	100	90 - 110	P	11/07/2024	13:49	LB133344
	Chromium	531	520	102	90 - 110	P	11/07/2024	13:49	LB133344
	Copper	531	510	104	90 - 110	P	11/07/2024	13:49	LB133344
	Lead	1020	1000	102	90 - 110	P	11/07/2024	13:49	LB133344
	Nickel	518	530	98	90 - 110	P	11/07/2024	13:49	LB133344
	Selenium	1040	1000	104	90 - 110	P	11/07/2024	13:49	LB133344
	Silver	256	250	102	90 - 110	P	11/07/2024	13:49	LB133344
	Zinc	1050	1000	105	90 - 110	P	11/07/2024	13:49	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT SDG No.: P4660
Contract: ENTA05 Lab Code: CHEM Case No.: P4660 SAS No.: P4660
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	18.0	20.0	90	80 - 120	P	11/07/2024	14:23	LB133344
	Barium	100	100	100	80 - 120	P	11/07/2024	14:23	LB133344
	Cadmium	6.52	6.0	109	80 - 120	P	11/07/2024	14:23	LB133344
	Chromium	9.52	10.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Copper	22.7	20.0	114	80 - 120	P	11/07/2024	14:23	LB133344
	Lead	11.4	12.0	95	80 - 120	P	11/07/2024	14:23	LB133344
	Nickel	38.5	40.0	96	80 - 120	P	11/07/2024	14:23	LB133344
	Selenium	20.8	20.0	104	80 - 120	P	11/07/2024	14:23	LB133344
	Silver	10.8	10.0	108	80 - 120	P	11/07/2024	14:23	LB133344
	Zinc	42.0	40.0	105	80 - 120	P	11/07/2024	14:23	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5040	5000	101	90 - 110	P	11/07/2024	15:01	LB133344
	Barium	9710	10000	97	90 - 110	P	11/07/2024	15:01	LB133344
	Cadmium	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Chromium	992	1000	99	90 - 110	P	11/07/2024	15:01	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	15:01	LB133344
	Lead	4920	5000	98	90 - 110	P	11/07/2024	15:01	LB133344
	Nickel	2450	2500	98	90 - 110	P	11/07/2024	15:01	LB133344
	Selenium	5150	5000	103	90 - 110	P	11/07/2024	15:01	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	15:01	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	15:01	LB133344
CCV02	Arsenic	4950	5000	99	90 - 110	P	11/07/2024	15:53	LB133344
	Barium	9390	10000	94	90 - 110	P	11/07/2024	15:53	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	15:53	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	15:53	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/07/2024	15:53	LB133344
	Selenium	5050	5000	101	90 - 110	P	11/07/2024	15:53	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	15:53	LB133344
	Zinc	2480	2500	99	90 - 110	P	11/07/2024	15:53	LB133344
CCV03	Arsenic	5190	5000	104	90 - 110	P	11/07/2024	16:45	LB133344
	Barium	9850	10000	98	90 - 110	P	11/07/2024	16:45	LB133344
	Cadmium	2470	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Chromium	993	1000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Copper	1300	1250	104	90 - 110	P	11/07/2024	16:45	LB133344
	Lead	4970	5000	99	90 - 110	P	11/07/2024	16:45	LB133344
	Nickel	2480	2500	99	90 - 110	P	11/07/2024	16:45	LB133344
	Selenium	5320	5000	106	90 - 110	P	11/07/2024	16:45	LB133344
	Silver	1240	1250	100	90 - 110	P	11/07/2024	16:45	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	16:45	LB133344
CCV04	Arsenic	5210	5000	104	90 - 110	P	11/07/2024	17:56	LB133344
	Barium	9810	10000	98	90 - 110	P	11/07/2024	17:56	LB133344
	Cadmium	2500	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Chromium	1020	1000	102	90 - 110	P	11/07/2024	17:56	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660

Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV04	Copper	1300	1250	104	90 - 110	P	11/07/2024	17:56	LB133344
	Lead	5010	5000	100	90 - 110	P	11/07/2024	17:56	LB133344
	Nickel	2490	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
	Selenium	5390	5000	108	90 - 110	P	11/07/2024	17:56	LB133344
	Silver	1260	1250	101	90 - 110	P	11/07/2024	17:56	LB133344
	Zinc	2510	2500	100	90 - 110	P	11/07/2024	17:56	LB133344
CCV05	Arsenic	5120	5000	102	90 - 110	P	11/07/2024	18:32	LB133344
	Barium	9790	10000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Cadmium	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Chromium	988	1000	99	90 - 110	P	11/07/2024	18:32	LB133344
	Copper	1280	1250	102	90 - 110	P	11/07/2024	18:32	LB133344
	Lead	4880	5000	98	90 - 110	P	11/07/2024	18:32	LB133344
	Nickel	2430	2500	97	90 - 110	P	11/07/2024	18:32	LB133344
	Selenium	5290	5000	106	90 - 110	P	11/07/2024	18:32	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	18:32	LB133344
	Zinc	2500	2500	100	90 - 110	P	11/07/2024	18:32	LB133344
	Arsenic	5110	5000	102	90 - 110	P	11/07/2024	19:08	LB133344
	Barium	9640	10000	96	90 - 110	P	11/07/2024	19:08	LB133344
CCV06	Cadmium	2460	2500	99	90 - 110	P	11/07/2024	19:08	LB133344
	Chromium	994	1000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	19:08	LB133344
	Lead	4930	5000	99	90 - 110	P	11/07/2024	19:08	LB133344
	Nickel	2460	2500	98	90 - 110	P	11/07/2024	19:08	LB133344
	Selenium	5260	5000	105	90 - 110	P	11/07/2024	19:08	LB133344
	Silver	1240	1250	99	90 - 110	P	11/07/2024	19:08	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:08	LB133344
	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	19:58	LB133344
	Barium	9400	10000	94	90 - 110	P	11/07/2024	19:58	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/07/2024	19:58	LB133344
	Chromium	986	1000	99	90 - 110	P	11/07/2024	19:58	LB133344
CCV07	Copper	1270	1250	101	90 - 110	P	11/07/2024	19:58	LB133344
	Lead	4850	5000	97	90 - 110	P	11/07/2024	19:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	19:58	LB133344
	Selenium	5270	5000	105	90 - 110	P	11/07/2024	19:58	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV07	Silver	1230	1250	98	90 - 110	P	11/07/2024	19:58	LB133344
	Zinc	2490	2500	100	90 - 110	P	11/07/2024	19:58	LB133344
CCV08	Arsenic	5100	5000	102	90 - 110	P	11/07/2024	20:58	LB133344
	Barium	9480	10000	95	90 - 110	P	11/07/2024	20:58	LB133344
	Cadmium	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Chromium	989	1000	99	90 - 110	P	11/07/2024	20:58	LB133344
	Copper	1270	1250	102	90 - 110	P	11/07/2024	20:58	LB133344
	Lead	4840	5000	97	90 - 110	P	11/07/2024	20:58	LB133344
	Nickel	2410	2500	96	90 - 110	P	11/07/2024	20:58	LB133344
	Selenium	5290	5000	106	90 - 110	P	11/07/2024	20:58	LB133344
	Silver	1250	1250	100	90 - 110	P	11/07/2024	20:58	LB133344
	Zinc	2530	2500	101	90 - 110	P	11/07/2024	20:58	LB133344
CCV09	Arsenic	5030	5000	100	90 - 110	P	11/07/2024	21:51	LB133344
	Barium	9510	10000	95	90 - 110	P	11/07/2024	21:51	LB133344
	Cadmium	2390	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Chromium	966	1000	97	90 - 110	P	11/07/2024	21:51	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	21:51	LB133344
	Lead	4780	5000	96	90 - 110	P	11/07/2024	21:51	LB133344
	Nickel	2380	2500	95	90 - 110	P	11/07/2024	21:51	LB133344
	Selenium	5170	5000	103	90 - 110	P	11/07/2024	21:51	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	21:51	LB133344
	Zinc	2670	2500	107	90 - 110	P	11/07/2024	21:51	LB133344
CCV10	Arsenic	5000	5000	100	90 - 110	P	11/07/2024	22:52	LB133344
	Barium	9460	10000	95	90 - 110	P	11/07/2024	22:52	LB133344
	Cadmium	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Chromium	963	1000	96	90 - 110	P	11/07/2024	22:52	LB133344
	Copper	1240	1250	99	90 - 110	P	11/07/2024	22:52	LB133344
	Lead	4710	5000	94	90 - 110	P	11/07/2024	22:52	LB133344
	Nickel	2340	2500	94	90 - 110	P	11/07/2024	22:52	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	22:52	LB133344
	Silver	1220	1250	98	90 - 110	P	11/07/2024	22:52	LB133344
	Zinc	2400	2500	96	90 - 110	P	11/07/2024	22:52	LB133344
CCV11	Arsenic	5020	5000	100	90 - 110	P	11/07/2024	23:43	LB133344
	Barium	9690	10000	97	90 - 110	P	11/07/2024	23:43	LB133344

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV11	Cadmium	2390	2500	96	90 - 110	P	11/07/2024	23:43	LB133344
	Chromium	962	1000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Copper	1250	1250	100	90 - 110	P	11/07/2024	23:43	LB133344
	Lead	4800	5000	96	90 - 110	P	11/07/2024	23:43	LB133344
	Nickel	2390	2500	95	90 - 110	P	11/07/2024	23:43	LB133344
	Selenium	5160	5000	103	90 - 110	P	11/07/2024	23:43	LB133344
	Silver	1210	1250	97	90 - 110	P	11/07/2024	23:43	LB133344
	Zinc	2430	2500	97	90 - 110	P	11/07/2024	23:43	LB133344
CCV12	Arsenic	5150	5000	103	90 - 110	P	11/08/2024	00:39	LB133344
	Barium	9830	10000	98	90 - 110	P	11/08/2024	00:39	LB133344
	Cadmium	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Chromium	965	1000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Copper	1280	1250	102	90 - 110	P	11/08/2024	00:39	LB133344
	Lead	4820	5000	96	90 - 110	P	11/08/2024	00:39	LB133344
	Nickel	2390	2500	96	90 - 110	P	11/08/2024	00:39	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:39	LB133344
CCV13	Silver	1230	1250	98	90 - 110	P	11/08/2024	00:39	LB133344
	Zinc	2380	2500	95	90 - 110	P	11/08/2024	00:39	LB133344
	Arsenic	5160	5000	103	90 - 110	P	11/08/2024	00:57	LB133344
	Barium	9650	10000	96	90 - 110	P	11/08/2024	00:57	LB133344
	Cadmium	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Chromium	980	1000	98	90 - 110	P	11/08/2024	00:57	LB133344
	Copper	1290	1250	103	90 - 110	P	11/08/2024	00:57	LB133344
	Lead	4870	5000	97	90 - 110	P	11/08/2024	00:57	LB133344
	Nickel	2420	2500	97	90 - 110	P	11/08/2024	00:57	LB133344
	Selenium	5340	5000	107	90 - 110	P	11/08/2024	00:57	LB133344
	Silver	1230	1250	99	90 - 110	P	11/08/2024	00:57	LB133344
	Zinc	2470	2500	99	90 - 110	P	11/08/2024	00:57	LB133344