



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	107	40 - 160	CV	11/05/2024	12:07	LB133297
CRI01	Arsenic	19.9	20.0	99	40 - 160	P	11/07/2024	14:42	LB133344
	Barium	99.8	100	100	40 - 160	P	11/07/2024	14:42	LB133344
	Cadmium	6.45	6.0	108	40 - 160	P	11/07/2024	14:42	LB133344
	Chromium	9.25	10.0	92	40 - 160	P	11/07/2024	14:42	LB133344
	Copper	22.5	20.0	112	40 - 160	P	11/07/2024	14:42	LB133344
	Lead	11.3	12.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Nickel	38.8	40.0	97	40 - 160	P	11/07/2024	14:42	LB133344
	Selenium	18.8	20.0	94	40 - 160	P	11/07/2024	14:42	LB133344
	Silver	10.7	10.0	107	40 - 160	P	11/07/2024	14:42	LB133344
	Zinc	41.4	40.0	104	40 - 160	P	11/07/2024	14:42	LB133344