



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	P4660				
Contract:	ENTA05	Lab Code:	CHEM		Case No.:	P4660	SAS No.:	P4660		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB69	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:00	LB133297	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: P4660							
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	P4660	SAS No.:	P4660		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB26	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:05	LB133297
CCB27	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:32	LB133297
CCB28	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	12:59	LB133297
CCB29	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	13:34	LB133297
CCB30	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:11	LB133297
CCB31	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:31	LB133297
CCB32	Mercury	0.20	+/-0.20	U	0.20	CV	11/05/2024	14:45	LB133297

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>P4660</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4660</u>	SAS No.: <u>P4660</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	14:38	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	14:38	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	14:38	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	14:38	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	14:38	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	14:38	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT			SDG No.:	P4660				
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	P4660		SAS No.:	P4660	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:05	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:05	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:05	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:05	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:05	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:05	LB133344
CCB02	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	15:58	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	15:58	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	15:58	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	15:58	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	15:58	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	15:58	LB133344
CCB03	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	16:49	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	16:49	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	16:49	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	16:49	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	16:49	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	16:49	LB133344
CCB04	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:04	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:04	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:04	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:04	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:04	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: P4660							
Contract: ENTA05	Lab Code: CHEM	Case No.: P4660	SAS No.: P4660						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:04	LB133344
CCB05	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	18:37	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	18:37	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	18:37	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	18:37	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	18:37	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	18:37	LB133344
CCB06	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	19:12	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	19:12	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	19:12	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	19:12	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	19:12	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	19:12	LB133344
CCB07	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	20:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	20:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	20:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	20:02	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	20:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	20:02	LB133344
CCB08	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:02	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:02	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:02	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:02	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	ENTACT				SDG No.:	P4660			
Contract:	ENTA05	Lab Code:	CHEM	Case No.:	P4660	SAS No.:	P4660		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:02	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:02	LB133344
CCB09	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	21:55	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	21:55	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	21:55	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	21:55	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	21:55	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	21:55	LB133344
CCB10	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	22:56	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	22:56	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	22:56	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	22:56	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	22:56	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	22:56	LB133344
CCB11	Arsenic	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Barium	100	+/-100	U	100	P	11/07/2024	23:47	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/07/2024	23:47	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/07/2024	23:47	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/07/2024	23:47	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/07/2024	23:47	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/07/2024	23:47	LB133344
CCB12	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	00:43	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	00:43	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	00:43	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>P4660</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P4660</u>	SAS No.: <u>P4660</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB12	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	00:43	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	00:43	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	00:43	LB133344
CCB13	Arsenic	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Barium	100	+/-100	U	100	P	11/08/2024	01:01	LB133344
	Cadmium	6.00	+/-6.00	U	6.00	P	11/08/2024	01:01	LB133344
	Chromium	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Copper	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Lead	12.0	+/-12.0	U	12.0	P	11/08/2024	01:01	LB133344
	Nickel	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344
	Selenium	20.0	+/-20.0	U	20.0	P	11/08/2024	01:01	LB133344
	Silver	10.0	+/-10.0	U	10.0	P	11/08/2024	01:01	LB133344
	Zinc	40.0	+/-40.0	U	40.0	P	11/08/2024	01:01	LB133344