

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	4.03			-20	20	11/07/2024	14:47	LB133344
	Barium	1.97	6.0	33	-94	106	11/07/2024	14:47	LB133344
	Cadmium	6.27	1.0	627	-5	7	11/07/2024	14:47	LB133344
	Chromium	54.7	52.0	105	42	62	11/07/2024	14:47	LB133344
	Copper	6.59	2.0	330	-18	22	11/07/2024	14:47	LB133344
	Lead	8.91			-12	12	11/07/2024	14:47	LB133344
	Nickel	2.14	2.0	107	-38	42	11/07/2024	14:47	LB133344
	Selenium	-18.3			-20	20	11/07/2024	14:47	LB133344
	Silver	0.85			-10	10	11/07/2024	14:47	LB133344
	Zinc	6.19			-40	40	11/07/2024	14:47	LB133344
ICSAB01	Arsenic	116	104	112	88.4	120	11/07/2024	14:51	LB133344
	Barium	504	537	94	437	637	11/07/2024	14:51	LB133344
	Cadmium	999	972	103	826	1120	11/07/2024	14:51	LB133344
	Chromium	556	542	103	460	624	11/07/2024	14:51	LB133344
	Copper	493	511	96	434	588	11/07/2024	14:51	LB133344
	Lead	56.6	49.0	116	37	61	11/07/2024	14:51	LB133344
	Nickel	991	954	104	810	1100	11/07/2024	14:51	LB133344
	Selenium	32.9	46.0	72	26	66	11/07/2024	14:51	LB133344
	Silver	201	201	100	170	232	11/07/2024	14:51	LB133344
	Zinc	835	952	88	809	1095	11/07/2024	14:51	LB133344