

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: P4660

Contract: ENTA05

Lab Code: CHEM

Case No.: P4660

SAS No.: P4660

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164660BS Mercury	ug/L	4.0	3.94		98	80 - 120	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** P4660
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** P4660 **SAS No.:** P4660

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB164665BS							
Arsenic	ug/L	4000	4120		103	80 - 120	P
Barium	ug/L	1000	970		97	80 - 120	P
Cadmium	ug/L	1000	963		96	80 - 120	P
Chromium	ug/L	2000	2010		100	80 - 120	P
Copper	ug/L	1500	1590		106	80 - 120	P
Lead	ug/L	5000	4920		98	80 - 120	P
Nickel	ug/L	2500	2460		98	80 - 120	P
Selenium	ug/L	10000	10400		104	80 - 120	P
Silver	ug/L	380	378		100	80 - 120	P
Zinc	ug/L	1000	1000		100	80 - 120	P