

Surrogate Summary

SDG No.: P4660

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107183.D	PIBLK-PO107183.D	Tetrachloro-m-xylene	1	20	22.5	112		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	23.4	117		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	22.1	110		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.8	114		70 (60)	130 (140)
I.BLK-PO107633.D	PIBLK-PO107633.D	Tetrachloro-m-xylene	1	20	19.4	97		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	21.3	106		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.4	97		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	22.3	111		70 (60)	130 (140)
PB164638BL	PB164638BL	Tetrachloro-m-xylene	1	20	20.4	102		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	22.3	111		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.9	115		30 (32)	150 (175)
PB164638BS	PB164638BS	Tetrachloro-m-xylene	1	20	20.2	101		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	21.9	109		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.9	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	22.6	113		30 (32)	150 (175)
P4660-02	WC-TA2-01-C	Tetrachloro-m-xylene	1	20	13.3	67		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.3	86		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	15.7	79		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	17.9	89		30 (32)	150 (175)
P4660-06	WC-WOOD-01-C	Tetrachloro-m-xylene	1	20	10.5	53		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	17.0	85		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	11.4	57		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	14.6	73		30 (32)	150 (175)
P4660-10	WC-CONCRETE-01-C	Tetrachloro-m-xylene	1	20	6.73	34		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	13.7	68		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	7.09	35		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	10.4	52		30 (32)	150 (175)
I.BLK-PO107648.D	PIBLK-PO107648.D	Tetrachloro-m-xylene	1	20	19.3	96		70 (60)	130 (140)
		Decachlorobiphenyl	1	20	20.8	104		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.7	98		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.3	107		70 (60)	130 (140)
P4680-01MS	BP-F26MS	Tetrachloro-m-xylene	1	20	19.6	98		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.1	95		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.9	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.1	96		30 (32)	150 (175)
P4680-01MSD	BP-F26MSD	Tetrachloro-m-xylene	1	20	18.8	94		30 (32)	150 (144)
		Decachlorobiphenyl	1	20	19.2	96		30 (32)	150 (175)
		Tetrachloro-m-xylene	2	20	18.9	95		30 (32)	150 (144)
		Decachlorobiphenyl	2	20	19.2	96		30 (32)	150 (175)
I.BLK-PO107663.D	PIBLK-PO107663.D	Tetrachloro-m-xylene	1	20	19.5	97		70 (60)	130 (140)

() = LABORATORY INHOUSE LIMIT

Surrogate Summary

SDG No.: P4660

Client: ENTACT

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO107663.D	PIBLK-PO107663.D	Decachlorobiphenyl	1	20	20.9	105		70 (60)	130 (140)
		Tetrachloro-m-xylene	2	20	19.7	98		70 (60)	130 (140)
		Decachlorobiphenyl	2	20	21.9	109		70 (60)	130 (140)