

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4660

Client: ENTACT

Analytical Method: **8082A** **Datafile :** PO107635.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB164638BS	AR1016	166.6	148	ug/kg	89				40 (71)	140 (120)	
	AR1260	166.6	150	ug/kg	90				40 (65)	140 (130)	